

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM081020\
 Data File : BM027025.D
 Acq On : 10 Aug 2020 17:19
 Operator : JU/CG
 Sample : PB130842BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK42

Quant Time: Aug 10 17:51:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM080720MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 07 03:56:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	94825	20.00	ng/ul	0.00
18) Naphthalene-d8	10.42	136	370181	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.28	164	270707	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.02	188	606602	20.00	ng/ul	0.00
78) Chrysene-d12	21.23	240	627207	20.00	ng/ul	0.01
86) Perylene-d12	23.43	264	579995	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.20	96	12467	5.97	ng/uL	0.00
5) Phenol-d5	6.82	99	242573	32.90	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.98	67	175815	34.81	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	195674	33.65	ng/ul	0.00
13) 4-Methylphenol-d8	8.35	113	198548	34.54	ng/ul	0.00
19) Nitrobenzene-d5	8.79	128	96105	32.76	ng/ul	0.00
22) 2-Nitrophenol-d4	9.50	143	106989	35.19	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.04	165	216797	33.92	ng/ul	0.00
29) 4-Chloroaniline-d4	10.55	131	287087	38.65	ng/ul	0.00
44) Dimethylphthalate-d6	13.69	166	765031	35.44	ng/ul	0.00
47) Acenaphthylene-d8	13.97	160	908664	35.32	ng/ul	0.00
52) 4-Nitrophenol-d4	14.47	143	109080	29.93	ng/ul	0.00
58) Fluorene-d10	15.27	176	663852	35.28	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.39	200	104265	31.47	ng/ul	0.00
71) Anthracene-d10	17.12	188	1067296	35.71	ng/ul	0.00
79) Pyrene-d10	19.42	212	1215270	38.85	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.29	264	1155210	35.68	ng/ul	0.02

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM081020\
Data File : BM027025.D
Acq On : 10 Aug 2020 17:19
Operator : JU/CG
Sample : PB130842BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK42

Quant Time: Aug 10 17:51:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM080720MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 07 03:56:08 2020
Response via : Initial Calibration

