

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
 Data File : BMCCC258.D
 Acq On : 13 Aug 2022 14:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 14 23:04:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 11:23:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.810	152	215461	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	921541	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	546937	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	1100809	20.000	ng/u1	-0.01
79) Chrysene-d12	21.374	240	1111929	20.000	ng/u1	-0.01
88) Perylene-d12	23.709	264	1147659	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	47770	8.343	ng/uL	0.00
4) Pyridine-d5	3.640	84	301245	19.204	ng/u1	0.00
7) Phenol-d5	6.993	99	353599	19.570	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.146	67	237191	19.692	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	310798	20.723	ng/u1	-0.01
15) 4-Methylphenol-d8	8.540	113	286668	19.584	ng/u1	-0.01
21) Nitrobenzene-d5	8.981	128	152780	20.922	ng/u1	0.00
24) 2-Nitrophenol-d4	9.698	143	153558	21.584	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.239	165	268073	20.833	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.763	131	440542	20.224	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	754824	20.260	ng/u1	-0.01
49) Acenaphthylene-d8	14.145	160	966077	20.323	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.680	143	136279	17.813	ng/u1	-0.02
60) Fluorene-d10	15.445	176	677650	20.128	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.574	200	109782	18.553	ng/u1	0.00
73) Anthracene-d10	17.292	188	1033312	20.762	ng/u1	-0.01
81) Pyrene-d10	19.586	212	1170792	18.999	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.556	264	1161518	19.935	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	48734	8.213	ng/uL	94
5) Pyridine	3.658	79	323015	19.623	ng/u1	91
6) Benzaldehyde	6.951	77	164781	21.729	ng/u1	94
8) Phenol	7.022	94	390916	19.770	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.240	93	310398	19.526	ng/u1	96
12) 2-Chlorophenol	7.375	128	318840	20.724	ng/u1	99
13) 2-Methylphenol	8.269	108	289941	19.628	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.334	45	200898	9.574	ng/u1	98
16) Acetophenone	8.640	105	467762	19.904	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.622	70	227737	19.432	ng/u1	95
18) 4-Methylphenol	8.604	108	315064	19.882	ng/u1	96
19) Hexachloroethane	8.881	117	130562	20.568	ng/u1	89
22) Nitrobenzene	9.022	77	343701	20.568	ng/u1	99
23) Isophorone	9.545	82	612566	19.419	ng/u1	97
25) 2-Nitrophenol	9.734	139	168898	21.399	ng/u1	96
26) 2,4-Dimethylphenol	9.798	107	333215	20.660	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.034	93	393523	19.780	ng/u1	99
29) 2,4-Dichlorophenol	10.269	162	280023	20.967	ng/u1	97
30) Naphthalene	10.663	128	1015672	20.881	ng/u1	99
32) 4-Chloroaniline	10.787	127	429396	19.844	ng/u1	99
33) Hexachlorobutadiene	10.939	225	170177	20.640	ng/u1	99
34) Caprolactam	11.575	113	82675	18.731	ng/u1	85
35) 4-Chloro-3-methylphenol	11.922	107	284170	20.463	ng/u1	95
36) 2-Methylnaphthalene	12.275	142	644240	20.244	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
 Data File : BMCCC258.D
 Acq On : 13 Aug 2022 14:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 45 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Aug 14 23:04:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 11:23:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	652889	20.353	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.639	216	311339	19.997	ng/ul	98
40) Hexachlorocyclopentadiene	12.616	237	163328	15.751	ng/ul	99
41) 2,4,6-Trichlorophenol	12.892	196	202465	20.321	ng/ul	98
42) 2,4,5-Trichlorophenol	12.969	196	212516	19.990	ng/ul	99
43) 1,1'-Biphenyl	13.286	154	844406	20.058	ng/ul	98
44) 2-Chloronaphthalene	13.328	162	660045	20.430	ng/ul	99
45) 2-Nitroaniline	13.551	65	176946	19.866	ng/ul	97
47) Dimethylphthalate	13.916	163	765504	20.427	ng/ul	99
48) 2,6-Dinitrotoluene	14.045	165	150152	20.181	ng/ul	94
50) Acenaphthylene	14.175	152	1082593	20.588	ng/ul	100
51) 3-Nitroaniline	14.375	138	119629	14.549	ng/ul	99
52) Acenaphthene	14.516	153	694968	20.372	ng/ul	99
53) 2,4-Dinitrophenol	14.586	184	67333	15.939	ng/ul	99
55) 4-Nitrophenol	14.698	109	106425	18.188	ng/ul	84
56) Dibenzofuran	14.851	168	975889	20.675	ng/ul	100
57) 2,4-Dinitrotoluene	14.827	165	218553	20.839	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.080	232	169719	20.319	ng/ul#	91
59) Diethylphthalate	15.280	149	761689	20.094	ng/ul	99
61) Fluorene	15.498	166	782731	20.300	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.498	204	365791	19.800	ng/ul	95
63) 4-Nitroaniline	15.539	138	93758	12.386	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.586	198	114074	19.266	ng/ul	96
67) N-Nitrosodiphenylamine	15.716	169	646347	19.972	ng/ul	99
68) 4-Bromophenyl-phenylether	16.386	248	209666	19.428	ng/ul	96
69) Hexachlorobenzene	16.492	284	260218	19.790	ng/ul	96
70) Atrazine	16.668	200	215400	19.190	ng/ul	98
71) Pentachlorophenol	16.851	266	129801	16.159	ng/ul	97
72) Phenanthrene	17.239	178	1213001	20.559	ng/ul	99
74) Anthracene	17.327	178	1230082	20.777	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.251	216	323135	18.876	ng/uL	97
76) Pentachlorobenzene	14.769	250	309438	19.941	ng/uL	99
77) Carbazole	17.604	167	1137099	20.795	ng/ul	99
78) Di-n-butylphthalate	18.168	149	1240553	20.108	ng/ul	100
80) Fluoranthene	19.251	202	1380818	18.754	ng/ul	99
82) Pyrene	19.615	202	1451630	19.128	ng/ul	97
83) Butylbenzylphthalate	20.515	149	550028	19.927	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.298	252	435499	21.224	ng/ul	98
85) Benzo(a)anthracene	21.362	228	1405300	20.080	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.292	149	768283	19.264	ng/ul	99
87) Chrysene	21.415	228	1365655	19.999	ng/ul	99
89) Di-n-octyl phthalate	22.192	149	1312182	17.705	ng/ul	100
90) Benzo(b)fluoranthene	22.997	252	1484127	19.377	ng/ul	98
91) Benzo(k)fluoranthene	23.045	252	1371414	18.869	ng/ul	98
93) Benzo(a)pyrene	23.603	252	1448342	19.853	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.127	276	1687950	21.707	ng/ul	94
95) Dibenzo(a,h)anthracene	26.144	278	1454161	21.797	ng/ul	97
96) Benzo(g,h,i)perylene	26.868	276	1430706	22.202	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 14 23:04:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 12 11:23:11 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.213 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 88 Resp: 48734
Ion Ratio Lower Upper
88 100
43 29.8 27.8 41.8
58 73.5 61.8 92.8



#5
Pyridine
Concen: 19.623 ng/uL
RT: 3.658 min Scan# 97
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion: 79 Resp: 323015
Ion Ratio Lower Upper
79 100
52 74.9 67.8 101.6
51 36.0 31.4 47.0









#6
Benzaldehyde
Concen: 21.729 ng/ul
RT: 6.951 min Scan# 659
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 164781
Ion Ratio Lower Upper
77 100
105 100.1 84.6 127.0
106 94.5 80.3 120.5



#8
Phenol
Concen: 19.770 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion: 94 Resp: 390916
Ion Ratio Lower Upper
94 100
65 28.8 21.3 31.9
66 35.1 27.4 41.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.526 ng/ul
 RT: 7.240 min Scan# 707
 Delta R.T. -0.006 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion: 93 Resp: 310398
 Ion Ratio Lower Upper
 93 100
 63 75.1 64.0 96.0
 95 31.5 25.8 38.6



#12
 2-Chlorophenol
 Concen: 20.724 ng/ul
 RT: 7.375 min Scan# 729
 Delta R.T. -0.012 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34

Tgt Ion:128 Resp: 318840
 Ion Ratio Lower Upper
 128 100
 64 45.6 36.9 55.3
 130 32.4 25.6 38.4









#13
2-Methylphenol
Concen: 19.628 ng/ul
RT: 8.269 min Scan# 881
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 289941
Ion Ratio Lower Upper
108 100
107 88.9 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 9.574 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. -0.018 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion: 45 Resp: 200898
Ion Ratio Lower Upper
45 100
77 14.6 10.6 15.8
79 10.4 8.3 12.5









#16
Acetophenone
Concen: 19.904 ng/ul
RT: 8.640 min Scan# 946
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 467762
Ion Ratio Lower Upper
105 100
77 76.9 59.8 89.6
51 26.2 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 19.432 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion: 70 Resp: 227737
Ion Ratio Lower Upper
70 100
42 57.3 49.6 74.4
101 9.5 8.1 12.1
130 21.2 18.4 27.6









#18
4-Methylphenol
Concen: 19.882 ng/ul
RT: 8.604 min Scan# 911
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 315064
Ion Ratio Lower Upper
108 100
107 108.3 90.3 135.5



#19
Hexachloroethane
Concen: 20.568 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:117 Resp: 130562
Ion Ratio Lower Upper
117 100
201 90.2 81.8 122.6
199 57.2 51.6 77.4









#22
Nitrobenzene
Concen: 20.568 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 343701
Ion Ratio Lower Upper
77 100
123 47.6 37.4 56.0
65 13.0 10.8 16.2



#23
Isophorone
Concen: 19.419 ng/ul
RT: 9.545 min Scan# 1098
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion: 82 Resp: 612566
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 17.6 15.3 22.9









#25
2-Nitrophenol
Concen: 21.399 ng/ul
RT: 9.734 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 168898
Ion Ratio Lower Upper
139 100
65 45.9 39.8 59.6
109 28.2 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 20.660 ng/ul
RT: 9.798 min Scan# 1141
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:107 Resp: 333215
Ion Ratio Lower Upper
107 100
121 50.5 42.7 64.1
122 87.8 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 19.780 ng/ul

RT: 10.034 min Scan# 1182

Delta R.T. -0.006 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Instrument :

BNA_M

ClientSampleId :





Tgt Ion: 93 Resp: 393523

Ion Ratio Lower Upper

93 100

95 33.0 26.1 39.1

123 11.8 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 20.967 ng/ul

RT: 10.269 min Scan# 1221

Delta R.T. -0.012 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34




Tgt Ion:162 Resp: 280023


Ion Ratio Lower Upper

162 100

164 65.8 51.8 77.6

98 34.1 29.5 44.3









#30
Naphthalene
Concen: 20.881 ng/ul
RT: 10.663 min Scan# 1
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1015672
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 12.8 10.6 16.0



#32
4-Chloroaniline
Concen: 19.844 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34


Tgt Ion:127 Resp: 429396
Ion Ratio Lower Upper
127 100
129 32.1 25.3 37.9








#33
Hexachlorobutadiene
Concen: 20.640 ng/ul
RT: 10.939 min Scan# 11
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:225 Resp: 170177
Ion Ratio Lower Upper
225 100
223 62.1 50.0 75.0
227 63.8 49.9 74.9




#34
Caprolactam
Concen: 18.731 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. -0.018 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34



Tgt Ion:113 Resp: 82675
Ion Ratio Lower Upper
113 100
55 160.3 150.2 225.4
56 117.8 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 20.463 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:107 Resp: 284170
Ion Ratio Lower Upper
107 100
144 27.4 20.3 30.5
142 84.6 64.4 96.6



#36
2-Methylnaphthalene
Concen: 20.244 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:142 Resp: 644240
Ion Ratio Lower Upper
142 100
141 90.3 70.7 106.1









#37
1-Methylnaphthalene
Concen: 20.353 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:142 Resp: 652889
Ion Ratio Lower Upper
142 100
141 93.5 75.2 112.8
116 3.7 2.6 3.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.997 ng/ul
RT: 12.639 min Scan# 1624
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34


Tgt Ion:216 Resp: 311339
Ion Ratio Lower Upper
216 100
214 80.2 62.4 93.6
179 22.2 17.3 25.9
108 18.9 14.0 21.0








#40
 Hexachlorocyclopentadiene
 Concen: 15.751 ng/ul
 RT: 12.616 min Scan# 1620
 Delta R.T. -0.006 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34


Instrument :
 BNA_M
 ClientSampleId :





Tgt Ion:237 Resp: 163328
 Ion Ratio Lower Upper
 237 100
 235 61.5 48.4 72.6
 272 13.9 11.6 17.4




#41
 2,4,6-Trichlorophenol
 Concen: 20.321 ng/ul
 RT: 12.892 min Scan# 1667
 Delta R.T. -0.012 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34




Tgt Ion:196 Resp: 202465
 Ion Ratio Lower Upper
 196 100
 198 95.1 77.8 116.6
 200 31.4 25.0 37.6








#42
2,4,5-Trichlorophenol
Concen: 19.990 ng/ul
RT: 12.969 min Scan# 1
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:196 Resp: 212516
Ion Ratio Lower Upper
196 100
198 95.0 76.8 115.2
200 31.8 24.9 37.3




#43
1,1'-Biphenyl
Concen: 20.058 ng/ul
RT: 13.286 min Scan# 1734
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Tgt Ion:154 Resp: 844406
Ion Ratio Lower Upper
154 100
153 39.3 32.4 48.6
76 13.4 10.1 15.1








#44
2-Chloronaphthalene
Concen: 20.430 ng/ul
RT: 13.328 min Scan# 1741
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:162 Resp: 660045
Ion Ratio Lower Upper
162 100
164 32.3 25.4 38.0
127 39.2 31.6 47.4




#45
2-Nitroaniline
Concen: 19.866 ng/ul
RT: 13.551 min Scan# 1779
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34



Tgt Ion: 65 Resp: 176946
Ion Ratio Lower Upper
65 100
92 71.9 54.5 81.7
138 115.7 91.2 136.8









#47
Dimethylphthalate
Concen: 20.427 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:163 Resp: 765504
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.181 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:165 Resp: 150152
Ion Ratio Lower Upper
165 100
89 58.3 42.5 63.7
121 22.3 17.2 25.8









#50
Acenaphthylene
Concen: 20.588 ng/ul
RT: 14.175 min Scan# 1885
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:152 Resp: 1082593
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1
153 13.5 10.6 15.8



#51
3-Nitroaniline
Concen: 14.549 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:138 Resp: 119629
Ion Ratio Lower Upper
138 100
108 9.3 8.1 12.1
92 114.5 90.4 135.6









#52
Acenaphthene
Concen: 20.372 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:153 Resp: 694968
Ion Ratio Lower Upper
153 100
152 47.3 37.1 55.7
154 88.3 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 15.939 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:184 Resp: 67333
Ion Ratio Lower Upper
184 100
63 67.9 54.2 81.4
154 63.5 49.3 73.9









#55
4-Nitrophenol
Concen: 18.188 ng/ul
RT: 14.698 min Scan# 1977
Delta R.T. -0.018 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 106425
Ion Ratio Lower Upper
109 100
139 138.8 124.9 187.3
65 128.2 122.2 183.4



#56
Dibenzofuran
Concen: 20.675 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:168 Resp: 975889
Ion Ratio Lower Upper
168 100
139 37.7 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 20.839 ng/ul
RT: 14.827 min Scan# 1996
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:165 Resp: 218553
Ion Ratio Lower Upper
165 100
63 46.1 35.9 53.9
182 2.4 2.5 3.7#



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.319 ng/ul
RT: 15.080 min Scan# 2039
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:232 Resp: 169719
Ion Ratio Lower Upper
232 100
131 54.1 36.7 55.1
130 2.8 1.7 2.5#
166 33.2 24.4 36.6









#59
Diethylphthalate
Concen: 20.094 ng/ul
RT: 15.280 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 761689
Ion Ratio Lower Upper
149 100
177 21.3 16.8 25.2
150 11.7 10.2 15.4



#61
Fluorene
Concen: 20.300 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:166 Resp: 782731
Ion Ratio Lower Upper
166 100
165 98.3 77.2 115.8
167 13.2 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.800 ng/ul
RT: 15.498 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 365791
Ion Ratio Lower Upper
204 100
206 34.0 25.9 38.9
141 58.0 50.6 75.8



#63
4-Nitroaniline
Concen: 12.386 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:138 Resp: 93758
Ion Ratio Lower Upper
138 100
92 49.1 39.2 58.8
108 73.1 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 19.266 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:198 Resp: 114074
Ion Ratio Lower Upper
198 100
182 5.3 4.2 6.4
77 27.3 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 19.972 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34


Tgt Ion:169 Resp: 646347
Ion Ratio Lower Upper
169 100
168 66.8 53.8 80.6
167 36.9 28.7 43.1








#68
4-Bromophenyl-phenylether
Concen: 19.428 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:248 Resp: 209666
Ion Ratio Lower Upper
248 100
250 99.7 77.3 115.9
141 79.8 59.8 89.8



#69
Hexachlorobenzene
Concen: 19.790 ng/ul
RT: 16.492 min Scan# 2279
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:284 Resp: 260218
Ion Ratio Lower Upper
284 100
142 39.7 28.6 42.8
249 31.4 24.6 37.0





#70

Atrazine

Concen: 19.190 ng/ul

RT: 16.668 min Scan# 21

Delta R.T. -0.006 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:200 Resp: 215400

Ion Ratio Lower Upper

200 100

173 26.3 21.3 31.9

215 47.2 39.0 58.6



#71

Pentachlorophenol

Concen: 16.159 ng/ul

RT: 16.851 min Scan# 2340

Delta R.T. -0.012 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34




Tgt Ion:266 Resp: 129801


Ion Ratio Lower Upper

266 100

264 63.9 49.5 74.3

268 65.8 51.0 76.4









#72
Phenanthrene
Concen: 20.559 ng/u1
RT: 17.239 min Scan# 2406
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:178 Resp: 1213001
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 19.2 14.9 22.3



#74
Anthracene
Concen: 20.777 ng/u1
RT: 17.327 min Scan# 2421
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:178 Resp: 1230082
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.3 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.876 ng/uL

RT: 13.251 min Scan# 1

Delta R.T. -0.006 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:216 Resp: 323135

Ion Ratio Lower Upper

216 100

214 79.8 60.9 91.3

218 48.2 37.5 56.3



#76

Pentachlorobenzene

Concen: 19.941 ng/uL

RT: 14.769 min Scan# 1986

Delta R.T. -0.006 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34




Tgt Ion:250 Resp: 309438


Ion Ratio Lower Upper

250 100

248 62.1 50.4 75.6

252 64.8 51.1 76.7





#77

Carbazole

Concen: 20.795 ng/ul

RT: 17.604 min Scan# 24

Delta R.T. -0.012 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:167 Resp: 1137099

Ion Ratio Lower Upper

167 100

166 21.5 17.4 26.0

139 13.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 20.108 ng/ul

RT: 18.168 min Scan# 2564

Delta R.T. -0.006 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34




Tgt Ion:149 Resp: 1240553


Ion Ratio Lower Upper

149 100

150 9.1 7.3 10.9

104 4.9 3.7 5.5









#80
 Fluoranthene
 Concen: 18.754 ng/u1
 RT: 19.251 min Scan# 21
 Delta R.T. -0.012 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:202 Resp: 1380818
 Ion Ratio Lower Upper
 202 100
 101 12.0 10.1 15.1
 100 9.2 7.8 11.6



#82
 Pyrene
 Concen: 19.128 ng/u1
 RT: 19.615 min Scan# 2810
 Delta R.T. -0.006 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34

Tgt Ion:202 Resp: 1451630
 Ion Ratio Lower Upper
 202 100
 101 13.6 11.9 17.9
 100 11.5 10.2 15.4









#83
 Butylbenzylphthalate
 Concen: 19.927 ng/ul
 RT: 20.515 min Scan# 2963
 Delta R.T. -0.012 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:149 Resp: 550028
 Ion Ratio Lower Upper
 149 100
 91 70.8 53.2 79.8
 206 17.6 16.0 24.0



#84
 3,3'-Dichlorobenzidine
 Concen: 21.224 ng/ul
 RT: 21.298 min Scan# 3096
 Delta R.T. -0.012 min
 Lab File: BMCCC258.D
 Acq: 13 Aug 2022 14:34

Tgt Ion:252 Resp: 435499
 Ion Ratio Lower Upper
 252 100
 254 64.5 52.4 78.6
 126 11.2 10.8 16.2









#85
Benzo(a)anthracene
Concen: 20.080 ng/ul
RT: 21.362 min Scan# 3108
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1405300
Ion Ratio Lower Upper
228 100
229 19.5 15.6 23.4
226 26.8 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.264 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:149 Resp: 768283
Ion Ratio Lower Upper
149 100
167 26.4 21.5 32.3
279 3.7 3.4 5.2





#87

Chrysene

Concen: 19.999 ng/ul

RT: 21.415 min Scan# 3117

Delta R.T. -0.006 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 1365655

Ion Ratio Lower Upper

228 100

226 30.3 23.7 35.5

229 19.3 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 17.705 ng/ul




RT: 22.192 min Scan# 3248


Delta R.T. -0.012 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Tgt Ion:149 Resp: 1312182









#90
Benzo(b)fluoranthene
Concen: 19.377 ng/ul
RT: 22.997 min Scan# 31
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1484127
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 11.1 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 18.869 ng/ul
RT: 23.045 min Scan# 3393
Delta R.T. -0.012 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:252 Resp: 1371414
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 10.7 10.6 15.8









#93
Benzo(a)pyrene
Concen: 19.853 ng/ul
RT: 23.603 min Scan# 3491
Delta R.T. -0.018 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1448342
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 12.7 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.707 ng/ul
RT: 26.127 min Scan# 3917
Delta R.T. -0.023 min
Lab File: BMCCC258.D
Acq: 13 Aug 2022 14:34

Tgt Ion:276 Resp: 1687950
Ion Ratio Lower Upper
276 100
138 26.3 25.6 38.4
277 24.8 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 21.797 ng/ul

RT: 26.144 min Scan# 3923

Delta R.T. -0.018 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:278 Resp: 1454161

Ion Ratio Lower Upper

278 100

139 17.2 16.2 24.2

279 23.5 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 22.202 ng/ul

RT: 26.868 min Scan# 4043

Delta R.T. -0.029 min

Lab File: BMCCC258.D

Acq: 13 Aug 2022 14:34

Tgt Ion:276 Resp: 1430706

Ion Ratio Lower Upper

276 100

138 23.6 21.3 31.9

277 23.6 18.6 28.0

