

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
 Data File : BM036234.D
 Acq On : 12 Aug 2022 17:32
 Operator : CG/JU
 Sample : N4095-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GBJM8

Quant Time: Aug 12 20:19:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 11:23:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	179260	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	721162	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	401191	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	801652	20.000	ng/u1	0.00
79) Chrysene-d12	21.386	240	815020	20.000	ng/u1	0.00
88) Perylene-d12	23.721	264	913652	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	21594	4.533	ng/uL	0.00
4) Pyridine-d5	3.657	84	67716	5.189	ng/u1	0.01
7) Phenol-d5	7.010	99	84633	5.630	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.151	67	262679	26.212	ng/u1	0.00
11) 2-Chlorophenol-d4	7.351	132	270598	21.686	ng/u1	0.00
15) 4-Methylphenol-d8	8.551	113	153561	12.609	ng/u1	0.00
21) Nitrobenzene-d5	8.986	128	167927	29.386	ng/u1	0.00
24) 2-Nitrophenol-d4	9.710	143	162964	29.270	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	256803	25.503	ng/u1	0.00
31) 4-Chloroaniline-d4	10.775	131	328289	19.258	ng/u1	0.00
46) Dimethylphthalate-d6	13.880	166	848520	31.049	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1068373	30.640	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	21157	3.770	ng/u1	0.02
60) Fluorene-d10	15.451	176	781294	31.638	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.580	200	115442	26.790	ng/u1	0.00
73) Anthracene-d10	17.304	188	1302964	35.951	ng/u1	0.00
81) Pyrene-d10	19.592	212	1508546	33.397	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.574	264	1687241	36.375	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
Data File : BM036234.D
Acq On : 12 Aug 2022 17:32
Operator : CG/JU
Sample : N4095-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GBJM8

Quant Time: Aug 12 20:19:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 12 11:23:11 2022
Response via : Initial Calibration

