

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
 Data File : BM036239.D
 Acq On : 12 Aug 2022 20:35
 Operator : CG/JU
 Sample : N4069-16
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GBJM0

Quant Time: Aug 12 21:05:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 11:23:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	159130	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	644395	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	363488	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.203	188	754285	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	850272	20.000	ng/u1	0.00
88) Perylene-d12	23.721	264	943612	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	19699	4.658	ng/uL	0.00
4) Pyridine-d5	3.657	84	62533	5.398	ng/u1	0.01
7) Phenol-d5	7.004	99	90288	6.766	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.151	67	255775	28.752	ng/u1	0.00
11) 2-Chlorophenol-d4	7.351	132	266707	24.078	ng/u1	0.00
15) 4-Methylphenol-d8	8.545	113	160347	14.832	ng/u1	0.00
21) Nitrobenzene-d5	8.986	128	164399	32.196	ng/u1	0.00
24) 2-Nitrophenol-d4	9.704	143	158741	31.909	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	259083	28.794	ng/u1	0.00
31) 4-Chloroaniline-d4	10.775	131	247939	16.277	ng/u1	0.00
46) Dimethylphthalate-d6	13.880	166	881503	35.602	ng/u1	0.00
49) Acenaphthylene-d8	14.151	160	1076516	34.076	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	27944	5.496	ng/u1	0.01
60) Fluorene-d10	15.451	176	795639	35.561	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.580	200	125879	31.047	ng/u1	0.00
73) Anthracene-d10	17.298	188	1390007	40.761	ng/u1	0.00
81) Pyrene-d10	19.592	212	1758585	37.318	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.568	264	1966015	41.039	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
Data File : BM036239.D
Acq On : 12 Aug 2022 20:35
Operator : CG/JU
Sample : N4069-16
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GBJM0

Quant Time: Aug 12 21:05:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 12 11:23:11 2022
Response via : Initial Calibration

