

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
 Data File : BM036253.D
 Acq On : 13 Aug 2022 05:34
 Operator : CG/JU
 Sample : N4165-10
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGBY7

Quant Time: Aug 13 06:04:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 11:23:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.810	152	163915	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	675373	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	371915	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.198	188	781299	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	862376	20.000	ng/u1	0.00
88) Perylene-d12	23.715	264	991310	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	19030	4.369	ng/uL	0.00
4) Pyridine-d5	3.646	84	101370	8.494	ng/u1	0.00
7) Phenol-d5	6.998	99	81155	5.904	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.146	67	261561	28.544	ng/u1	0.00
11) 2-Chlorophenol-d4	7.346	132	256811	22.508	ng/u1	0.00
15) 4-Methylphenol-d8	8.540	113	141659	12.721	ng/u1	-0.01
21) Nitrobenzene-d5	8.981	128	161771	30.228	ng/u1	0.00
24) 2-Nitrophenol-d4	9.704	143	153308	29.403	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.245	165	242175	25.681	ng/u1	0.00
31) 4-Chloroaniline-d4	10.763	131	347477	21.766	ng/u1	0.00
46) Dimethylphthalate-d6	13.875	166	809687	31.960	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	973889	30.129	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.710	143	22818	4.386	ng/u1	0.01
60) Fluorene-d10	15.445	176	758120	33.116	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.574	200	109141	25.988	ng/u1	0.00
73) Anthracene-d10	17.292	188	1296881	36.715	ng/u1	-0.01
81) Pyrene-d10	19.586	212	1610079	33.687	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.562	264	1854565	36.850	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
Data File : BM036253.D
Acq On : 13 Aug 2022 05:34
Operator : CG/JU
Sample : N4165-10
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGBY7

Quant Time: Aug 13 06:04:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 12 11:23:11 2022
Response via : Initial Calibration

