

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
 Data File : BM036236.D
 Acq On : 12 Aug 2022 18:46
 Operator : CG/JU
 Sample : N4069-22
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 GBJM6

Quant Time: Aug 12 20:22:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 11:23:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	186485	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	765952	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	432447	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	856228	20.000	ng/u1	0.00
79) Chrysene-d12	21.386	240	843095	20.000	ng/u1	# 0.00
88) Perylene-d12	23.721	264	931180	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	20374	4.111	ng/uL	0.00
4) Pyridine-d5	3.657	84	47853	3.525	ng/u1	0.01
7) Phenol-d5	7.010	99	117376	7.506	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.151	67	279467	26.807	ng/u1	0.00
11) 2-Chlorophenol-d4	7.351	132	307620	23.698	ng/u1	0.00
15) 4-Methylphenol-d8	8.545	113	207004	16.339	ng/u1	0.00
21) Nitrobenzene-d5	8.987	128	189708	31.256	ng/u1	0.00
24) 2-Nitrophenol-d4	9.710	143	188230	31.832	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	305608	28.575	ng/u1	0.00
31) 4-Chloroaniline-d4	10.775	131	183198	10.118	ng/u1	0.00
46) Dimethylphthalate-d6	13.880	166	1030457	34.981	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1232482	32.792	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	31962	5.284	ng/u1	0.01
60) Fluorene-d10	15.451	176	925176	34.756	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.580	200	140828	30.598	ng/u1	0.00
73) Anthracene-d10	17.304	188	1463092	37.796	ng/u1	0.00
81) Pyrene-d10	19.592	212	1730204	37.029	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.574	264	1774682	37.540	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081222\
Data File : BM036236.D
Acq On : 12 Aug 2022 18:46
Operator : CG/JU
Sample : N4069-22
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
GBJM6

Quant Time: Aug 12 20:22:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 12 11:23:11 2022
Response via : Initial Calibration

