

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081621\
 Data File : BM031753.D
 Acq On : 16 Aug 2021 16:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081621

Quant Time: Aug 16 17:13:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 16:42:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	39.199	2.0	102	0.00
3	Pyridine	40.000	41.798	-4.5	105	0.00
4	n-Nitrosodimethylamine	40.000	42.162	-5.4	106	0.00
5 S	2-Fluorophenol	80.000	82.720	-3.4	105	0.00
6	Aniline	40.000	41.578	-3.9	104	0.00
7 S	Phenol-d6	80.000	84.790	-6.0	104	0.00
8	2-Chlorophenol	40.000	41.672	-4.2	104	0.00
9	Benzaldehyde	40.000	39.761	0.6	103	0.00
10 C	Phenol	40.000	42.254	-5.6	105	0.00
11	bis(2-Chloroethyl)ether	40.000	40.308	-0.8	101	0.00
12	1,3-Dichlorobenzene	40.000	41.697	-4.2	104	0.00
13 C	1,4-Dichlorobenzene	40.000	41.017	-2.5	104	0.00
14	1,2-Dichlorobenzene	40.000	41.784	-4.5	108	0.00
15	Benzyl Alcohol	40.000	42.082	-5.2	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.420	-1.1	101	0.00
17	2-Methylphenol	40.000	42.027	-5.1	102	0.00
18	Hexachloroethane	40.000	41.411	-3.5	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.398	-3.5	103	0.00
20	3+4-Methylphenols	40.000	42.276	-5.7	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	41.334	-3.3	105	0.00
23 S	Nitrobenzene-d5	80.000	82.878	-3.6	104	0.00
24	Nitrobenzene	40.000	40.496	-1.2	103	0.00
25	Isophorone	40.000	41.585	-4.0	104	0.00
26 C	2-Nitrophenol	40.000	43.198	-8.0	107	0.00
27	2,4-Dimethylphenol	40.000	41.853	-4.6	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.814	-4.5	105	0.00
29 C	2,4-Dichlorophenol	40.000	42.587	-6.5	104	0.00
30	1,2,4-Trichlorobenzene	40.000	41.350	-3.4	105	0.00
31	Naphthalene	40.000	41.765	-4.4	105	0.00
32	Benzoic acid	40.000	42.907	-7.3	105	0.00
33	4-Chloroaniline	40.000	42.172	-5.4	105	0.00
34 C	Hexachlorobutadiene	40.000	41.360	-3.4	103	0.00
35	Caprolactam	40.000	43.596	-9.0	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.366	-5.9	104	0.00
37	2-Methylnaphthalene	40.000	41.973	-4.9	104	0.00
38	1-Methylnaphthalene	40.000	41.625	-4.1	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.143	-5.4	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.996	-10.0	103	0.00
42 S	2,4,6-Tribromophenol	80.000	85.421	-6.8	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.992	-7.5	105	0.00
44	2,4,5-Trichlorophenol	40.000	43.221	-8.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	83.596	-4.5	104	0.00
46	1,1'-Biphenyl	40.000	41.584	-4.0	104	0.00
47	2-Chloronaphthalene	40.000	42.168	-5.4	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081621\
 Data File : BM031753.D
 Acq On : 16 Aug 2021 16:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM081621

Quant Time: Aug 16 17:13:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 16:42:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.841	-7.1	103	0.00
49	Acenaphthylene	40.000	42.261	-5.7	105	0.00
50	Dimethylphthalate	40.000	41.601	-4.0	104	0.00
51	2,6-Dinitrotoluene	40.000	42.395	-6.0	105	0.00
52 C	Acenaphthene	40.000	41.857	-4.6	104	0.00
53	3-Nitroaniline	40.000	43.190	-8.0	105	0.00
54 P	2,4-Dinitrophenol	40.000	41.689	-4.2	100	0.00
55	Dibenzofuran	40.000	42.064	-5.2	105	0.00
56 P	4-Nitrophenol	40.000	44.249	-10.6	105	0.00
57	2,4-Dinitrotoluene	40.000	43.431	-8.6	105	0.00
58	Fluorene	40.000	42.144	-5.4	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.731	-6.8	106	0.00
60	Diethylphthalate	40.000	42.127	-5.3	104	0.00
61	4-Chlorophenyl-phenylether	40.000	41.874	-4.7	105	0.00
62	4-Nitroaniline	40.000	43.066	-7.7	103	0.00
63	Azobenzene	40.000	41.985	-5.0	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.465	-8.7	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.237	-5.6	104	0.00
67	4-Bromophenyl-phenylether	40.000	41.960	-4.9	103	0.00
68	Hexachlorobenzene	40.000	41.630	-4.1	103	0.00
69	Atrazine	40.000	41.995	-5.0	103	0.00
70 C	Pentachlorophenol	40.000	44.437	-11.1	107	0.00
71	Phenanthrene	40.000	42.134	-5.3	104	0.00
72	Anthracene	40.000	42.137	-5.3	104	0.00
73	Carbazole	40.000	41.947	-4.9	102	0.00
74	Di-n-butylphthalate	40.000	42.998	-7.5	105	0.00
75 C	Fluoranthene	40.000	42.380	-6.0	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	45.964	-14.9	119	0.00
78	Pyrene	40.000	43.048	-7.6	105	0.00
79 S	Terphenyl-d14	80.000	85.896	-7.4	104	0.00
80	Butylbenzylphthalate	40.000	43.361	-8.4	104	0.00
81	Benzo(a)anthracene	40.000	42.498	-6.2	104	0.00
82	3,3'-Dichlorobenzidine	40.000	42.441	-6.1	102	0.00
83	Chrysene	40.000	41.736	-4.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.878	-9.7	105	0.00
85 c	Di-n-octyl phthalate	40.000	42.994	-7.5	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.155	-5.4	103	0.00
88	Benzo(b)fluoranthene	40.000	44.179	-10.4	106	0.00
89	Benzo(k)fluoranthene	40.000	40.924	-2.3	100	0.00
90 C	Benzo(a)pyrene	40.000	42.799	-7.0	103	0.00
91	Dibenzo(a,h)anthracene	40.000	42.060	-5.2	103	0.00
92	Benzo(g,h,i)perylene	40.000	41.774	-4.4	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081621\
Data File : BM031753.D
Acq On : 16 Aug 2021 16:32
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM081621

Quant Time: Aug 16 17:13:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 16:42:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0