

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081621\
 Data File : BM031756.D
 Acq On : 16 Aug 2021 18:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 16 19:17:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 16:42:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.521	0.537	-3.1	86	0.00
3	Pyridine	1.440	1.482	-2.9	83	0.00
4	n-Nitrosodimethylamine	0.840	0.951	-13.2	91	0.00
5 S	2-Fluorophenol	1.246	1.300	-4.3	85	0.00
6	Aniline	1.974	2.050	-3.9	83	0.00
7 S	Phenol-d6	1.806	1.862	-3.1	81	0.00
8	2-Chlorophenol	1.459	1.524	-4.5	84	0.00
9	Benzaldehyde	1.282	1.271	0.9	82	0.00
10 C	Phenol	1.794	1.869	-4.2	83	0.00
11	bis(2-Chloroethyl)ether	1.347	1.329	1.3	79	0.00
12	1,3-Dichlorobenzene	1.561	1.632	-4.5	84	0.00
13 C	1,4-Dichlorobenzene	1.608	1.705	-6.0	87	0.00
14	1,2-Dichlorobenzene	1.580	1.664	-5.3	87	0.00
15	Benzyl Alcohol	1.555	1.674	-7.7	85	0.00
16	2,2'-oxybis(1-Chloropropane	1.925	1.966	-2.1	82	-0.01
17	2-Methylphenol	1.268	1.339	-5.6	82	0.00
18	Hexachloroethane	0.667	0.728	-9.1	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.426	1.505	-5.5	84	0.00
20	3+4-Methylphenols	1.783	1.888	-5.9	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22	Acetophenone	0.568	0.574	-1.1	85	0.00
23 S	Nitrobenzene-d5	0.478	0.489	-2.3	85	0.00
24	Nitrobenzene	0.489	0.494	-1.0	85	0.00
25	Isophorone	0.857	0.867	-1.2	84	0.00
26 C	2-Nitrophenol	0.192	0.197	-2.6	84	0.00
27	2,4-Dimethylphenol	0.294	0.303	-3.1	86	0.00
28	bis(2-Chloroethoxy)methane	0.412	0.422	-2.4	85	0.00
29 C	2,4-Dichlorophenol	0.334	0.350	-4.8	85	0.00
30	1,2,4-Trichlorobenzene	0.357	0.376	-5.3	89	0.00
31	Naphthalene	1.145	1.186	-3.6	87	0.00
32	Benzoic acid	0.263	0.269	-2.3	83	0.00
33	4-Chloroaniline	0.476	0.500	-5.0	86	0.00
34 C	Hexachlorobutadiene	0.228	0.241	-5.7	88	0.00
35	Caprolactam	0.116	0.125	-7.8	85	0.00
36 C	4-Chloro-3-methylphenol	0.415	0.436	-5.1	85	0.00
37	2-Methylnaphthalene	0.857	0.898	-4.8	87	0.00
38	1-Methylnaphthalene	0.818	0.862	-5.4	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.568	0.583	-2.6	89	0.00
41 P	Hexachlorocyclopentadiene	0.284	0.299	-5.3	86	0.00
42 S	2,4,6-Tribromophenol	0.246	0.269	-9.3	93	0.00
43 C	2,4,6-Trichlorophenol	0.419	0.438	-4.5	89	0.00
44	2,4,5-Trichlorophenol	0.458	0.470	-2.6	87	0.00
45 S	2-Fluorobiphenyl	1.503	1.536	-2.2	89	0.00
46	1,1'-Biphenyl	1.574	1.598	-1.5	89	0.00
47	2-Chloronaphthalene	1.236	1.260	-1.9	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081621\
 Data File : BM031756.D
 Acq On : 16 Aug 2021 18:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 16 19:17:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 16:42:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.459	0.483	-5.2	89	0.00
49	Acenaphthylene	2.023	2.115	-4.5	91	0.00
50	Dimethylphthalate	1.667	1.695	-1.7	89	0.00
51	2,6-Dinitrotoluene	0.370	0.386	-4.3	90	0.00
52 C	Acenaphthene	1.232	1.290	-4.7	91	0.00
53	3-Nitroaniline	0.380	0.403	-6.1	90	0.00
54 P	2,4-Dinitrophenol	0.218	0.217	0.5	83	0.00
55	Dibenzofuran	2.059	2.167	-5.2	92	0.00
56 P	4-Nitrophenol	0.324	0.352	-8.6	90	0.00
57	2,4-Dinitrotoluene	0.533	0.571	-7.1	91	0.00
58	Fluorene	1.710	1.801	-5.3	91	0.00
59	2,3,4,6-Tetrachlorophenol	0.415	0.439	-5.8	92	0.00
60	Diethylphthalate	1.832	1.925	-5.1	91	0.00
61	4-Chlorophenyl-phenylether	0.825	0.880	-6.7	93	0.00
62	4-Nitroaniline	0.390	0.433	-11.0	93	0.00
63	Azobenzene	2.028	2.191	-8.0	92	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	0.137	0.140	-2.2	88	0.00
66 c	n-Nitrosodiphenylamine	0.658	0.675	-2.6	92	0.00
67	4-Bromophenyl-phenylether	0.213	0.221	-3.8	92	0.00
68	Hexachlorobenzene	0.233	0.240	-3.0	92	0.00
69	Atrazine	0.234	0.244	-4.3	93	0.00
70 C	Pentachlorophenol	0.156	0.167	-7.1	93	0.00
71	Phenanthrene	1.220	1.272	-4.3	93	0.00
72	Anthracene	1.196	1.251	-4.6	94	0.00
73	Carbazole	1.198	1.269	-5.9	94	0.00
74	Di-n-butylphthalate	1.492	1.590	-6.6	94	0.00
75 C	Fluoranthene	1.461	1.536	-5.1	94	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
77	Benzydine	0.594	0.719	-21.0	114	0.00
78	Pyrene	1.449	1.543	-6.5	95	0.00
79 S	Terphenyl-d14	1.187	1.275	-7.4	95	0.00
80	Butylbenzylphthalate	0.694	0.751	-8.2	95	0.00
81	Benzo(a)anthracene	1.465	1.562	-6.6	95	0.00
82	3,3'-Dichlorobenzidine	0.453	0.489	-7.9	94	0.00
83	Chrysene	1.428	1.485	-4.0	93	0.00
84	Bis(2-ethylhexyl)phthalate	1.073	1.163	-8.4	95	0.00
85 c	Di-n-octyl phthalate	1.773	1.906	-7.5	96	0.00
86 I	Perylene-d12	1.000	1.000	0.0	95	0.00
87	Indeno(1,2,3-cd)pyrene	1.409	1.475	-4.7	95	0.00
88	Benzo(b)fluoranthene	1.502	1.619	-7.8	96	0.00
89	Benzo(k)fluoranthene	1.449	1.513	-4.4	95	0.00
90 C	Benzo(a)pyrene	1.330	1.419	-6.7	95	0.00
91	Dibenzo(a,h)anthracene	1.240	1.293	-4.3	95	0.00
92	Benzo(g,h,i)perylene	1.129	1.175	-4.1	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081621\
Data File : BM031756.D
Acq On : 16 Aug 2021 18:24
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Aug 16 19:17:41 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 16:42:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0