




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081721\
Data File : BM031783.D
Acq On : 17 Aug 2021 16:04
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 17:07:46 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 16:42:38 2021
Response via : Initial Calibration






TIC: BM031783.D\data.ms

(77) Benzidine**19.209min (0.000) 131955.97 ng****response 584281**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.60	13.45
183.00	12.30	11.94
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081721\
Data File : BM031783.D
Acq On : 17 Aug 2021 16:04
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 17:07:46 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 16:42:38 2021
Response via : Initial Calibration



TIC: BM031783.D\data.ms

(70) Pentachlorophenol (C)

16.592min (-0.006) 0.00 ng

response 99464

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.00	62.21
264.00	64.90	62.79
0.00	0.00	0.00