

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081721\
 Data File : BM031784.D
 Acq On : 17 Aug 2021 16:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 17 18:30:22 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 16:42:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	40.794	-2.0	107	0.00
3	Pyridine	40.000	40.560	-1.4	103	0.00
4	n-Nitrosodimethylamine	40.000	41.773	-4.4	106	0.00
5 S	2-Fluorophenol	80.000	84.181	-5.2	108	0.00
6	Aniline	40.000	39.501	1.2	100	0.00
7 S	Phenol-d6	80.000	83.131	-3.9	103	0.00
8	2-Chlorophenol	40.000	40.882	-2.2	103	0.00
9	Benzaldehyde	40.000	39.684	0.8	104	0.00
10 C	Phenol	40.000	41.684	-4.2	104	0.00
11	bis(2-Chloroethyl)ether	40.000	39.790	0.5	101	0.00
12	1,3-Dichlorobenzene	40.000	41.074	-2.7	103	0.00
13 C	1,4-Dichlorobenzene	40.000	40.949	-2.4	105	0.00
14	1,2-Dichlorobenzene	40.000	40.556	-1.4	106	0.00
15	Benzyl Alcohol	40.000	40.565	-1.4	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.956	5.1	96	0.00
17	2-Methylphenol	40.000	40.608	-1.5	100	0.00
18	Hexachloroethane	40.000	40.211	-0.5	100	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	38.211	4.5	96	0.00
20	3+4-Methylphenols	40.000	40.441	-1.1	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	41.816	-4.5	100	0.00
23 S	Nitrobenzene-d5	80.000	83.600	-4.5	99	0.00
24	Nitrobenzene	40.000	41.073	-2.7	98	0.00
25	Isophorone	40.000	40.404	-1.0	95	0.00
26 C	2-Nitrophenol	40.000	43.968	-9.9	102	0.00
27	2,4-Dimethylphenol	40.000	41.711	-4.3	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.435	-3.6	98	0.00
29 C	2,4-Dichlorophenol	40.000	43.591	-9.0	101	0.00
30	1,2,4-Trichlorobenzene	40.000	42.599	-6.5	102	0.00
31	Naphthalene	40.000	41.970	-4.9	100	0.00
32	Benzoic acid	40.000	40.684	-1.7	94	0.00
33	4-Chloroaniline	40.000	41.641	-4.1	97	-0.01
34 C	Hexachlorobutadiene	40.000	42.127	-5.3	99	0.00
35	Caprolactam	40.000	43.646	-9.1	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.263	-5.7	97	0.00
37	2-Methylnaphthalene	40.000	41.588	-4.0	97	0.00
38	1-Methylnaphthalene	40.000	41.436	-3.6	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.657	-6.6	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.177	7.1	82	0.00
42 S	2,4,6-Tribromophenol	80.000	85.394	-6.7	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.428	-8.6	100	0.00
44	2,4,5-Trichlorophenol	40.000	42.875	-7.2	98	0.00
45 S	2-Fluorobiphenyl	80.000	83.790	-4.7	98	0.00
46	1,1'-Biphenyl	40.000	41.476	-3.7	98	0.00
47	2-Chloronaphthalene	40.000	41.967	-4.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081721\
 Data File : BM031784.D
 Acq On : 17 Aug 2021 16:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 17 18:30:22 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 16:42:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.019	-5.0	95	0.00
49	Acenaphthylene	40.000	41.498	-3.7	97	0.00
50	Dimethylphthalate	40.000	40.590	-1.5	95	0.00
51	2,6-Dinitrotoluene	40.000	40.963	-2.4	95	0.00
52 C	Acenaphthene	40.000	41.703	-4.3	97	0.00
53	3-Nitroaniline	40.000	42.362	-5.9	97	0.00
54 P	2,4-Dinitrophenol	40.000	30.911	22.7	70	0.00
55	Dibenzofuran	40.000	41.731	-4.3	98	0.00
56 P	4-Nitrophenol	40.000	42.098	-5.2	94	0.00
57	2,4-Dinitrotoluene	40.000	42.527	-6.3	97	0.00
58	Fluorene	40.000	41.233	-3.1	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.429	-6.1	99	0.00
60	Diethylphthalate	40.000	40.337	-0.8	94	0.00
61	4-Chlorophenyl-phenylether	40.000	40.972	-2.4	96	0.00
62	4-Nitroaniline	40.000	41.516	-3.8	94	0.00
63	Azobenzene	40.000	41.026	-2.6	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	40.000	34.126	14.7	75	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.909	-4.8	96	0.00
67	4-Bromophenyl-phenylether	40.000	41.619	-4.0	95	0.00
68	Hexachlorobenzene	40.000	41.824	-4.6	96	0.00
69	Atrazine	40.000	42.111	-5.3	96	0.00
70 C	Pentachlorophenol	40.000	42.235	-5.6	94	0.00
71	Phenanthrene	40.000	41.783	-4.5	96	0.00
72	Anthracene	40.000	42.147	-5.4	97	0.00
73	Carbazole	40.000	42.635	-6.6	97	0.00
74	Di-n-butylphthalate	40.000	41.508	-3.8	94	0.00
75 C	Fluoranthene	40.000	41.930	-4.8	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	33.651	15.9	80	0.00
78	Pyrene	40.000	43.024	-7.6	96	0.00
79 S	Terphenyl-d14	80.000	85.829	-7.3	95	0.00
80	Butylbenzylphthalate	40.000	43.757	-9.4	96	0.00
81	Benzo(a)anthracene	40.000	42.289	-5.7	95	0.00
82	3,3'-Dichlorobenzidine	40.000	41.236	-3.1	90	0.00
83	Chrysene	40.000	41.917	-4.8	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.528	-6.3	93	0.00
85 c	Di-n-octyl phthalate	40.000	42.782	-7.0	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.034	-2.6	91	0.00
88	Benzo(b)fluoranthene	40.000	43.712	-9.3	95	-0.01
89	Benzo(k)fluoranthene	40.000	41.319	-3.3	92	-0.01
90 C	Benzo(a)pyrene	40.000	42.303	-5.8	93	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.608	-1.5	91	-0.01
92	Benzo(g,h,i)perylene	40.000	40.909	-2.3	93	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081721\
Data File : BM031784.D
Acq On : 17 Aug 2021 16:41
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Aug 17 18:30:22 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 16:42:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0