

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090120\
 Data File : BM027290.D
 Acq On : 01 Sep 2020 10:25
 Operator : JU/CG
 Sample : PB131292BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK92

Quant Time: Sep 01 11:13:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM082520MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 01 11:00:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	93879	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	354361	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.23	164	239170	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.97	188	527576	20.00	ng/ul	0.00
78) Chrysene-d12	21.17	240	487123	20.00	ng/ul	0.00
86) Perylene-d12	23.35	264	449664	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.16	96	12734	7.68	ng/uL	0.00
5) Phenol-d5	6.77	99	196548	25.82	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.93	67	137213	27.86	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	159836	27.26	ng/ul	0.00
13) 4-Methylphenol-d8	8.30	113	157673	25.80	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	76700	27.01	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	84473	25.80	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.99	165	168034	25.16	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	212255	29.49	ng/ul	0.00
44) Dimethylphthalate-d6	13.64	166	551693	29.07	ng/ul	0.00
47) Acenaphthylene-d8	13.92	160	649955	28.97	ng/ul	0.00
52) 4-Nitrophenol-d4	14.43	143	66069	20.64	ng/ul	0.00
58) Fluorene-d10	15.23	176	468315	28.59	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.35	200	64191	18.65	ng/ul	0.00
71) Anthracene-d10	17.07	188	770806	30.78	ng/ul	0.00
79) Pyrene-d10	19.37	212	843684	30.03	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.21	264	696978	27.80	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM090120\
Data File : BM027290.D
Acq On : 01 Sep 2020 10:25
Operator : JU/CG
Sample : PB131292BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK92

Quant Time: Sep 01 11:13:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM082520MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 01 11:00:31 2020
Response via : Initial Calibration

