

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102922\
 Data File : BM037271.D
 Acq On : 29 Oct 2022 15:16
 Operator : CG/JU
 Sample : N5236-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 LSP-GW-03

Quant Time: Oct 31 02:06:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 31 01:34:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	481809	20.000	ng	0.00
21) Naphthalene-d8	10.674	136	1832655	20.000	ng	0.00
39) Acenaphthene-d10	14.510	164	996263	20.000	ng	0.00
64) Phenanthrene-d10	17.245	188	1778938	20.000	ng	0.00
76) Chrysene-d12	21.421	240	1509350	20.000	ng	0.00
86) Perylene-d12	23.780	264	1713976	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.445	112	2421039	86.812	ng	0.00
7) Phenol-d6	7.045	99	2013936	53.208	ng	0.00
23) Nitrobenzene-d5	9.045	82	3905251	107.511	ng	0.00
42) 2,4,6-Tribromophenol	15.998	330	2082708	177.393	ng	0.00
45) 2-Fluorobiphenyl	13.133	172	8289063	110.549	ng	0.00
79) Terphenyl-d14	19.868	244	10091192	119.894	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102922\
Data File : BM037271.D
Acq On : 29 Oct 2022 15:16
Operator : CG/JU
Sample : N5236-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
LSP-GW-03

Quant Time: Oct 31 02:06:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 31 01:34:15 2022
Response via : Initial Calibration

