

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110222\
 Data File : BM037310.D
 Acq On : 02 Nov 2022 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020194

Quant Time: Nov 02 22:52:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	296771	20.000	ng/ul	0.00
20) Naphthalene-d8	10.657	136	1287444	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.492	164	784259	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.227	188	1505126	20.000	ng/ul	0.00
79) Chrysene-d12	21.403	240	1284463	20.000	ng/ul	0.00
88) Perylene-d12	23.750	264	1159639	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	61100	9.076	ng/uL	0.00
4) Pyridine-d5	3.699	84	436092	21.700	ng/ul	0.00
7) Phenol-d5	7.028	99	495294	19.177	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	326278	20.587	ng/ul	0.00
11) 2-Chlorophenol-d4	7.387	132	410529	20.804	ng/ul	0.00
15) 4-Methylphenol-d8	8.575	113	399904	18.962	ng/ul	0.00
21) Nitrobenzene-d5	9.022	128	197786	20.566	ng/ul	0.00
24) 2-Nitrophenol-d4	9.745	143	211790	20.217	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	399788	20.700	ng/ul	0.00
31) 4-Chloroaniline-d4	10.804	131	583199	19.760	ng/ul	0.00
46) Dimethylphthalate-d6	13.910	166	1104391	20.367	ng/ul	0.00
49) Acenaphthylene-d8	14.186	160	1382292	21.204	ng/ul	0.00
54) 4-Nitrophenol-d4	14.704	143	190228	17.551	ng/ul	0.00
60) Fluorene-d10	15.480	176	1013692	20.963	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	170022	17.835	ng/ul	0.00
73) Anthracene-d10	17.327	188	1472677	21.245	ng/ul	0.00
81) Pyrene-d10	19.615	212	1461282	21.771	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.597	264	1261335	21.130	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.310	88	65631	9.279	ng/uL	98
5) Pyridine	3.716	79	430980	21.411	ng/ul	98
6) Benzaldehyde	6.998	77	304357	25.473	ng/ul	97
8) Phenol	7.051	94	531087	19.544	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.287	93	443059	20.323	ng/ul	100
12) 2-Chlorophenol	7.416	128	444561	20.820	ng/ul	98
13) 2-Methylphenol	8.304	108	395947	19.098	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.392	45	643178	20.772	ng/ul	100
16) Acetophenone	8.687	105	664977	19.660	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.669	70	342167	20.366	ng/ul	99
18) 4-Methylphenol	8.634	108	436499	19.113	ng/ul	97
19) Hexachloroethane	8.928	117	185361	21.874	ng/ul	92
22) Nitrobenzene	9.069	77	500835	21.577	ng/ul	99
23) Isophorone	9.592	82	911763	20.366	ng/ul	99
25) 2-Nitrophenol	9.775	139	247946	20.708	ng/ul	97
26) 2,4-Dimethylphenol	9.839	107	486851	21.625	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.075	93	598192	21.492	ng/ul	99
29) 2,4-Dichlorophenol	10.310	162	411419	20.700	ng/ul	99
30) Naphthalene	10.704	128	1460527	21.263	ng/ul	100
32) 4-Chloroaniline	10.828	127	600802	19.645	ng/ul	98
33) Hexachlorobutadiene	10.980	225	261918	21.398	ng/ul	99
34) Caprolactam	11.622	113	105350	16.108	ng/ul	98
35) 4-Chloro-3-methylphenol	11.951	107	417125	20.233	ng/ul	97
36) 2-Methylnaphthalene	12.316	142	978195	20.621	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110222\
 Data File : BM037310.D
 Acq On : 02 Nov 2022 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020194

Quant Time: Nov 02 22:52:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.533	142	982305	20.579	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.680	216	501780	21.318	ng/ul	99
40) Hexachlorocyclopentadiene	12.657	237	263204	20.002	ng/ul	98
41) 2,4,6-Trichlorophenol	12.927	196	318534	20.964	ng/ul	97
42) 2,4,5-Trichlorophenol	12.998	196	340071	20.785	ng/ul	98
43) 1,1'-Biphenyl	13.327	154	1269335	21.776	ng/ul	100
44) 2-Chloronaphthalene	13.369	162	1015308	21.677	ng/ul	100
45) 2-Nitroaniline	13.586	65	259145	20.983	ng/ul	97
47) Dimethylphthalate	13.957	163	1191729	20.633	ng/ul	100
48) 2,6-Dinitrotoluene	14.080	165	228991	20.248	ng/ul	99
50) Acenaphthylene	14.216	152	1521588	21.354	ng/ul	100
51) 3-Nitroaniline	14.410	138	241300	19.772	ng/ul	99
52) Acenaphthene	14.557	153	1087569	21.450	ng/ul	98
53) 2,4-Dinitrophenol	14.616	184	115486	15.207	ng/ul	94
55) 4-Nitrophenol	14.716	109	146674	18.402	ng/ul	96
56) Dibenzofuran	14.886	168	1451810	21.207	ng/ul	97
57) 2,4-Dinitrotoluene	14.863	165	323835	19.542	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.116	232	284980	20.008	ng/ul	98
59) Diethylphthalate	15.316	149	1177492	20.468	ng/ul	98
61) Fluorene	15.533	166	1220074	21.069	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.533	204	599555	21.068	ng/ul	100
63) 4-Nitroaniline	15.568	138	206216	18.174	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.621	198	187259	18.491	ng/ul	95
67) N-Nitrosodiphenylamine	15.745	169	1001586	22.448	ng/ul	100
68) 4-Bromophenyl-phenylether	16.421	248	345412	21.572	ng/ul	96
69) Hexachlorobenzene	16.533	284	420154	21.501	ng/ul	98
70) Atrazine	16.698	200	305570	19.346	ng/ul	99
71) Pentachlorophenol	16.880	266	219264	18.335	ng/ul	99
72) Phenanthrene	17.274	178	1787071	21.235	ng/ul	99
74) Anthracene	17.362	178	1817496	21.464	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.286	216	497853	22.826	ng/uL	99
76) Pentachlorobenzene	14.804	250	534957	22.542	ng/uL	99
77) Carbazole	17.639	167	1540678	20.169	ng/ul	100
78) Di-n-butylphthalate	18.198	149	1815949	20.634	ng/ul	100
80) Fluoranthene	19.286	202	1866608	22.600	ng/ul	100
82) Pyrene	19.645	202	1961181	22.539	ng/ul	97
83) Butylbenzylphthalate	20.545	149	700901	20.740	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.327	252	591995	19.886	ng/ul	98
85) Benzo(a)anthracene	21.386	228	1773596	21.516	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.315	149	990834	21.026	ng/ul	99
87) Chrysene	21.439	228	1676753	21.618	ng/ul	100
89) Di-n-octyl phthalate	22.221	149	1611985	19.106	ng/ul	100
90) Benzo(b)fluoranthene	23.033	252	1657533	20.546	ng/ul	99
91) Benzo(k)fluoranthene	23.080	252	1690148	21.399	ng/ul	99
93) Benzo(a)pyrene	23.644	252	1448954	21.476	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.180	276	1683868	23.320	ng/ul	99
95) Dibenzo(a,h)anthracene	26.191	278	1522877	23.453	ng/ul	100
96) Benzo(g,h,i)perylene	26.921	276	1479051	23.907	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110222\
 Data File : BM037310.D
 Acq On : 02 Nov 2022 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020194

Quant Time: Nov 02 22:52:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.279 ng/uL
RT: 3.310 min Scan# 38
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion: 88 Resp: 65631
Ion Ratio Lower Upper
88 100
43 29.5 23.4 35.0
58 76.8 59.4 89.0



#5
Pyridine
Concen: 21.411 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 79 Resp: 430980
Ion Ratio Lower Upper
79 100
52 76.5 61.8 92.6
51 36.0 30.8 46.2









#6
Benzaldehyde
Concen: 25.473 ng/ul
RT: 6.998 min Scan# 665
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion: 77 Resp: 304357
Ion Ratio Lower Upper
77 100
105 98.5 81.3 121.9
106 95.0 78.8 118.2



#8
Phenol
Concen: 19.544 ng/ul
RT: 7.051 min Scan# 674
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 94 Resp: 531087
Ion Ratio Lower Upper
94 100
65 28.0 22.6 33.8
66 36.0 27.4 41.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.323 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion: 93 Resp: 443059
Ion Ratio Lower Upper
93 100
63 78.2 63.0 94.4
95 32.6 26.0 39.0



#12
2-Chlorophenol
Concen: 20.820 ng/ul
RT: 7.416 min Scan# 736
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 128 Resp: 444561
Ion Ratio Lower Upper
128 100
64 46.6 36.0 54.0
130 31.9 25.9 38.9









#13
2-Methylphenol
Concen: 19.098 ng/ul
RT: 8.304 min Scan# 887
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:108 Resp: 395947
Ion Ratio Lower Upper
108 100
107 89.3 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.772 ng/ul
RT: 8.392 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 45 Resp: 643178
Ion Ratio Lower Upper
45 100
77 14.0 11.1 16.7
79 10.7 8.4 12.6









#16
Acetophenone
Concen: 19.660 ng/ul
RT: 8.687 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:105 Resp: 664977
Ion Ratio Lower Upper
105 100
77 76.3 59.8 89.6
51 27.6 21.3 31.9



#17
N-Nitroso-di-n-propylamine
Concen: 20.366 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 70 Resp: 342167
Ion Ratio Lower Upper
70 100
42 58.7 46.8 70.2
101 9.1 8.2 12.2
130 20.8 17.7 26.5









#18
4-Methylphenol
Concen: 19.113 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:108 Resp: 436499
Ion Ratio Lower Upper
108 100
107 114.5 89.3 133.9



#19
Hexachloroethane
Concen: 21.874 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:117 Resp: 185361
Ion Ratio Lower Upper
117 100
201 91.0 80.1 120.1
199 57.8 50.0 75.0









#22
Nitrobenzene
Concen: 21.577 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion: 77 Resp: 500835
Ion Ratio Lower Upper
77 100
123 46.9 37.3 55.9
65 13.6 10.1 15.1



#23
Isophorone
Concen: 20.366 ng/ul
RT: 9.592 min Scan# 1106
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 82 Resp: 911763
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 18.2 15.0 22.6









#25
2-Nitrophenol
Concen: 20.708 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:139 Resp: 247946
Ion Ratio Lower Upper
139 100
65 49.4 37.9 56.9
109 28.5 21.4 32.2



#26
2,4-Dimethylphenol
Concen: 21.625 ng/ul
RT: 9.839 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:107 Resp: 486851
Ion Ratio Lower Upper
107 100
121 52.0 42.4 63.6
122 87.8 71.8 107.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.492 ng/ul
 RT: 10.075 min Scan# 1188
 Delta R.T. -0.006 min
 Lab File: BM037310.D
 Acq: 02 Nov 2022 10:10




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020194


Tgt Ion: 93 Resp: 598192
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.6 38.4
 123 11.2 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 20.700 ng/ul
 RT: 10.310 min Scan# 1228
 Delta R.T. 0.000 min
 Lab File: BM037310.D
 Acq: 02 Nov 2022 10:10

Tgt Ion: 162 Resp: 411419
 Ion Ratio Lower Upper
 162 100
 164 64.9 52.7 79.1
 98 34.5 27.4 41.2









#30
Naphthalene
Concen: 21.263 ng/ul
RT: 10.704 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:128 Resp: 1460527
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.2 10.4 15.6



#32
4-Chloroaniline
Concen: 19.645 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:127 Resp: 600802
Ion Ratio Lower Upper
127 100
129 33.0 25.4 38.2





#33

Hexachlorobutadiene

Concen: 21.398 ng/ul

RT: 10.980 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:225 Resp: 261918

Ion Ratio Lower Upper

225 100

223 64.5 50.4 75.6

227 64.8 51.7 77.5



#34

Caprolactam

Concen: 16.108 ng/ul

RT: 11.622 min Scan# 1451

Delta R.T. 0.000 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10




Tgt Ion:113 Resp: 105350


Ion Ratio Lower Upper

113 100

55 173.4 137.7 206.5

56 123.9 97.0 145.6









#35
4-Chloro-3-methylphenol
Concen: 20.233 ng/ul
RT: 11.951 min Scan# 1507
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:107 Resp: 417125
Ion Ratio Lower Upper
107 100
144 27.4 23.3 34.9
142 84.4 69.4 104.2



#36
2-Methylnaphthalene
Concen: 20.621 ng/ul
RT: 12.316 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:142 Resp: 978195
Ion Ratio Lower Upper
142 100
141 87.5 71.4 107.2





#37

1-Methylnaphthalene

Concen: 20.579 ng/ul

RT: 12.533 min Scan# 1606

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:142 Resp: 982305

Ion Ratio Lower Upper

142 100

141 91.4 73.0 109.6

116 3.5 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.318 ng/ul

RT: 12.680 min Scan# 1631

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Tgt Ion:216 Resp: 501780




Ion Ratio Lower Upper


216 100

214 78.1 62.2 93.4

179 21.1 16.2 24.4

108 17.7 12.8 19.2









#40
Hexachlorocyclopentadiene
Concen: 20.002 ng/ul
RT: 12.657 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:237 Resp: 263204
Ion Ratio Lower Upper
237 100
235 61.7 50.6 75.8
272 13.9 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.964 ng/ul
RT: 12.927 min Scan# 1673
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:196 Resp: 318534
Ion Ratio Lower Upper
196 100
198 94.8 78.5 117.7
200 30.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.785 ng/ul
RT: 12.998 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:196 Resp: 340071
Ion Ratio Lower Upper
196 100
198 94.8 78.0 117.0
200 30.3 24.9 37.3



#43
1,1'-Biphenyl
Concen: 21.776 ng/ul
RT: 13.327 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:154 Resp: 1269335
Ion Ratio Lower Upper
154 100
153 39.6 31.9 47.9
76 13.1 10.6 15.8









#44
2-Chloronaphthalene
Concen: 21.677 ng/ul
RT: 13.369 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:162 Resp: 1015308
Ion Ratio Lower Upper
162 100
164 33.0 26.6 39.8
127 37.5 29.8 44.8



#45
2-Nitroaniline
Concen: 20.983 ng/ul
RT: 13.586 min Scan# 1785
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion: 65 Resp: 259145
Ion Ratio Lower Upper
65 100
92 69.1 53.9 80.9
138 114.5 89.1 133.7





#47

Dimethylphthalate

Concen: 20.633 ng/ul

RT: 13.957 min Scan# 1848

Delta R.T. 0.000 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:163 Resp: 1191729

Ion Ratio Lower Upper

163 100

194 3.9 3.3 4.9

164 9.9 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 20.248 ng/ul

RT: 14.080 min Scan# 1869

Delta R.T. 0.000 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10




Tgt Ion:165 Resp: 228991


Ion Ratio Lower Upper

165 100

89 57.5 45.4 68.0

121 20.8 17.0 25.6









#50
Acenaphthylene
Concen: 21.354 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:152 Resp: 1521588
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.0 10.5 15.7



#51
3-Nitroaniline
Concen: 19.772 ng/ul
RT: 14.410 min Scan# 1925
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:138 Resp: 241300
Ion Ratio Lower Upper
138 100
108 9.4 7.1 10.7
92 108.2 85.5 128.3









#52
Acenaphthene
Concen: 21.450 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:153 Resp: 1087569
Ion Ratio Lower Upper
153 100
152 46.9 38.2 57.2
154 86.8 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 15.207 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:184 Resp: 115486
Ion Ratio Lower Upper
184 100
63 68.4 49.0 73.6
154 59.5 45.9 68.9









#55
4-Nitrophenol
Concen: 18.402 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:109 Resp: 146674
Ion Ratio Lower Upper
109 100
139 144.0 117.1 175.7
65 139.6 106.2 159.4



#56
Dibenzofuran
Concen: 21.207 ng/ul
RT: 14.886 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:168 Resp: 1451810
Ion Ratio Lower Upper
168 100
139 37.3 28.4 42.6





#57

2,4-Dinitrotoluene

Concen: 19.542 ng/ul

RT: 14.863 min Scan# 2002

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:165 Resp: 323835

Ion Ratio Lower Upper

165 100

63 46.1 32.8 49.2

182 2.6 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.008 ng/ul

RT: 15.116 min Scan# 2045

Delta R.T. 0.000 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Tgt Ion:232 Resp: 284980




Ion Ratio Lower Upper


232 100

131 47.6 36.5 54.7

130 2.6 1.8 2.8

166 31.1 25.0 37.6









#59
Diethylphthalate
Concen: 20.468 ng/ul
RT: 15.316 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:149 Resp: 1177492
Ion Ratio Lower Upper
149 100
177 21.4 17.8 26.8
150 11.9 9.8 14.8



#61
Fluorene
Concen: 21.069 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:166 Resp: 1220074
Ion Ratio Lower Upper
166 100
165 98.3 77.9 116.9
167 14.0 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 21.068 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:204 Resp: 599555
Ion Ratio Lower Upper
204 100
206 33.3 26.2 39.2
141 55.4 44.2 66.4



#63
4-Nitroaniline
Concen: 18.174 ng/ul
RT: 15.568 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:138 Resp: 206216
Ion Ratio Lower Upper
138 100
92 50.7 38.5 57.7
108 75.1 57.0 85.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.491 ng/ul
 RT: 15.621 min Scan# 2131
 Delta R.T. -0.006 min
 Lab File: BM037310.D
 Acq: 02 Nov 2022 10:10




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020194


Tgt Ion:198 Resp: 187259
 Ion Ratio Lower Upper
 198 100
 182 4.9 3.4 5.0
 77 25.0 18.1 27.1



#67
 N-Nitrosodiphenylamine
 Concen: 22.448 ng/ul
 RT: 15.745 min Scan# 2152
 Delta R.T. -0.006 min
 Lab File: BM037310.D
 Acq: 02 Nov 2022 10:10

Tgt Ion:169 Resp: 1001586
 Ion Ratio Lower Upper
 169 100
 168 67.6 54.0 81.0
 167 36.6 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 21.572 ng/ul
RT: 16.421 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:248 Resp: 345412
Ion Ratio Lower Upper
248 100
250 96.4 78.6 117.8
141 70.7 52.5 78.7



#69
Hexachlorobenzene
Concen: 21.501 ng/ul
RT: 16.533 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:284 Resp: 420154
Ion Ratio Lower Upper
284 100
142 33.2 27.2 40.8
249 28.1 23.9 35.9





#70

Atrazine

Concen: 19.346 ng/ul

RT: 16.698 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:200 Resp: 305570

Ion Ratio Lower Upper

200 100

173 27.1 21.8 32.8

215 49.6 40.2 60.4



#71

Pentachlorophenol

Concen: 18.335 ng/ul

RT: 16.880 min Scan# 2345

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10




Tgt Ion:266 Resp: 219264


Ion Ratio Lower Upper

266 100

264 61.0 48.3 72.5

268 63.2 49.5 74.3





#72

Phenanthrene

Concen: 21.235 ng/ul

RT: 17.274 min Scan# 2412

Delta R.T. 0.000 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:178 Resp: 1787071

Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 19.6 15.4 23.2



#74

Anthracene

Concen: 21.464 ng/ul

RT: 17.362 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10




Tgt Ion:178 Resp: 1817496


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 18.6 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.826 ng/uL
RT: 13.286 min Scan# 1734
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:216 Resp: 497853
Ion Ratio Lower Upper
216 100
214 79.4 63.4 95.2
218 47.2 38.5 57.7



#76
Pentachlorobenzene
Concen: 22.542 ng/uL
RT: 14.804 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:250 Resp: 534957
Ion Ratio Lower Upper
250 100
248 63.1 50.1 75.1
252 64.5 51.0 76.6









#77
 Carbazole
 Concen: 20.169 ng/ul
 RT: 17.639 min Scan# 2474
 Delta R.T. -0.006 min
 Lab File: BM037310.D
 Acq: 02 Nov 2022 10:10




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020194


Tgt Ion:167 Resp: 1540678
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.0 25.6
 139 12.5 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 20.634 ng/ul
 RT: 18.198 min Scan# 2569
 Delta R.T. 0.000 min
 Lab File: BM037310.D
 Acq: 02 Nov 2022 10:10

Tgt Ion:149 Resp: 1815949
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 4.9 3.8 5.8









#80
Fluoranthene
Concen: 22.600 ng/ul
RT: 19.286 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:202 Resp: 1866608
Ion Ratio Lower Upper
202 100
101 11.6 9.2 13.8
100 8.7 7.1 10.7



#82
Pyrene
Concen: 22.539 ng/ul
RT: 19.645 min Scan# 2815
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:202 Resp: 1961181
Ion Ratio Lower Upper
202 100
101 14.1 10.4 15.6
100 11.6 8.5 12.7









#83
Butylbenzylphthalate
Concen: 20.740 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:149 Resp: 700901
Ion Ratio Lower Upper
149 100
91 67.9 54.2 81.4
206 20.3 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 19.886 ng/ul
RT: 21.327 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:252 Resp: 591995
Ion Ratio Lower Upper
252 100
254 65.8 51.3 76.9
126 10.6 8.9 13.3





#85

Benzo(a)anthracene

Concen: 21.516 ng/ul

RT: 21.386 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:228 Resp: 1773596

Ion Ratio Lower Upper

228 100

229 19.8 16.0 24.0

226 27.2 22.2 33.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.026 ng/ul

RT: 21.315 min Scan# 3099

Delta R.T. 0.000 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10




Tgt Ion:149 Resp: 990834


Ion Ratio Lower Upper

149 100

167 27.1 21.9 32.9

279 4.4 3.4 5.2









#87
Chrysene
Concen: 21.618 ng/ul
RT: 21.439 min Scan# 3120
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:228 Resp: 1676753
Ion Ratio Lower Upper
228 100
226 29.8 24.1 36.1
229 19.6 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 19.106 ng/ul
RT: 22.221 min Scan# 3253
Delta R.T. 0.000 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:149 Resp: 1611985









#90
Benzo(b)fluoranthene
Concen: 20.546 ng/ul
RT: 23.033 min Scan# 3391
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10




Instrument :
BNA_M
ClientSampleId :
SSTD020194


Tgt Ion:252 Resp: 1657533
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.6 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 21.399 ng/ul
RT: 23.080 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BM037310.D
Acq: 02 Nov 2022 10:10

Tgt Ion:252 Resp: 1690148
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 10.9 8.2 12.4





#93

Benzo(a)pyrene

Concen: 21.476 ng/ul

RT: 23.644 min Scan# 3495

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:252 Resp: 1448954

Ion Ratio Lower Upper

252 100

253 21.6 17.6 26.4

125 11.8 9.0 13.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.320 ng/ul

RT: 26.180 min Scan# 3926

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10




Tgt Ion:276 Resp: 1683868


Ion Ratio Lower Upper

276 100

138 26.1 20.4 30.6

277 25.1 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 23.453 ng/ul

RT: 26.191 min Scan# 3928

Delta R.T. -0.006 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Instrument :

BNA_M

ClientSampleId :

SSTD020194





Tgt Ion:278 Resp: 1522877

Ion Ratio Lower Upper

278 100

139 16.6 13.4 20.2

279 23.8 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 23.907 ng/ul

RT: 26.921 min Scan# 4052

Delta R.T. -0.012 min

Lab File: BM037310.D

Acq: 02 Nov 2022 10:10

Tgt Ion:276 Resp: 1479051

Ion Ratio Lower Upper

276 100

138 22.0 17.4 26.2

277 23.8 19.4 29.0

