

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110222\
 Data File : BM037319.D
 Acq On : 02 Nov 2022 16:53
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020195

Quant Time: Nov 02 22:54:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	299065	20.000	ng/ul	0.00
20) Naphthalene-d8	10.657	136	1378880	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.492	164	878619	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.233	188	1811853	20.000	ng/ul	0.00
79) Chrysene-d12	21.403	240	1733297	20.000	ng/ul	0.00
88) Perylene-d12	23.756	264	1399880	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	57117	8.419	ng/uL	0.00
4) Pyridine-d5	3.699	84	432008	21.332	ng/ul	0.00
7) Phenol-d5	7.028	99	510028	19.596	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	329032	20.602	ng/ul	0.00
11) 2-Chlorophenol-d4	7.387	132	415246	20.882	ng/ul	0.00
15) 4-Methylphenol-d8	8.575	113	419911	19.757	ng/ul	0.00
21) Nitrobenzene-d5	9.028	128	214190	20.795	ng/ul	0.00
24) 2-Nitrophenol-d4	9.745	143	227632	20.289	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	432670	20.917	ng/ul	0.00
31) 4-Chloroaniline-d4	10.804	131	639068	20.217	ng/ul	0.00
46) Dimethylphthalate-d6	13.910	166	1268869	20.887	ng/ul	0.00
49) Acenaphthylene-d8	14.186	160	1558063	21.333	ng/ul	0.00
54) 4-Nitrophenol-d4	14.704	143	237300	19.542	ng/ul	0.00
60) Fluorene-d10	15.480	176	1159269	21.399	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	214562	18.697	ng/ul	0.00
73) Anthracene-d10	17.333	188	1781373	21.348	ng/ul	0.00
81) Pyrene-d10	19.621	212	1950417	21.534	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.603	264	1539117	21.359	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.310	88	62209	8.728	ng/uL	99
5) Pyridine	3.716	79	424577	20.931	ng/ul	97
6) Benzaldehyde	7.004	77	314586	26.127	ng/ul	97
8) Phenol	7.057	94	539112	19.687	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.287	93	449605	20.465	ng/ul	99
12) 2-Chlorophenol	7.416	128	453864	21.093	ng/ul	99
13) 2-Methylphenol	8.304	108	413133	19.774	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.392	45	651774	20.889	ng/ul	99
16) Acetophenone	8.687	105	699020	20.508	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.675	70	354508	20.939	ng/ul	98
18) 4-Methylphenol	8.640	108	461853	20.068	ng/ul	100
19) Hexachloroethane	8.928	117	186364	21.823	ng/ul	96
22) Nitrobenzene	9.069	77	530611	21.344	ng/ul	99
23) Isophorone	9.592	82	981830	20.477	ng/ul	99
25) 2-Nitrophenol	9.775	139	266317	20.768	ng/ul	98
26) 2,4-Dimethylphenol	9.839	107	512672	21.262	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.081	93	629231	21.108	ng/ul	99
29) 2,4-Dichlorophenol	10.310	162	447109	21.004	ng/ul	99
30) Naphthalene	10.710	128	1560581	21.213	ng/ul	100
32) 4-Chloroaniline	10.828	127	668636	20.413	ng/ul	99
33) Hexachlorobutadiene	10.986	225	280145	21.370	ng/ul	99
34) Caprolactam	11.622	113	131960	18.838	ng/ul	98
35) 4-Chloro-3-methylphenol	11.957	107	465393	21.078	ng/ul	99
36) 2-Methylnaphthalene	12.316	142	1068121	21.024	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110222\
 Data File : BM037319.D
 Acq On : 02 Nov 2022 16:53
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020195

Quant Time: Nov 02 22:54:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.539	142	1075070	21.029	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.686	216	553378	20.985	ng/ul	100
40) Hexachlorocyclopentadiene	12.657	237	279416	18.953	ng/ul	99
41) 2,4,6-Trichlorophenol	12.933	196	356374	20.935	ng/ul	99
42) 2,4,5-Trichlorophenol	13.004	196	387832	21.159	ng/ul	99
43) 1,1'-Biphenyl	13.327	154	1403989	21.500	ng/ul	100
44) 2-Chloronaphthalene	13.369	162	1118150	21.309	ng/ul	99
45) 2-Nitroaniline	13.586	65	306952	22.185	ng/ul	98
47) Dimethylphthalate	13.957	163	1366698	21.121	ng/ul	100
48) 2,6-Dinitrotoluene	14.080	165	268111	21.161	ng/ul	97
50) Acenaphthylene	14.216	152	1712196	21.449	ng/ul	100
51) 3-Nitroaniline	14.410	138	293478	21.464	ng/ul	98
52) Acenaphthene	14.557	153	1211908	21.336	ng/ul	99
53) 2,4-Dinitrophenol	14.621	184	146879	17.263	ng/ul	96
55) 4-Nitrophenol	14.721	109	185638	20.790	ng/ul	96
56) Dibenzofuran	14.892	168	1639445	21.376	ng/ul	99
57) 2,4-Dinitrotoluene	14.869	165	393961	21.221	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.116	232	326473	20.459	ng/ul	97
59) Diethylphthalate	15.316	149	1364743	21.175	ng/ul	99
61) Fluorene	15.539	166	1393691	21.482	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.533	204	682678	21.412	ng/ul	99
63) 4-Nitroaniline	15.574	138	264859	20.835	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.627	198	230013	18.867	ng/ul	97
67) N-Nitrosodiphenylamine	15.751	169	1152309	21.454	ng/ul	99
68) 4-Bromophenyl-phenylether	16.427	248	400479	20.777	ng/ul	99
69) Hexachlorobenzene	16.533	284	490221	20.839	ng/ul	99
70) Atrazine	16.704	200	382145	20.098	ng/ul	99
71) Pentachlorophenol	16.886	266	273592	19.005	ng/ul	97
72) Phenanthrene	17.274	178	2154499	21.267	ng/ul	100
74) Anthracene	17.362	178	2189265	21.478	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.292	216	552072	21.027	ng/ul	99
76) Pentachlorobenzene	14.804	250	604342	21.154	ng/ul	99
77) Carbazole	17.639	167	1936798	21.062	ng/ul	99
78) Di-n-butylphthalate	18.198	149	2268375	21.411	ng/ul	100
80) Fluoranthene	19.286	202	2411503	21.636	ng/ul	99
82) Pyrene	19.651	202	2569918	21.887	ng/ul	100
83) Butylbenzylphthalate	20.545	149	985254	21.604	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.327	252	841031	20.936	ng/ul	99
85) Benzo(a)anthracene	21.392	228	2409967	21.665	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.315	149	1378766	21.682	ng/ul	100
87) Chrysene	21.445	228	2270673	21.694	ng/ul	100
89) Di-n-octyl phthalate	22.221	149	2119373	20.809	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	2117494	21.744	ng/ul	99
91) Benzo(k)fluoranthene	23.086	252	2057887	21.583	ng/ul	99
93) Benzo(a)pyrene	23.650	252	1750847	21.497	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.185	276	1846394	21.183	ng/ul	99
95) Dibenzo(a,h)anthracene	26.197	278	1663741	21.225	ng/ul	99
96) Benzo(g,h,i)perylene	26.932	276	1591535	21.310	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.728 ng/uL
RT: 3.310 min Scan# 38
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion: 88 Resp: 62209
Ion Ratio Lower Upper
88 100
43 30.2 23.4 35.0
58 75.1 59.4 89.0



#5
Pyridine
Concen: 20.931 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion: 79 Resp: 424577
Ion Ratio Lower Upper
79 100
52 75.8 61.8 92.6
51 35.4 30.8 46.2









#6
Benzaldehyde
Concen: 26.127 ng/ul
RT: 7.004 min Scan# 665
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion: 77 Resp: 314586
Ion Ratio Lower Upper
77 100
105 100.2 81.3 121.9
106 94.6 78.8 118.2



#8
Phenol
Concen: 19.687 ng/ul
RT: 7.057 min Scan# 675
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion: 94 Resp: 539112
Ion Ratio Lower Upper
94 100
65 27.9 22.6 33.8
66 35.0 27.4 41.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.465 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion: 93 Resp: 449605
Ion Ratio Lower Upper
93 100
63 77.9 63.0 94.4
95 32.7 26.0 39.0



#12
2-Chlorophenol
Concen: 21.093 ng/ul
RT: 7.416 min Scan# 736
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:128 Resp: 453864
Ion Ratio Lower Upper
128 100
64 45.4 36.0 54.0
130 31.9 25.9 38.9









#13
2-Methylphenol
Concen: 19.774 ng/ul
RT: 8.304 min Scan# 887
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:108 Resp: 413133
Ion Ratio Lower Upper
108 100
107 89.5 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.889 ng/ul
RT: 8.392 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion: 45 Resp: 651774
Ion Ratio Lower Upper
45 100
77 14.1 11.1 16.7
79 10.8 8.4 12.6









#16
Acetophenone
Concen: 20.508 ng/ul
RT: 8.687 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:105 Resp: 699020
Ion Ratio Lower Upper
105 100
77 77.3 59.8 89.6
51 28.0 21.3 31.9



#17
N-Nitroso-di-n-propylamine
Concen: 20.939 ng/ul
RT: 8.675 min Scan# 950
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion: 70 Resp: 354508
Ion Ratio Lower Upper
70 100
42 56.7 46.8 70.2
101 9.7 8.2 12.2
130 21.7 17.7 26.5









#18
4-Methylphenol
Concen: 20.068 ng/ul
RT: 8.640 min Scan# 943
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:108 Resp: 461853
Ion Ratio Lower Upper
108 100
107 112.1 89.3 133.9



#19
Hexachloroethane
Concen: 21.823 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:117 Resp: 186364
Ion Ratio Lower Upper
117 100
201 94.5 80.1 120.1
199 61.1 50.0 75.0









#22
Nitrobenzene
Concen: 21.344 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion: 77 Resp: 530611
Ion Ratio Lower Upper
77 100
123 47.3 37.3 55.9
65 12.9 10.1 15.1



#23
Isophorone
Concen: 20.477 ng/ul
RT: 9.592 min Scan# 1106
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion: 82 Resp: 981830
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.0
138 18.0 15.0 22.6









#25
2-Nitrophenol
Concen: 20.768 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:139 Resp: 266317
Ion Ratio Lower Upper
139 100
65 49.7 37.9 56.9
109 26.5 21.4 32.2



#26
2,4-Dimethylphenol
Concen: 21.262 ng/ul
RT: 9.839 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:107 Resp: 512672
Ion Ratio Lower Upper
107 100
121 51.7 42.4 63.6
122 91.4 71.8 107.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.108 ng/ul

RT: 10.081 min Scan# 1189

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Instrument :

BNA_M

ClientSampleId :

SSTD020195





Tgt Ion: 93 Resp: 629231

Ion Ratio Lower Upper

93 100

95 33.1 25.6 38.4

123 11.7 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 21.004 ng/ul

RT: 10.310 min Scan# 1228

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53




Tgt Ion:162 Resp: 447109


Ion Ratio Lower Upper

162 100

164 65.5 52.7 79.1

98 34.9 27.4 41.2









#30
Naphthalene
Concen: 21.213 ng/ul
RT: 10.710 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:128 Resp: 1560581
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 20.413 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:127 Resp: 668636
Ion Ratio Lower Upper
127 100
129 32.3 25.4 38.2









#33
Hexachlorobutadiene
Concen: 21.370 ng/ul
RT: 10.986 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:225 Resp: 280145
Ion Ratio Lower Upper
225 100
223 61.5 50.4 75.6
227 65.0 51.7 77.5



#34
Caprolactam
Concen: 18.838 ng/ul
RT: 11.622 min Scan# 1451
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:113 Resp: 131960
Ion Ratio Lower Upper
113 100
55 173.2 137.7 206.5
56 126.4 97.0 145.6









#35
4-Chloro-3-methylphenol
Concen: 21.078 ng/ul
RT: 11.957 min Scan# 1508
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:107 Resp: 465393
Ion Ratio Lower Upper
107 100
144 28.6 23.3 34.9
142 88.0 69.4 104.2



#36
2-Methylnaphthalene
Concen: 21.024 ng/ul
RT: 12.316 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:142 Resp: 1068121
Ion Ratio Lower Upper
142 100
141 88.4 71.4 107.2





#37

1-Methylnaphthalene

Concen: 21.029 ng/ul

RT: 12.539 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Instrument :

BNA_M

ClientSampleId :

SSTD020195





Tgt Ion:142 Resp: 1075070

Ion Ratio Lower Upper

142 100

141 92.5 73.0 109.6

116 3.6 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.985 ng/ul

RT: 12.686 min Scan# 1632

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Tgt Ion:216 Resp: 553378




Ion Ratio Lower Upper


216 100

214 77.6 62.2 93.4

179 20.4 16.2 24.4

108 16.5 12.8 19.2









#40
Hexachlorocyclopentadiene
Concen: 18.953 ng/ul
RT: 12.657 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:237 Resp: 279416
Ion Ratio Lower Upper
237 100
235 62.9 50.6 75.8
272 14.0 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.935 ng/ul
RT: 12.933 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:196 Resp: 356374
Ion Ratio Lower Upper
196 100
198 97.2 78.5 117.7
200 31.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 21.159 ng/ul
RT: 13.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:196 Resp: 387832
Ion Ratio Lower Upper
196 100
198 96.0 78.0 117.0
200 31.5 24.9 37.3



#43
1,1'-Biphenyl
Concen: 21.500 ng/ul
RT: 13.327 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:154 Resp: 1403989
Ion Ratio Lower Upper
154 100
153 39.8 31.9 47.9
76 13.1 10.6 15.8





#44

2-Chloronaphthalene

Concen: 21.309 ng/ul

RT: 13.369 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Instrument :

BNA_M

ClientSampleId :

SSTD020195





Tgt Ion:162 Resp: 1118150

Ion Ratio Lower Upper

162 100

164 32.3 26.6 39.8

127 37.4 29.8 44.8



#45

2-Nitroaniline

Concen: 22.185 ng/ul

RT: 13.586 min Scan# 1785

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53




Tgt Ion: 65 Resp: 306952


Ion Ratio Lower Upper

65 100

92 69.3 53.9 80.9

138 109.1 89.1 133.7









#47
Dimethylphthalate
Concen: 21.121 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:163 Resp: 1366698
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.161 ng/ul
RT: 14.080 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:165 Resp: 268111
Ion Ratio Lower Upper
165 100
89 59.3 45.4 68.0
121 21.8 17.0 25.6









#50
Acenaphthylene
Concen: 21.449 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:152 Resp: 1712196
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 21.464 ng/ul
RT: 14.410 min Scan# 1925
Delta R.T. -0.006 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:138 Resp: 293478
Ion Ratio Lower Upper
138 100
108 9.0 7.1 10.7
92 109.2 85.5 128.3









#52
Acenaphthene
Concen: 21.336 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:153 Resp: 1211908
Ion Ratio Lower Upper
153 100
152 47.7 38.2 57.2
154 87.7 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 17.263 ng/ul
RT: 14.621 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:184 Resp: 146879
Ion Ratio Lower Upper
184 100
63 64.9 49.0 73.6
154 59.1 45.9 68.9









#55
4-Nitrophenol
Concen: 20.790 ng/ul
RT: 14.721 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:109 Resp: 185638
Ion Ratio Lower Upper
109 100
139 142.1 117.1 175.7
65 137.9 106.2 159.4



#56
Dibenzofuran
Concen: 21.376 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:168 Resp: 1639445
Ion Ratio Lower Upper
168 100
139 36.2 28.4 42.6





#57

2,4-Dinitrotoluene

Concen: 21.221 ng/ul

RT: 14.869 min Scan# 2002

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Instrument :

BNA_M

ClientSampleId :

SSTD020195





Tgt Ion:165 Resp: 393961

Ion Ratio Lower Upper

165 100

63 43.6 32.8 49.2

182 2.6 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.459 ng/ul

RT: 15.116 min Scan# 2045

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Tgt Ion:232 Resp: 326473




Ion Ratio Lower Upper


232 100

131 47.7 36.5 54.7

130 2.6 1.8 2.8

166 32.5 25.0 37.6









#59
Diethylphthalate
Concen: 21.175 ng/ul
RT: 15.316 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:149 Resp: 1364743
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 12.3 9.8 14.8



#61
Fluorene
Concen: 21.482 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:166 Resp: 1393691
Ion Ratio Lower Upper
166 100
165 98.4 77.9 116.9
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 21.412 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:204 Resp: 682678
Ion Ratio Lower Upper
204 100
206 32.5 26.2 39.2
141 53.8 44.2 66.4



#63
4-Nitroaniline
Concen: 20.835 ng/ul
RT: 15.574 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:138 Resp: 264859
Ion Ratio Lower Upper
138 100
92 47.7 38.5 57.7
108 71.4 57.0 85.4









#66
4,6-Dinitro-2-methylphenol
Concen: 18.867 ng/ul
RT: 15.627 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:198 Resp: 230013
Ion Ratio Lower Upper
198 100
182 4.4 3.4 5.0
77 24.0 18.1 27.1



#67
N-Nitrosodiphenylamine
Concen: 21.454 ng/ul
RT: 15.751 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:169 Resp: 1152309
Ion Ratio Lower Upper
169 100
168 67.6 54.0 81.0
167 37.0 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 20.777 ng/ul
RT: 16.427 min Scan# 2268
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:248 Resp: 400479
Ion Ratio Lower Upper
248 100
250 96.7 78.6 117.8
141 65.2 52.5 78.7



#69
Hexachlorobenzene
Concen: 20.839 ng/ul
RT: 16.533 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:284 Resp: 490221
Ion Ratio Lower Upper
284 100
142 33.9 27.2 40.8
249 28.8 23.9 35.9









#70
Atrazine
Concen: 20.098 ng/ul
RT: 16.704 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:200 Resp: 382145
Ion Ratio Lower Upper
200 100
173 26.7 21.8 32.8
215 49.4 40.2 60.4



#71
Pentachlorophenol
Concen: 19.005 ng/ul
RT: 16.886 min Scan# 2346
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:266 Resp: 273592
Ion Ratio Lower Upper
266 100
264 62.3 48.3 72.5
268 64.6 49.5 74.3





#72

Phenanthrene

Concen: 21.267 ng/ul

RT: 17.274 min Scan# 2412

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Instrument :

BNA_M

ClientSampleId :

SSTD020195





Tgt Ion:178 Resp: 2154499

Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 19.2 15.4 23.2



#74

Anthracene

Concen: 21.478 ng/ul

RT: 17.362 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53




Tgt Ion:178 Resp: 2189265


Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 18.7 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.027 ng/uL

RT: 13.292 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53

Instrument :

BNA_M

ClientSampleId :

SSTD020195





Tgt Ion:216 Resp: 552072

Ion Ratio Lower Upper

216 100

214 77.6 63.4 95.2

218 48.2 38.5 57.7



#76

Pentachlorobenzene

Concen: 21.154 ng/uL

RT: 14.804 min Scan# 1992

Delta R.T. 0.000 min

Lab File: BM037319.D

Acq: 02 Nov 2022 16:53




Tgt Ion:250 Resp: 604342


Ion Ratio Lower Upper

250 100

248 63.9 50.1 75.1

252 63.2 51.0 76.6









#77
 Carbazole
 Concen: 21.062 ng/ul
 RT: 17.639 min Scan# 2474
 Delta R.T. -0.006 min
 Lab File: BM037319.D
 Acq: 02 Nov 2022 16:53




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020195


Tgt Ion:167 Resp: 1936798
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.0 25.6
 139 12.7 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 21.411 ng/ul
 RT: 18.198 min Scan# 2569
 Delta R.T. 0.000 min
 Lab File: BM037319.D
 Acq: 02 Nov 2022 16:53

Tgt Ion:149 Resp: 2268375
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 4.8 3.8 5.8









#80
Fluoranthene
Concen: 21.636 ng/ul
RT: 19.286 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:202 Resp: 2411503
Ion Ratio Lower Upper
202 100
101 11.8 9.2 13.8
100 9.1 7.1 10.7



#82
Pyrene
Concen: 21.887 ng/ul
RT: 19.651 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:202 Resp: 2569918
Ion Ratio Lower Upper
202 100
101 13.1 10.4 15.6
100 10.5 8.5 12.7









#83
Butylbenzylphthalate
Concen: 21.604 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:149 Resp: 985254
Ion Ratio Lower Upper
149 100
91 67.9 54.2 81.4
206 19.7 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 20.936 ng/ul
RT: 21.327 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:252 Resp: 841031
Ion Ratio Lower Upper
252 100
254 64.7 51.3 76.9
126 10.9 8.9 13.3









#85
Benzo(a)anthracene
Concen: 21.665 ng/ul
RT: 21.392 min Scan# 3111
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:228 Resp: 2409967
Ion Ratio Lower Upper
228 100
229 19.7 16.0 24.0
226 27.4 22.2 33.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.682 ng/ul
RT: 21.315 min Scan# 3099
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:149 Resp: 1378766
Ion Ratio Lower Upper
149 100
167 27.2 21.9 32.9
279 4.1 3.4 5.2









#87
Chrysene
Concen: 21.694 ng/ul
RT: 21.445 min Scan# 3121
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:228 Resp: 2270673
Ion Ratio Lower Upper
228 100
226 30.1 24.1 36.1
229 19.7 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 20.809 ng/ul
RT: 22.221 min Scan# 3253
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:149 Resp: 2119373









#90
Benzo(b)fluoranthene
Concen: 21.744 ng/ul
RT: 23.039 min Scan# 311
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:252 Resp: 2117494
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 10.6 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 21.583 ng/ul
RT: 23.086 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:252 Resp: 2057887
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 10.7 8.2 12.4









#93
Benzo(a)pyrene
Concen: 21.497 ng/ul
RT: 23.650 min Scan# 3495
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53




Instrument :
BNA_M
ClientSampleId :
SSTD020195


Tgt Ion:252 Resp: 1750847
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 11.8 9.0 13.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.183 ng/ul
RT: 26.185 min Scan# 3927
Delta R.T. 0.000 min
Lab File: BM037319.D
Acq: 02 Nov 2022 16:53

Tgt Ion:276 Resp: 1846394
Ion Ratio Lower Upper
276 100
138 25.5 20.4 30.6
277 25.1 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 21.225 ng/ul
 RT: 26.197 min Scan# 3928
 Delta R.T. 0.000 min
 Lab File: BM037319.D
 Acq: 02 Nov 2022 16:53

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020195

Tgt Ion:278 Resp: 1663741
 Ion Ratio Lower Upper
 278 100
 139 16.3 13.4 20.2
 279 24.3 19.2 28.8



#96
 Benzo(g,h,i)perylene
 Concen: 21.310 ng/ul
 RT: 26.932 min Scan# 4054
 Delta R.T. 0.000 min
 Lab File: BM037319.D
 Acq: 02 Nov 2022 16:53

Tgt Ion:276 Resp: 1591535
 Ion Ratio Lower Upper
 276 100
 138 21.5 17.4 26.2
 277 23.9 19.4 29.0

