

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111020\
 Data File : BM027773.D
 Acq On : 10 Nov 2020 17:03
 Operator : JU/CG
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.483

Quant Time: Nov 10 17:39:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM111020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 15:45:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	507	0.40	ng/ul	0.00
2) Naphthalene-d8	11.22	136	1825	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.99	164	1168	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	2493	0.40	ng/ul	0.00
16) Chrysene-d12	21.81	240	1916	0.40	ng/ul	0.00
20) Perylene-d12	24.25	264	1586	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.78	152	1250	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	2615	0.35	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.27	128	2166	0.371	ng/ul	98
5) 2-Methylnaphthalene	12.86	142	1559	0.384	ng/ul	98
7) Acenaphthylene	14.72	152	2203	0.383	ng/ul	98
8) Acenaphthene	15.06	153	1635	0.371	ng/ul	96
9) Fluorene	16.03	166	1949	0.368	ng/ul	96
11) Pentachlorophenol	17.35	266	319	0.355	ng/ul	97
12) Phenanthrene	17.75	178	3021	0.372	ng/ul	99
13) Anthracene	17.84	178	2924	0.370	ng/ul	99
15) Fluoranthene	19.72	202	3365	0.362	ng/ul	99
17) Pyrene	20.08	202	3307	0.389	ng/ul	97
18) Benzo(a)anthracene	21.79	228	2898	0.378	ng/ul	99
19) Chrysene	21.85	228	2938	0.377	ng/ul	99
21) Benzo(b)fluoranthene	23.51	252	2806	0.387	ng/ul	98
22) Benzo(k)fluoranthene	23.55	252	2766	0.388	ng/ul	98
23) Benzo(a)pyrene	24.14	252	2321	0.364	ng/ul	95
24) Indeno(1,2,3-cd)pyrene	26.77	276	2719	0.374	ng/ul#	99
25) Dibenzo(a,h)anthracene	26.78	278	2276	0.375	ng/ul#	95
26) Benzo(g,h,i)perylene	27.55	276	2321	0.377	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111020\
Data File : BM027773.D
Acq On : 10 Nov 2020 17:03
Operator : JU/CG
Sample : SSTDCCC0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.483

Quant Time: Nov 10 17:39:58 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM111020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 15:45:47 2020
Response via : Initial Calibration

