

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111020\
 Data File : BM027777.D
 Acq On : 10 Nov 2020 19:29
 Operator : JU/CG
 Sample : L4661-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F4R04

Manual Integrations
 APPROVED

mohammad
 11/11/2020 10:33:29 AM

Quant Time: Nov 11 01:31:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM111020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 11 01:20:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	1160	0.40	ng/ul	0.00
2) Naphthalene-d8	11.22	136	4628	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.99	164	3045	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	5797m	0.40	ng/ul	0.00
16) Chrysene-d12	21.81	240	4776	0.40	ng/ul	0.00
20) Perylene-d12	24.25	264	4074	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.78	152	2722	0.32	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	6170	0.36	ng/ul	0.00
Target Compounds						
8) Acenaphthene	15.06	153	1211	0.105	ng/ul#	Ovalue 26

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111020\
Data File : BM027777.D
Acq On : 10 Nov 2020 19:29
Operator : JU/CG
Sample : L4661-05
Misc :
ALS Vial : 14 Sample Multiplier: 1




Instrument :
BNA_M
ClientSampleId :
F4R04

Quant Time: Nov 11 01:31:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM111020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 11 01:20:37 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

mohammad
11/11/2020 10:33:29 AM





#8

Acenaphthene

Concen: 0.105 ng/ul

RT: 15.06 min Scan# 3182

Delta R.T. 0.00 min

Lab File: BM027777.D

Acq: 10 Nov 2020 19:29

Instrument :

BNA_M

ClientSampleId :

F4R04

Tot Ion:	153	Resp:	1211
Ion Ratio	Lower	Upper	
153	100		
152	6.7	38.6	58.0#
154	13.5	72.0	108.0#

Manual Integrations
APPROVED

mohammad
11/11/2020 10:33:29 AM

