


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM111020\
Data File : BM027764.D
Acq On : 10 Nov 2020 09:33
Operator : JU/CG
Sample : DFTPP88
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP88

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM111020.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Nov 10 15:45:47 2020



AutoFind: Scans 2571, 2572, 2573; Background Corrected with Scan 2565

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	64.6	28733	PASS
68	69	0.00	2	1.8	460	PASS
69	198	0.00	100	58.6	26056	PASS
70	69	0.00	2	0.3	72	PASS
127	198	10	80	52.7	23456	PASS
197	198	0.00	2	0.7	326	PASS
198	198	100	100	100.0	44498	PASS
199	198	5	9	7.0	3108	PASS
275	198	10	60	28.7	12779	PASS
365	198	1	100	5.0	2208	PASS
441	198	0.01	100	13.5	5995	PASS
442	198	50	100	85.3	37957	PASS
443	442	15	24	19.7	7495	PASS

Scan 2565 (18.074 min): BM027764.D
DFTPP88

m/z	Abundance
44.00	218.0
206.90	288.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

m/z	Abundance
37.00	167.0
38.10	393.0
39.10	2096.0
40.00	223.0
50.10	8055.0
51.10	26656.0
52.10	1470.0
54.90	170.0
55.90	1011.0
57.00	1582.0
61.00	254.0
62.00	507.0
63.00	1279.0
65.10	262.0
68.10	482.0
69.00	24232.0
73.00	298.0
74.00	2522.0
75.00	3906.0
76.00	1378.0
77.00	23840.0
78.10	1520.0
79.00	2090.0
80.00	1549.0
81.00	1926.0
82.00	516.0
82.90	289.0
84.90	363.0
86.00	684.0
87.00	266.0
90.90	528.0
92.00	462.0
93.00	4011.0
93.90	206.0
98.00	3033.0
99.00	2549.0
100.00	233.0
101.00	1045.0
103.00	288.0
103.90	793.0
105.00	554.0
105.90	152.0
107.00	7829.0
108.00	1199.0
108.90	225.0
110.00	11928.0
111.00	1835.0
112.00	219.0
116.10	403.0
117.00	9436.0
117.90	665.0
121.90	531.0
122.90	677.0
124.00	287.0
125.00	268.0
126.20	201.0
127.00	19896.0
128.00	1865.0
129.00	9405.0
130.00	606.0
130.90	199.0
133.90	227.0
134.90	708.0
135.90	364.0
136.90	341.0
139.90	229.0
141.00	1266.0
142.00	584.0
143.00	317.0
145.90	322.0
147.00	829.0
148.00	1687.0
148.90	372.0
150.90	222.0
151.90	157.0
152.80	348.0
153.90	309.0
155.00	691.0
156.00	1052.0
156.80	198.0
157.90	200.0
158.80	151.0
159.90	367.0
161.00	548.0
161.90	152.0
165.00	676.0
166.00	314.0
167.00	2751.0
168.00	1787.0
168.90	278.0
172.00	261.0
172.90	480.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

174.00	475.0
175.00	893.0
175.90	242.0
176.90	255.0
178.00	206.0
179.00	1722.0
180.00	1487.0
181.00	652.0
184.90	978.0
186.00	5906.0
187.00	1841.0
187.90	343.0
188.90	305.0
191.00	323.0
191.90	720.0
193.00	696.0
195.90	842.0
197.10	421.0
197.90	35560.0
199.00	2430.0
199.90	203.0
201.20	222.0
203.00	312.0
203.90	1764.0
205.00	2792.0
206.00	11042.0
207.00	1499.0
207.90	569.0
208.90	176.0
210.10	286.0
211.10	533.0
214.90	191.0
215.90	404.0
217.00	3293.0
217.90	463.0
220.30	154.0
221.00	1336.0
221.70	296.0
222.10	268.0
223.00	705.0
224.00	6097.0
225.00	1650.0
225.90	222.0
227.00	3216.0
228.00	377.0
228.90	601.0
230.90	257.0
234.00	180.0
234.80	210.0
236.80	417.0
239.80	178.0
240.90	311.0
241.90	382.0
242.90	465.0
244.00	3767.0
244.90	565.0
246.00	881.0
247.00	301.0
248.90	154.0
254.00	251.0
255.00	21888.0
256.00	3691.0
256.90	409.0
257.90	1690.0
259.00	316.0
265.00	637.0
272.90	633.0
274.00	1835.0
275.00	8713.0
276.00	1308.0
277.00	1098.0
277.80	254.0
293.00	204.0
296.00	3310.0
296.90	443.0
303.00	395.0
303.90	164.0
314.90	268.0
315.90	160.0
322.90	927.0
326.80	336.0
333.90	582.0
341.00	211.0
346.00	319.0
352.00	216.0
354.00	246.0
364.90	1414.0
365.90	153.0
371.90	342.0
402.90	236.0
422.10	294.0
423.00	1267.0
424.00	324.0
441.00	3858.0
442.00	21552.0
443.00	4518.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

443.90

551.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

m/z	Abundance
36.90	217.0
38.00	493.0
39.10	2690.0
40.00	215.0
44.00	278.0
49.00	356.0
50.10	11074.0
51.10	34976.0
52.10	1897.0
53.10	160.0
54.90	248.0
55.90	1112.0
57.00	2239.0
61.00	497.0
62.10	668.0
63.00	1408.0
64.00	198.0
65.00	619.0
68.10	478.0
69.00	32008.0
70.00	217.0
73.00	442.0
74.00	3924.0
75.00	5811.0
76.00	1583.0
77.00	31016.0
78.00	1822.0
79.00	2906.0
80.00	2017.0
81.00	2733.0
82.00	587.0
83.00	483.0
85.00	645.0
85.90	721.0
86.90	424.0
90.90	662.0
91.90	789.0
93.00	5314.0
93.90	450.0
98.00	4326.0
99.00	2796.0
101.00	1530.0
102.90	685.0
104.00	1169.0
105.00	1102.0
106.00	257.0
107.00	10837.0
108.00	1653.0
109.00	336.0
110.00	16560.0
111.00	2729.0
111.90	371.0
113.00	211.0
116.00	497.0
117.00	12199.0
118.00	1163.0
121.90	730.0
122.90	1115.0
124.00	447.0
124.90	428.0
126.20	341.0
127.00	28648.0
128.00	2246.0
129.00	13619.0
130.00	1376.0
131.00	250.0
133.90	366.0
135.00	1150.0
136.00	563.0
137.10	730.0
140.00	224.0
141.00	1798.0
142.00	592.0
143.00	404.0
144.00	158.0
145.90	418.0
146.90	1129.0
148.00	2511.0
149.00	442.0
150.90	151.0
151.90	179.0
152.90	683.0
154.00	539.0
155.10	984.0
156.00	1520.0
156.80	237.0
158.00	423.0
158.90	294.0
159.90	654.0
161.00	856.0
161.80	221.0
164.90	666.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

165.90	527.0
167.00	4149.0
168.00	2128.0
169.00	237.0
170.80	229.0
171.90	444.0
173.00	300.0
174.00	858.0
175.10	1512.0
175.90	414.0
176.90	706.0
178.00	173.0
178.90	2583.0
180.00	1806.0
181.00	839.0
183.90	184.0
185.00	1423.0
186.00	9879.0
187.00	2646.0
188.10	270.0
189.00	694.0
191.00	420.0
192.00	740.0
193.00	1083.0
194.00	268.0
196.00	1394.0
197.10	557.0
198.00	54160.0
198.90	3686.0
199.90	282.0
201.20	197.0
202.80	374.0
204.00	2359.0
205.00	4409.0
206.00	16704.0
207.00	2584.0
208.00	769.0
208.90	300.0
209.90	253.0
210.90	578.0
214.90	215.0
215.90	387.0
216.90	5554.0
218.00	908.0
221.00	1610.0
222.00	738.0
223.00	1194.0
224.00	10225.0
225.00	2613.0
226.10	322.0
227.00	4972.0
228.00	719.0
228.90	959.0
231.00	349.0
233.90	384.0
234.90	270.0
235.90	237.0
237.00	445.0
238.90	245.0
239.90	177.0
240.90	215.0
242.00	344.0
242.90	570.0
244.00	6651.0
245.10	970.0
245.90	1695.0
247.00	361.0
248.90	182.0
252.90	181.0
254.20	517.0
255.00	36888.0
256.00	5529.0
256.90	632.0
257.90	2526.0
259.00	521.0
265.00	1241.0
265.90	176.0
272.00	162.0
272.90	921.0
274.00	3472.0
275.00	15629.0
276.00	2242.0
277.00	1576.0
278.00	430.0
283.00	217.0
285.00	348.0
293.00	321.0
296.00	5895.0
297.00	699.0
303.10	709.0
304.00	175.0
314.10	220.0
314.90	738.0
316.00	373.0
321.00	180.0
323.00	1452.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

323.90	226.0
326.90	293.0
331.90	198.0
334.00	992.0
335.00	158.0
340.90	163.0
346.00	445.0
352.00	509.0
352.90	176.0
354.00	335.0
364.90	2396.0
365.80	546.0
372.00	604.0
373.10	294.0
382.80	201.0
401.90	313.0
403.00	400.0
421.00	226.0
421.90	257.0
423.00	2455.0
424.00	965.0
441.00	6071.0
442.00	44336.0
443.00	8443.0
443.90	744.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

m/z	Abundance
38.00	246.0
39.00	1855.0
44.00	189.0
49.00	285.0
50.10	7220.0
51.10	24568.0
52.00	1090.0
55.00	205.0
56.00	826.0
57.00	1589.0
61.00	282.0
62.00	341.0
63.00	1041.0
64.00	164.0
65.00	323.0
68.10	420.0
69.00	21928.0
73.00	338.0
74.00	2393.0
75.00	3921.0
76.00	1140.0
77.00	22008.0
78.10	1603.0
79.00	1674.0
79.90	1589.0
81.00	2013.0
82.00	540.0
82.80	414.0
84.90	293.0
86.00	808.0
87.10	261.0
91.00	609.0
92.00	563.0
93.00	3968.0
94.00	244.0
98.00	3105.0
99.00	2197.0
100.90	1072.0
102.90	267.0
103.90	501.0
105.00	748.0
106.10	205.0
107.00	7955.0
108.00	1157.0
109.00	328.0
110.00	11917.0
111.00	2102.0
112.00	208.0
116.00	559.0
117.00	9309.0
117.90	734.0
122.00	652.0
122.90	837.0
124.00	548.0
125.00	213.0
126.20	260.0
127.00	21824.0
127.90	1638.0
129.00	10396.0
129.90	785.0
130.90	219.0
133.90	387.0
134.90	669.0
136.00	314.0
136.90	324.0
140.90	1619.0
141.90	541.0
142.90	382.0
146.00	315.0
146.90	977.0
148.00	2069.0
149.00	410.0
150.00	155.0
152.90	419.0
153.80	417.0
155.00	905.0
156.00	1027.0
157.10	239.0
158.00	327.0
158.90	180.0
159.90	374.0
161.00	699.0
162.00	171.0
164.90	662.0
165.90	427.0
167.00	3052.0
168.00	1414.0
168.90	301.0
171.00	187.0
171.80	151.0
172.90	327.0
174.00	585.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

175.00	1387.0
176.00	249.0
176.90	602.0
177.80	175.0
179.00	2405.0
180.00	1504.0
180.90	880.0
183.90	227.0
185.00	1140.0
186.00	7161.0
187.00	2125.0
187.90	254.0
188.90	552.0
190.80	188.0
192.00	879.0
193.00	882.0
193.90	157.0
195.90	1147.0
198.00	43776.0
198.90	3209.0
199.90	294.0
203.00	533.0
204.00	2041.0
205.00	3719.0
206.00	13063.0
207.00	2328.0
208.00	604.0
209.00	359.0
209.90	269.0
210.90	637.0
215.90	381.0
217.00	5202.0
218.00	678.0
220.20	150.0
220.90	1382.0
221.80	600.0
222.90	933.0
224.00	8858.0
225.00	2265.0
226.00	295.0
227.00	4092.0
228.00	545.0
228.90	838.0
230.90	283.0
233.90	171.0
234.90	312.0
235.90	274.0
237.00	349.0
238.90	232.0
240.80	199.0
241.90	644.0
242.90	547.0
244.00	6318.0
245.00	782.0
245.90	1556.0
246.90	276.0
248.90	206.0
254.20	316.0
255.00	33192.0
255.90	4955.0
256.90	471.0
258.00	2248.0
258.90	493.0
264.90	890.0
273.00	964.0
274.00	2890.0
275.00	13995.0
275.90	1757.0
277.00	1647.0
278.00	210.0
285.00	272.0
293.00	219.0
296.00	5381.0
296.90	671.0
303.00	517.0
304.00	257.0
308.00	168.0
314.00	311.0
315.00	626.0
316.00	210.0
323.00	1365.0
324.00	208.0
326.90	308.0
328.00	237.0
332.90	195.0
333.90	987.0
335.00	235.0
340.90	193.0
345.90	386.0
351.90	480.0
353.10	247.0
354.10	521.0
365.00	2816.0
365.80	263.0
372.00	589.0
373.00	235.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88

383.00	297.0
401.80	267.0
402.90	434.0
420.90	260.0
421.80	293.0
422.90	2304.0
424.00	733.0
441.00	8056.0
442.00	47984.0
443.00	9524.0
444.00	945.0

Instrument :
BNA_M
ClientSampleId :
DFTPP88