

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
 Data File : BM033259.D
 Acq On : 26 Nov 2021 11:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 12:20:35 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	120	0.00
2	1,4-Dioxane	40.000	34.905	12.7	111	0.00
3	Pyridine	40.000	39.242	1.9	118	0.00
4	n-Nitrosodimethylamine	40.000	37.441	6.4	116	0.00
5 S	2-Fluorophenol	80.000	76.897	3.9	121	0.00
6	Aniline	40.000	39.192	2.0	121	0.00
7 S	Phenol-d6	80.000	79.439	0.7	124	0.00
8	2-Chlorophenol	40.000	39.188	2.0	123	0.00
9	Benzaldehyde	40.000	41.409	-3.5	125	0.00
10 C	Phenol	40.000	39.302	1.7	124	0.01
11	bis(2-Chloroethyl)ether	40.000	38.724	3.2	120	0.00
12	1,3-Dichlorobenzene	40.000	37.455	6.4	119	0.00
13 C	1,4-Dichlorobenzene	40.000	37.344	6.6	119	0.00
14	1,2-Dichlorobenzene	40.000	38.117	4.7	120	0.00
15	Benzyl Alcohol	40.000	40.209	-0.5	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.494	6.3	118	0.00
17	2-Methylphenol	40.000	39.848	0.4	123	0.00
18	Hexachloroethane	40.000	38.103	4.7	118	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.637	0.9	122	0.00
20	3+4-Methylphenols	40.000	40.157	-0.4	123	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	118	0.00
22	Acetophenone	40.000	38.263	4.3	121	0.00
23 S	Nitrobenzene-d5	80.000	77.753	2.8	120	0.00
24	Nitrobenzene	40.000	38.512	3.7	121	0.00
25	Isophorone	40.000	39.609	1.0	122	0.00
26 C	2-Nitrophenol	40.000	42.972	-7.4	129	0.00
27	2,4-Dimethylphenol	40.000	39.383	1.5	122	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.602	3.5	119	0.00
29 C	2,4-Dichlorophenol	40.000	40.022	-0.1	122	0.00
30	1,2,4-Trichlorobenzene	40.000	37.139	7.2	117	0.00
31	Naphthalene	40.000	37.905	5.2	119	0.00
32	Benzoic acid	40.000	39.417	1.5	122	0.01
33	4-Chloroaniline	40.000	39.708	0.7	121	0.00
34 C	Hexachlorobutadiene	40.000	36.368	9.1	115	0.00
35	Caprolactam	40.000	43.735	-9.3	129	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.096	-0.2	121	0.00
37	2-Methylnaphthalene	40.000	37.507	6.2	117	0.00
38	1-Methylnaphthalene	40.000	38.058	4.9	118	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.809	8.0	117	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.152	-7.9	127	0.00
42 S	2,4,6-Tribromophenol	80.000	77.870	2.7	117	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.290	-0.7	122	0.00
44	2,4,5-Trichlorophenol	40.000	39.409	1.5	121	0.00
45 S	2-Fluorobiphenyl	80.000	74.126	7.3	116	0.00
46	1,1'-Biphenyl	40.000	37.525	6.2	118	0.00
47	2-Chloronaphthalene	40.000	37.692	5.8	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
 Data File : BM033259.D
 Acq On : 26 Nov 2021 11:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 12:20:35 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.367	-3.4	122	0.00
49	Acenaphthylene	40.000	38.716	3.2	118	0.00
50	Dimethylphthalate	40.000	37.835	5.4	118	0.00
51	2,6-Dinitrotoluene	40.000	39.543	1.1	118	0.00
52 C	Acenaphthene	40.000	37.783	5.5	118	0.00
53	3-Nitroaniline	40.000	41.554	-3.9	120	0.00
54 P	2,4-Dinitrophenol	40.000	39.159	2.1	115	0.00
55	Dibenzofuran	40.000	37.152	7.1	116	0.00
56 P	4-Nitrophenol	40.000	42.812	-7.0	121	0.01
57	2,4-Dinitrotoluene	40.000	40.925	-2.3	120	0.00
58	Fluorene	40.000	37.328	6.7	116	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.532	3.7	118	0.00
60	Diethylphthalate	40.000	38.319	4.2	119	0.00
61	4-Chlorophenyl-phenylether	40.000	36.887	7.8	115	0.00
62	4-Nitroaniline	40.000	42.926	-7.3	124	0.00
63	Azobenzene	40.000	38.898	2.8	118	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.942	7.6	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.212	2.0	118	0.00
67	4-Bromophenyl-phenylether	40.000	37.969	5.1	115	0.00
68	Hexachlorobenzene	40.000	37.982	5.0	116	0.00
69	Atrazine	40.000	41.927	-4.8	120	0.00
70 C	Pentachlorophenol	40.000	40.280	-0.7	114	0.00
71	Phenanthrene	40.000	37.636	5.9	115	0.00
72	Anthracene	40.000	39.192	2.0	118	0.00
73	Carbazole	40.000	38.845	2.9	117	0.00
74	Di-n-butylphthalate	40.000	40.975	-2.4	122	0.00
75 C	Fluoranthene	40.000	37.516	6.2	114	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	112	0.00
77	Benzydine	40.000	62.125	-55.3#	159	0.00
78	Pyrene	40.000	38.909	2.7	113	0.00
79 S	Terphenyl-d14	80.000	77.185	3.5	113	0.00
80	Butylbenzylphthalate	40.000	44.217	-10.5	123	0.00
81	Benzo(a)anthracene	40.000	38.072	4.8	114	0.00
82	3,3'-Dichlorobenzidine	40.000	40.136	-0.3	115	0.00
83	Chrysene	40.000	37.551	6.1	112	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.674	-9.2	123	0.00
85 c	Di-n-octyl phthalate	40.000	43.197	-8.0	124	0.00
86 I	Perylene-d12	20.000	20.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.842	5.4	113	0.00
88	Benzo(b)fluoranthene	40.000	38.169	4.6	111	0.00
89	Benzo(k)fluoranthene	40.000	36.843	7.9	111	0.00
90 C	Benzo(a)pyrene	40.000	38.295	4.3	113	0.00
91	Dibenzo(a,h)anthracene	40.000	38.265	4.3	115	0.00
92	Benzo(g,h,i)perylene	40.000	37.833	5.4	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
Data File : BM033259.D
Acq On : 26 Nov 2021 11:33
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 12:20:35 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 24 15:16:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0