

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
 Data File : BM033263.D
 Acq On : 26 Nov 2021 14:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 14:50:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	36.024	9.9	108	0.00
3	Pyridine	40.000	39.910	0.2	113	0.00
4	n-Nitrosodimethylamine	40.000	38.730	3.2	113	0.00
5 S	2-Fluorophenol	80.000	78.819	1.5	117	0.00
6	Aniline	40.000	39.531	1.2	114	0.00
7 S	Phenol-d6	80.000	81.137	-1.4	119	0.00
8	2-Chlorophenol	40.000	39.299	1.8	116	0.00
9	Benzaldehyde	40.000	42.165	-5.4	120	0.00
10 C	Phenol	40.000	40.351	-0.9	119	0.00
11	bis(2-Chloroethyl)ether	40.000	39.003	2.5	114	0.00
12	1,3-Dichlorobenzene	40.000	37.413	6.5	112	0.00
13 C	1,4-Dichlorobenzene	40.000	37.460	6.3	112	0.00
14	1,2-Dichlorobenzene	40.000	37.870	5.3	112	0.00
15	Benzyl Alcohol	40.000	40.717	-1.8	117	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.161	4.6	113	0.00
17	2-Methylphenol	40.000	40.587	-1.5	118	0.00
18	Hexachloroethane	40.000	38.781	3.0	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.471	-1.2	118	0.00
20	3+4-Methylphenols	40.000	41.125	-2.8	118	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	112	0.00
22	Acetophenone	40.000	38.343	4.1	116	0.00
23 S	Nitrobenzene-d5	80.000	78.041	2.4	115	0.00
24	Nitrobenzene	40.000	38.621	3.4	116	0.00
25	Isophorone	40.000	39.449	1.4	116	0.00
26 C	2-Nitrophenol	40.000	42.402	-6.0	121	0.00
27	2,4-Dimethylphenol	40.000	39.353	1.6	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.831	2.9	115	0.00
29 C	2,4-Dichlorophenol	40.000	40.194	-0.5	117	0.00
30	1,2,4-Trichlorobenzene	40.000	37.016	7.5	111	0.00
31	Naphthalene	40.000	37.669	5.8	113	0.00
32	Benzoic acid	40.000	41.828	-4.6	124	0.01
33	4-Chloroaniline	40.000	39.336	1.7	114	0.00
34 C	Hexachlorobutadiene	40.000	36.717	8.2	111	0.00
35	Caprolactam	40.000	43.772	-9.4	124	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.243	-0.6	116	0.00
37	2-Methylnaphthalene	40.000	38.088	4.8	114	0.00
38	1-Methylnaphthalene	40.000	37.941	5.1	113	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	113	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.931	7.7	114	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.800	-14.5	131	0.00
42 S	2,4,6-Tribromophenol	80.000	78.577	1.8	114	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.977	0.1	116	0.00
44	2,4,5-Trichlorophenol	40.000	39.680	0.8	117	0.00
45 S	2-Fluorobiphenyl	80.000	74.598	6.8	113	0.00
46	1,1'-Biphenyl	40.000	37.846	5.4	115	0.00
47	2-Chloronaphthalene	40.000	38.027	4.9	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
 Data File : BM033263.D
 Acq On : 26 Nov 2021 14:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 14:50:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.235	-5.6	120	0.00
49	Acenaphthylene	40.000	38.488	3.8	113	0.00
50	Dimethylphthalate	40.000	37.879	5.3	114	0.00
51	2,6-Dinitrotoluene	40.000	39.747	0.6	114	0.00
52 C	Acenaphthene	40.000	37.574	6.1	113	0.00
53	3-Nitroaniline	40.000	41.834	-4.6	117	0.00
54 P	2,4-Dinitrophenol	40.000	39.720	0.7	113	0.00
55	Dibenzofuran	40.000	37.219	7.0	112	0.00
56 P	4-Nitrophenol	40.000	43.049	-7.6	118	0.01
57	2,4-Dinitrotoluene	40.000	41.269	-3.2	117	0.00
58	Fluorene	40.000	37.782	5.5	114	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.458	1.4	117	0.00
60	Diethylphthalate	40.000	38.560	3.6	115	0.00
61	4-Chlorophenyl-phenylether	40.000	37.363	6.6	113	0.00
62	4-Nitroaniline	40.000	43.352	-8.4	121	0.00
63	Azobenzene	40.000	39.297	1.8	115	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	114	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.264	6.8	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.560	3.6	115	0.00
67	4-Bromophenyl-phenylether	40.000	37.432	6.4	113	0.00
68	Hexachlorobenzene	40.000	36.696	8.3	111	0.00
69	Atrazine	40.000	40.509	-1.3	115	0.00
70 C	Pentachlorophenol	40.000	40.441	-1.1	114	0.00
71	Phenanthrene	40.000	37.408	6.5	114	0.00
72	Anthracene	40.000	37.963	5.1	113	0.00
73	Carbazole	40.000	38.599	3.5	116	0.00
74	Di-n-butylphthalate	40.000	40.658	-1.6	120	0.00
75 C	Fluoranthene	40.000	37.500	6.3	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzydine	40.000	57.094	-42.7#	145	0.00
78	Pyrene	40.000	38.676	3.3	112	0.00
79 S	Terphenyl-d14	80.000	77.137	3.6	112	0.00
80	Butylbenzylphthalate	40.000	43.393	-8.5	120	0.00
81	Benzo(a)anthracene	40.000	38.049	4.9	113	0.00
82	3,3'-Dichlorobenzidine	40.000	40.200	-0.5	115	0.00
83	Chrysene	40.000	37.281	6.8	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.098	-7.7	121	0.00
85 c	Di-n-octyl phthalate	40.000	43.151	-7.9	123	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.308	6.7	110	0.00
88	Benzo(b)fluoranthene	40.000	38.630	3.4	110	0.00
89	Benzo(k)fluoranthene	40.000	36.770	8.1	109	0.00
90 C	Benzo(a)pyrene	40.000	37.920	5.2	110	0.00
91	Dibenzo(a,h)anthracene	40.000	37.705	5.7	111	0.00
92	Benzo(g,h,i)perylene	40.000	37.549	6.1	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
Data File : BM033263.D
Acq On : 26 Nov 2021 14:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 14:50:10 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 24 15:16:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0