

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
 Data File : BM033265.D
 Acq On : 26 Nov 2021 15:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 17:27:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	35.249	11.9	108	0.00
3	Pyridine	40.000	39.050	2.4	113	0.00
4	n-Nitrosodimethylamine	40.000	38.741	3.1	116	0.00
5 S	2-Fluorophenol	80.000	76.107	4.9	115	0.00
6	Aniline	40.000	38.601	3.5	114	0.00
7 S	Phenol-d6	80.000	78.816	1.5	119	0.00
8	2-Chlorophenol	40.000	38.408	4.0	116	0.00
9	Benzaldehyde	40.000	41.330	-3.3	120	0.00
10 C	Phenol	40.000	39.557	1.1	120	0.00
11	bis(2-Chloroethyl)ether	40.000	38.941	2.6	116	0.00
12	1,3-Dichlorobenzene	40.000	37.138	7.2	114	0.00
13 C	1,4-Dichlorobenzene	40.000	37.253	6.9	114	0.00
14	1,2-Dichlorobenzene	40.000	37.390	6.5	113	0.00
15	Benzyl Alcohol	40.000	40.377	-0.9	118	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.018	5.0	115	0.00
17	2-Methylphenol	40.000	39.779	0.6	118	0.00
18	Hexachloroethane	40.000	38.391	4.0	115	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.821	0.4	118	0.00
20	3+4-Methylphenols	40.000	40.400	-1.0	119	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	38.045	4.9	117	0.00
23 S	Nitrobenzene-d5	80.000	78.323	2.1	118	0.00
24	Nitrobenzene	40.000	38.166	4.6	116	0.00
25	Isophorone	40.000	39.858	0.4	119	0.00
26 C	2-Nitrophenol	40.000	42.433	-6.1	123	0.00
27	2,4-Dimethylphenol	40.000	39.431	1.4	118	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.778	3.1	116	0.00
29 C	2,4-Dichlorophenol	40.000	40.100	-0.3	118	0.00
30	1,2,4-Trichlorobenzene	40.000	37.422	6.4	114	0.00
31	Naphthalene	40.000	37.780	5.5	115	0.00
32	Benzoic acid	40.000	41.773	-4.4	126	0.01
33	4-Chloroaniline	40.000	39.920	0.2	118	0.00
34 C	Hexachlorobutadiene	40.000	36.256	9.4	112	0.00
35	Caprolactam	40.000	42.760	-6.9	123	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.769	-1.9	119	0.00
37	2-Methylnaphthalene	40.000	38.054	4.9	115	0.00
38	1-Methylnaphthalene	40.000	38.484	3.8	116	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.469	8.8	116	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.652	-14.1	134	0.00
42 S	2,4,6-Tribromophenol	80.000	78.539	1.8	117	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.155	-0.4	121	0.00
44	2,4,5-Trichlorophenol	40.000	39.281	1.8	120	0.00
45 S	2-Fluorobiphenyl	80.000	74.091	7.4	116	0.00
46	1,1'-Biphenyl	40.000	37.429	6.4	117	0.00
47	2-Chloronaphthalene	40.000	37.656	5.9	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
 Data File : BM033265.D
 Acq On : 26 Nov 2021 15:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 17:27:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 24 15:16:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.410	-6.0	124	0.00
49	Acenaphthylene	40.000	38.685	3.3	117	0.00
50	Dimethylphthalate	40.000	37.935	5.2	118	0.00
51	2,6-Dinitrotoluene	40.000	39.774	0.6	118	0.00
52 C	Acenaphthene	40.000	37.961	5.1	118	0.00
53	3-Nitroaniline	40.000	42.357	-5.9	122	0.00
54 P	2,4-Dinitrophenol	40.000	39.915	0.2	117	0.00
55	Dibenzofuran	40.000	37.769	5.6	117	0.00
56 P	4-Nitrophenol	40.000	43.762	-9.4	123	0.01
57	2,4-Dinitrotoluene	40.000	41.714	-4.3	122	0.00
58	Fluorene	40.000	38.225	4.4	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.798	0.5	121	0.00
60	Diethylphthalate	40.000	39.040	2.4	120	0.00
61	4-Chlorophenyl-phenylether	40.000	37.640	5.9	117	0.00
62	4-Nitroaniline	40.000	43.025	-7.6	124	0.00
63	Azobenzene	40.000	39.935	0.2	121	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.209	4.5	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.325	4.2	118	0.00
67	4-Bromophenyl-phenylether	40.000	37.712	5.7	117	0.00
68	Hexachlorobenzene	40.000	37.112	7.2	116	0.00
69	Atrazine	40.000	41.381	-3.5	121	0.00
70 C	Pentachlorophenol	40.000	41.567	-3.9	121	0.00
71	Phenanthrene	40.000	37.768	5.6	119	0.00
72	Anthracene	40.000	38.739	3.2	119	0.00
73	Carbazole	40.000	38.871	2.8	120	0.00
74	Di-n-butylphthalate	40.000	41.658	-4.1	127	0.00
75 C	Fluoranthene	40.000	37.989	5.0	119	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	57.059	-42.6#	152	0.00
78	Pyrene	40.000	39.355	1.6	119	0.00
79 S	Terphenyl-d14	80.000	77.112	3.6	117	0.00
80	Butylbenzylphthalate	40.000	43.953	-9.9	128	0.00
81	Benzo(a)anthracene	40.000	38.398	4.0	119	0.00
82	3,3'-Dichlorobenzidine	40.000	39.763	0.6	119	0.00
83	Chrysene	40.000	37.608	6.0	116	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.349	-10.9	130	0.00
85 c	Di-n-octyl phthalate	40.000	43.666	-9.2	130	0.00
86 I	Perylene-d12	20.000	20.000	0.0	117	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.367	6.6	116	0.00
88	Benzo(b)fluoranthene	40.000	38.393	4.0	116	0.00
89	Benzo(k)fluoranthene	40.000	37.061	7.3	116	0.00
90 C	Benzo(a)pyrene	40.000	37.924	5.2	116	0.00
91	Dibenzo(a,h)anthracene	40.000	37.369	6.6	116	0.00
92	Benzo(g,h,i)perylene	40.000	37.462	6.3	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112621\
Data File : BM033265.D
Acq On : 26 Nov 2021 15:17
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 17:27:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM112421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 24 15:16:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0