


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM120320\
 Data File : BM028012.D
 Acq On : 03 Dec 2020 10:52
 Operator : JU/CG
 Sample : DFTPP99
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP099

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM120320.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 04 11:00:53 2020



AutoFind: Scans 2523, 2524, 2525; Background Corrected with Scan 2517

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.5	129960	PASS
68	69	0.00	2	1.7	2107	PASS
69	198	0.00	100	46.5	121965	PASS
70	69	0.00	2	0.5	552	PASS
127	198	10	80	53.2	139565	PASS
197	198	0.00	2	0.7	1910	PASS
198	198	100	100	100.0	262314	PASS
199	198	5	9	7.0	18286	PASS
275	198	10	60	27.9	73282	PASS
365	198	1	100	3.4	8918	PASS
441	198	0.01	100	13.3	34925	PASS
442	198	50	100	90.2	236608	PASS
443	442	15	24	19.3	45709	PASS

m/z	Abundance
39.90	703.0
43.90	553.0
48.90	263.0
51.10	293.0
73.10	1060.0
76.90	234.0
95.90	231.0
132.90	264.0
147.00	405.0
190.90	357.0
193.10	364.0
207.00	3001.0
207.90	536.0
208.90	263.0
221.00	179.0
266.90	165.0
281.10	772.0
281.90	274.0
340.90	175.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

m/z	Abundance
36.00	201.0
37.00	632.0
38.10	1576.0
39.10	9765.0
40.00	1328.0
41.00	329.0
43.00	173.0
44.10	523.0
45.00	335.0
49.00	1259.0
50.10	31632.0
51.10	122408.0
52.10	5760.0
52.90	275.0
55.00	641.0
56.00	3704.0
57.00	8672.0
57.90	527.0
59.00	225.0
61.00	1425.0
62.00	1463.0
63.00	4737.0
64.00	586.0
65.00	2258.0
66.00	178.0
67.10	190.0
68.00	1998.0
69.00	112712.0
70.00	498.0
73.00	2294.0
74.10	11111.0
75.00	18152.0
76.10	5914.0
77.10	120520.0
78.00	8276.0
79.00	8292.0
80.00	6838.0
81.00	8448.0
82.00	2121.0
83.10	1868.0
84.00	297.0
85.00	1336.0
86.00	2839.0
87.10	969.0
88.00	434.0
89.00	257.0
91.00	2050.0
92.00	2806.0
93.00	14506.0
94.00	978.0
96.00	1124.0
96.90	426.0
98.00	10771.0
99.00	8643.0
100.00	845.0
101.00	4829.0
102.10	339.0
103.00	1844.0
104.00	3439.0
105.00	3144.0
106.00	924.0
107.00	35112.0
108.00	5385.0
109.00	1051.0
110.00	60488.0
111.00	10387.0
112.00	1284.0
112.90	324.0
114.90	248.0
116.00	2003.0
117.00	29600.0
118.00	2434.0
119.00	546.0
119.90	368.0
122.00	2406.0
123.00	3820.0
124.00	1292.0
125.00	1486.0
127.00	120776.0
128.00	9468.0
129.00	46664.0
130.00	4159.0
131.00	984.0
132.00	486.0
133.10	446.0
134.00	1424.0
135.00	3829.0
136.00	1423.0
137.00	2265.0
138.00	467.0
138.80	223.0
140.00	767.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

141.00	6258.0
142.00	1706.0
143.10	1310.0
144.10	414.0
145.00	348.0
146.00	1029.0
147.00	3565.0
148.00	7570.0
149.00	1961.0
149.90	436.0
151.10	829.0
151.80	584.0
152.90	1836.0
154.00	1152.0
155.00	3233.0
156.10	4836.0
156.90	966.0
158.00	1127.0
159.00	825.0
160.00	1848.0
161.00	2879.0
162.00	654.0
163.00	221.0
163.90	262.0
165.00	2191.0
166.10	1664.0
167.00	10964.0
168.00	7363.0
169.00	1128.0
169.90	453.0
170.90	517.0
172.00	1073.0
172.90	1304.0
174.00	2397.0
175.10	4335.0
176.00	1097.0
176.90	1972.0
178.00	801.0
179.00	7923.0
180.00	6344.0
181.00	2822.0
182.00	388.0
183.00	318.0
184.00	731.0
185.00	4249.0
186.00	29464.0
187.00	8316.0
188.10	1252.0
189.00	2232.0
190.00	385.0
191.00	1180.0
192.00	3042.0
193.00	3283.0
194.00	667.0
195.00	419.0
196.00	6280.0
197.10	1555.0
198.00	212864.0
199.00	15850.0
200.00	1155.0
201.60	1202.0
202.90	1852.0
204.00	8412.0
205.00	12856.0
206.00	54168.0
207.00	10310.0
208.00	2658.0
209.00	1037.0
210.20	925.0
211.10	2416.0
211.90	573.0
212.80	293.0
215.00	699.0
216.00	1129.0
217.00	15860.0
218.00	2224.0
219.00	221.0
220.30	439.0
221.00	10661.0
221.90	2034.0
223.00	3531.0
224.00	29592.0
225.00	7953.0
226.00	1223.0
227.00	12460.0
228.10	1754.0
229.00	2434.0
230.00	482.0
230.90	1124.0
232.10	227.0
232.90	272.0
233.90	1014.0
235.00	882.0
236.00	530.0
237.00	2219.0
238.00	458.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

238.90	376.0
239.90	589.0
241.00	777.0
242.00	1880.0
243.10	1813.0
244.10	23360.0
245.00	3150.0
246.00	3838.0
247.00	838.0
247.80	242.0
249.00	755.0
250.10	224.0
251.10	151.0
252.00	206.0
252.90	696.0
254.20	1008.0
255.00	120040.0
256.00	18120.0
257.00	2041.0
258.00	6578.0
259.00	1212.0
259.90	243.0
261.00	323.0
264.10	244.0
265.00	2619.0
265.90	254.0
270.00	212.0
271.00	296.0
272.00	211.0
273.00	3315.0
274.00	8882.0
275.00	53864.0
276.00	7585.0
277.00	4242.0
278.00	959.0
279.00	205.0
281.00	1032.0
282.00	270.0
283.00	710.0
283.90	278.0
285.00	709.0
289.00	173.0
291.90	170.0
292.90	882.0
294.10	257.0
295.00	439.0
296.00	14472.0
297.00	1903.0
301.00	216.0
302.10	328.0
303.00	1681.0
304.00	567.0
307.90	217.0
313.00	204.0
314.10	591.0
314.90	1542.0
315.90	957.0
321.00	437.0
322.00	303.0
323.00	3860.0
324.10	626.0
327.00	867.0
328.00	472.0
332.10	172.0
332.90	421.0
334.00	2931.0
335.00	1075.0
336.00	228.0
338.90	203.0
341.00	666.0
342.10	205.0
345.90	881.0
352.00	1327.0
353.00	1139.0
354.00	1420.0
355.00	524.0
365.00	5592.0
366.00	1150.0
371.00	248.0
372.00	2178.0
373.00	554.0
382.90	601.0
390.00	297.0
402.10	886.0
403.00	1348.0
404.00	547.0
421.00	1145.0
422.00	1377.0
423.00	8087.0
424.00	3735.0
424.90	527.0
441.00	21560.0
442.00	140992.0
443.00	26984.0
444.00	2823.0
445.00	162.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

m/z	Abundance
37.00	639.0
38.10	1802.0
39.10	11048.0
40.00	1144.0
41.00	357.0
44.00	418.0
45.00	358.0
49.00	1298.0
50.10	39968.0
51.10	155840.0
52.10	8234.0
53.00	365.0
55.10	792.0
56.00	4626.0
57.00	10410.0
58.00	494.0
59.90	152.0
61.00	1879.0
62.00	1883.0
63.00	5959.0
64.10	974.0
65.10	2868.0
68.10	2606.0
69.00	145920.0
70.00	711.0
73.00	1967.0
74.00	14869.0
75.00	23344.0
76.10	7886.0
77.10	155968.0
78.10	10481.0
79.00	10966.0
80.00	8043.0
81.00	11724.0
82.00	2939.0
83.00	2636.0
83.80	280.0
85.00	1709.0
86.00	3787.0
87.00	1577.0
88.00	597.0
88.80	365.0
91.00	2423.0
92.00	3060.0
93.00	19096.0
94.00	1192.0
95.00	288.0
96.00	1086.0
97.00	310.0
98.00	15264.0
99.00	11410.0
100.00	1185.0
101.00	6405.0
102.00	379.0
103.00	2479.0
104.00	4349.0
105.00	3848.0
106.00	1364.0
107.00	46296.0
108.00	8078.0
109.00	1392.0
110.00	84552.0
111.00	12696.0
112.00	1523.0
112.90	580.0
114.10	217.0
114.90	260.0
116.10	2315.0
117.00	41544.0
118.00	3143.0
119.00	480.0
120.00	560.0
120.90	214.0
122.00	3457.0
123.00	4231.0
124.00	2041.0
125.00	1569.0
126.20	1266.0
127.00	167360.0
128.00	12802.0
129.00	65776.0
130.00	5610.0
131.00	994.0
131.90	580.0
133.00	655.0
134.00	2143.0
135.00	5198.0
136.00	1957.0
137.10	2907.0
137.90	520.0
138.90	291.0
139.90	867.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

141.00	8703.0
142.00	2453.0
143.00	1705.0
143.90	400.0
144.90	627.0
146.00	1466.0
147.00	4557.0
148.00	11021.0
149.00	1992.0
149.90	664.0
151.10	1033.0
151.70	576.0
153.00	2940.0
154.00	2006.0
155.00	4577.0
156.00	6652.0
157.00	1420.0
158.00	1692.0
158.90	936.0
160.00	2782.0
161.00	3960.0
161.90	1052.0
162.90	389.0
163.90	493.0
165.00	3408.0
166.00	2491.0
167.00	17432.0
168.00	9697.0
169.00	1588.0
169.90	495.0
171.10	771.0
171.90	1798.0
173.00	2015.0
174.00	3663.0
175.10	6391.0
176.00	1924.0
177.00	2860.0
178.00	826.0
179.00	11619.0
180.00	8932.0
181.00	4097.0
182.00	720.0
182.90	303.0
184.00	1456.0
185.00	6309.0
186.00	44248.0
187.00	12846.0
188.00	1620.0
189.00	3047.0
190.00	580.0
191.00	1545.0
192.00	3830.0
193.00	4440.0
194.00	1102.0
195.00	743.0
196.00	10451.0
197.10	2446.0
198.00	310784.0
199.00	21896.0
200.00	2104.0
200.80	474.0
201.50	1976.0
202.20	296.0
202.90	2309.0
204.00	12394.0
205.00	19248.0
206.00	85664.0
207.10	14473.0
208.00	3854.0
209.00	1309.0
210.20	1543.0
211.00	3143.0
211.90	679.0
212.80	219.0
215.00	962.0
216.00	2004.0
217.00	23288.0
218.00	2704.0
219.00	192.0
221.00	15142.0
223.00	5414.0
224.00	44840.0
225.00	12218.0
226.00	1333.0
227.00	19544.0
228.00	2522.0
229.00	4055.0
229.90	650.0
231.00	2022.0
232.00	220.0
233.00	384.0
234.00	1153.0
235.00	1456.0
236.00	1018.0
237.00	2044.0
237.90	339.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

239.00	612.0
240.00	621.0
241.00	1165.0
242.10	2244.0
243.10	2791.0
244.10	35592.0
245.00	5170.0
246.00	6635.0
247.00	1528.0
247.90	375.0
249.00	1223.0
249.90	256.0
251.00	263.0
252.10	378.0
252.90	760.0
254.20	1970.0
255.00	187712.0
256.00	26616.0
257.00	2156.0
258.00	10149.0
259.00	1875.0
260.00	293.0
261.00	397.0
264.00	418.0
265.00	4009.0
266.00	533.0
269.90	214.0
271.10	378.0
272.00	398.0
273.00	5549.0
274.00	14701.0
275.00	87336.0
276.00	11622.0
277.00	6973.0
278.00	1281.0
279.00	427.0
281.10	1133.0
281.80	286.0
283.00	938.0
283.90	442.0
285.10	1552.0
285.90	239.0
290.10	252.0
290.90	241.0
292.00	354.0
293.10	1488.0
293.90	500.0
295.10	638.0
296.00	23952.0
297.00	2999.0
297.90	229.0
300.90	362.0
302.00	582.0
303.00	2356.0
303.90	837.0
308.10	373.0
309.00	238.0
310.10	211.0
313.10	182.0
314.00	1133.0
314.90	2692.0
316.00	1271.0
317.00	316.0
320.90	758.0
322.20	446.0
323.00	7701.0
324.00	1569.0
325.00	171.0
325.90	210.0
327.00	1266.0
328.00	767.0
329.00	152.0
332.00	640.0
333.10	649.0
334.00	4546.0
335.00	1219.0
336.00	307.0
339.00	269.0
339.90	206.0
341.10	949.0
341.90	435.0
346.00	1536.0
346.90	320.0
352.10	2056.0
353.00	1573.0
354.00	2482.0
355.00	834.0
358.00	160.0
358.90	185.0
364.10	181.0
365.00	10795.0
366.00	1320.0
367.00	276.0
369.90	216.0
371.00	564.0
372.00	3864.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

373.00	1325.0
374.10	151.0
383.00	1348.0
383.90	326.0
384.90	170.0
390.00	489.0
391.00	379.0
400.80	228.0
402.00	1366.0
402.90	2451.0
404.10	892.0
404.80	169.0
421.00	2564.0
422.00	2126.0
423.00	14288.0
424.00	4599.0
425.00	746.0
438.00	162.0
440.00	317.0
441.00	39528.0
442.00	271168.0
443.00	53928.0
444.00	5221.0
444.90	239.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

m/z	Abundance
37.00	501.0
38.00	1332.0
39.10	9159.0
39.90	1037.0
40.90	366.0
44.00	631.0
44.90	242.0
49.00	911.0
50.10	29704.0
51.10	112512.0
52.00	5624.0
53.00	185.0
55.00	520.0
56.00	3169.0
57.00	7410.0
58.00	222.0
59.00	202.0
61.00	1461.0
62.00	1229.0
63.00	3994.0
64.00	561.0
65.10	1844.0
66.00	150.0
67.00	226.0
68.10	1719.0
69.00	107264.0
70.00	448.0
73.00	1951.0
74.00	11070.0
75.00	17288.0
76.10	5664.0
77.10	114544.0
78.10	7819.0
79.00	8042.0
80.00	6268.0
81.00	8600.0
82.00	1869.0
83.00	1888.0
84.00	421.0
85.10	1356.0
86.00	2437.0
87.00	1093.0
88.00	307.0
88.90	228.0
91.00	1731.0
92.00	2238.0
93.00	14442.0
93.90	1043.0
95.00	232.0
95.90	836.0
96.90	303.0
98.00	10875.0
99.00	9279.0
100.00	755.0
101.00	4802.0
103.00	1864.0
104.00	3423.0
105.00	2872.0
106.00	985.0
107.00	35576.0
108.00	6007.0
109.20	1035.0
110.00	63512.0
111.00	10592.0
112.00	1162.0
112.90	557.0
115.00	167.0
116.00	2413.0
117.00	31576.0
118.00	2290.0
118.90	450.0
120.00	529.0
122.00	2547.0
123.00	3995.0
123.90	1639.0
125.00	1604.0
126.20	1105.0
127.00	130560.0
128.00	9629.0
129.00	51504.0
130.00	4103.0
131.00	709.0
131.90	620.0
132.90	632.0
134.00	1513.0
135.00	4292.0
136.00	1776.0
137.00	2622.0
138.00	472.0
139.10	296.0
140.00	603.0
141.00	6732.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

142.10	2506.0
142.90	1563.0
144.00	494.0
145.00	411.0
146.00	1114.0
147.00	3799.0
148.00	8211.0
149.00	1347.0
149.90	342.0
151.00	946.0
152.10	523.0
153.00	1864.0
154.00	1659.0
155.00	3415.0
156.00	5463.0
157.00	1161.0
158.00	1219.0
158.90	1108.0
159.90	1921.0
161.00	3089.0
162.00	1087.0
163.00	363.0
163.90	385.0
165.00	2905.0
166.00	1802.0
167.00	13421.0
168.00	7076.0
169.00	1202.0
170.00	620.0
170.90	835.0
172.00	1182.0
173.00	1751.0
174.00	2826.0
175.00	4848.0
176.00	1467.0
176.90	2580.0
177.80	841.0
179.00	11445.0
180.00	7170.0
181.00	3744.0
182.00	601.0
182.90	468.0
183.90	571.0
185.00	4770.0
186.00	35160.0
187.00	10318.0
188.00	1242.0
189.00	2365.0
189.90	476.0
190.90	1470.0
192.00	3398.0
193.00	3733.0
193.90	1122.0
195.00	531.0
196.00	8530.0
197.10	1729.0
198.00	263296.0
199.00	17112.0
200.00	1621.0
201.40	1087.0
202.90	1906.0
204.00	10489.0
205.10	16824.0
206.00	71128.0
207.00	12722.0
208.00	2952.0
209.10	1389.0
210.20	1369.0
211.00	2992.0
212.90	172.0
215.00	690.0
216.00	1566.0
217.00	20896.0
218.00	2394.0
219.00	243.0
219.80	215.0
221.00	12941.0
223.00	4428.0
224.00	40072.0
225.00	9898.0
226.00	1277.0
227.00	16382.0
228.00	2652.0
229.00	3216.0
230.00	511.0
231.10	1542.0
231.90	224.0
233.00	417.0
234.00	1039.0
235.00	1286.0
236.00	967.0
237.00	1725.0
237.90	273.0
239.00	880.0
240.00	516.0
241.00	1025.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

242.00	2138.0
243.00	2471.0
244.10	31080.0
245.10	3939.0
246.00	6476.0
247.00	1273.0
247.90	349.0
248.90	1138.0
249.90	256.0
251.00	333.0
252.00	267.0
252.90	484.0
254.10	1601.0
255.00	167552.0
256.00	23832.0
257.00	1954.0
258.00	9633.0
259.00	1741.0
260.00	267.0
260.80	298.0
263.90	257.0
265.00	3820.0
266.00	567.0
266.90	215.0
270.00	224.0
270.90	270.0
272.00	503.0
273.00	4704.0
274.00	12933.0
275.00	78648.0
276.00	9768.0
277.00	6358.0
278.00	1439.0
278.90	264.0
281.00	878.0
282.10	304.0
283.00	770.0
284.00	354.0
285.00	1026.0
286.00	202.0
289.00	184.0
289.90	311.0
291.00	217.0
291.90	272.0
293.00	1583.0
294.00	597.0
295.10	551.0
296.00	21160.0
297.00	3241.0
298.10	197.0
300.90	279.0
302.00	237.0
303.00	2948.0
304.10	276.0
308.00	350.0
309.00	285.0
309.90	245.0
313.10	196.0
314.00	984.0
314.90	2507.0
315.90	1037.0
317.00	277.0
321.00	680.0
322.00	322.0
323.00	7273.0
324.10	1048.0
325.80	189.0
326.90	1441.0
328.00	709.0
328.90	226.0
332.00	586.0
333.00	941.0
334.00	4611.0
335.00	1430.0
336.10	192.0
341.00	953.0
342.00	233.0
346.00	1491.0
346.90	319.0
352.00	2303.0
353.00	1490.0
354.00	2349.0
355.10	410.0
365.00	10368.0
366.00	1599.0
367.00	212.0
370.00	190.0
371.00	832.0
372.00	3688.0
373.10	1127.0
383.00	1311.0
384.00	399.0
389.90	568.0
391.00	442.0
392.00	361.0
401.00	534.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099

402.00	1669.0
403.00	2735.0
404.00	1099.0
405.00	164.0
414.80	150.0
421.00	2143.0
422.00	2046.0
423.00	15206.0
424.00	4464.0
425.10	362.0
440.10	400.0
441.00	43688.0
442.00	297664.0
443.00	56216.0
444.00	5394.0
444.90	454.0

Instrument :
BNA_M
ClientSampleId :
DFTPP099