

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120522\
 Data File : BM037839.D
 Acq On : 05 Dec 2022 11:55
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062

Quant Time: Dec 05 23:24:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.092	152	242044	20.000	ng/u1	0.00
20) Naphthalene-d8	10.916	136	1031880	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.721	164	604924	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.457	188	1191249	20.000	ng/u1	0.00
79) Chrysene-d12	21.627	240	1094664	20.000	ng/u1	0.00
88) Perylene-d12	24.115	264	1182476	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	47214	9.152	ng/uL	0.00
4) Pyridine-d5	3.863	84	353224	21.169	ng/u1	0.00
7) Phenol-d5	7.245	99	420476	19.282	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	254950	20.039	ng/u1	0.00
11) 2-Chlorophenol-d4	7.622	132	334465	20.286	ng/u1	0.00
15) 4-Methylphenol-d8	8.804	113	316771	18.266	ng/u1	0.00
21) Nitrobenzene-d5	9.263	128	167628	20.450	ng/u1	0.00
24) 2-Nitrophenol-d4	9.992	143	156132	18.408	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.533	165	301692	19.179	ng/u1	0.00
31) 4-Chloroaniline-d4	11.051	131	452367	18.749	ng/u1	0.00
46) Dimethylphthalate-d6	14.127	166	834596	20.092	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	1043492	20.655	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	142557	16.534	ng/u1	0.00
60) Fluorene-d10	15.710	176	751980	19.954	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.821	200	92216	14.695	ng/u1	0.00
73) Anthracene-d10	17.557	188	1128176	20.593	ng/u1	0.00
81) Pyrene-d10	19.839	212	1229906	21.691	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.950	264	1222175	20.631	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	52404	9.478	ng/uL#	72
5) Pyridine	3.887	79	357449	21.848	ng/u1	91
6) Benzaldehyde	7.228	77	223934	21.322	ng/u1	98
8) Phenol	7.275	94	437554	19.634	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.510	93	358228	20.150	ng/u1	98
12) 2-Chlorophenol	7.651	128	361701	20.382	ng/u1	97
13) 2-Methylphenol	8.540	108	334473	18.820	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.616	45	605059	18.316	ng/u1	98
16) Acetophenone	8.922	105	521400	19.060	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.910	70	254518	18.472	ng/u1	98
18) 4-Methylphenol	8.869	108	352789	18.398	ng/u1	98
19) Hexachloroethane	9.181	117	144719	21.350	ng/u1	95
22) Nitrobenzene	9.310	77	364153	21.199	ng/u1	99
23) Isophorone	9.839	82	716789	19.394	ng/u1	99
25) 2-Nitrophenol	10.022	139	179003	19.474	ng/u1	98
26) 2,4-Dimethylphenol	10.081	107	362975	20.543	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.316	93	456618	20.460	ng/u1	97
29) 2,4-Dichlorophenol	10.557	162	317156	19.933	ng/u1	96
30) Naphthalene	10.963	128	1171473	20.918	ng/u1	99
32) 4-Chloroaniline	11.075	127	463392	18.979	ng/u1	98
33) Hexachlorobutadiene	11.245	225	201916	20.487	ng/u1	99
34) Caprolactam	11.857	113	92174	15.953	ng/u1	80
35) 4-Chloro-3-methylphenol	12.186	107	309381	18.633	ng/u1	97
36) 2-Methylnaphthalene	12.563	142	758761	19.584	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120522\
 Data File : BM037839.D
 Acq On : 05 Dec 2022 11:55
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062

Quant Time: Dec 05 23:24:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.780	142	756196	19.500	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.927	216	371000	21.415	ng/ul	100
40) Hexachlorocyclopentadiene	12.904	237	174128	18.277	ng/ul	98
41) 2,4,6-Trichlorophenol	13.163	196	222429	19.666	ng/ul	100
42) 2,4,5-Trichlorophenol	13.233	196	238175	19.773	ng/ul	100
43) 1,1'-Biphenyl	13.563	154	955420	21.728	ng/ul	99
44) 2-Chloronaphthalene	13.604	162	748247	21.685	ng/ul	99
45) 2-Nitroaniline	13.810	65	196535	20.130	ng/ul	98
47) Dimethylphthalate	14.174	163	868092	20.064	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	162323	18.376	ng/ul	99
50) Acenaphthylene	14.445	152	1172391	20.931	ng/ul	99
51) 3-Nitroaniline	14.627	138	187475	17.781	ng/ul#	99
52) Acenaphthene	14.786	153	806718	20.785	ng/ul	98
53) 2,4-Dinitrophenol	14.827	184	62421	12.099	ng/ul#	85
55) 4-Nitrophenol	14.927	109	93487	17.591	ng/ul	88
56) Dibenzofuran	15.116	168	1068892	20.455	ng/ul	99
57) 2,4-Dinitrotoluene	15.080	165	226790	18.383	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.339	232	188693	17.809	ng/ul	92
59) Diethylphthalate	15.527	149	867086	19.542	ng/ul	99
61) Fluorene	15.763	166	876355	19.921	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.751	204	418409	19.950	ng/ul	98
63) 4-Nitroaniline	15.780	138	183919	17.339	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.839	198	106521	16.467	ng/ul#	99
67) N-Nitrosodiphenylamine	15.968	169	733955	21.380	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	250632	20.502	ng/ul	96
69) Hexachlorobenzene	16.762	284	292457	20.487	ng/ul	95
70) Atrazine	16.910	200	231471	19.579	ng/ul	99
71) Pentachlorophenol	17.110	266	139679	15.909	ng/ul	99
72) Phenanthrene	17.504	178	1355729	20.999	ng/ul	99
74) Anthracene	17.592	178	1377762	20.919	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.527	216	367243	22.864	ng/uL	98
76) Pentachlorobenzene	15.039	250	379619	21.724	ng/uL	99
77) Carbazole	17.862	167	1252755	20.669	ng/ul	99
78) Di-n-butylphthalate	18.409	149	1524503	20.006	ng/ul	100
80) Fluoranthene	19.503	202	1510889	21.815	ng/ul	98
82) Pyrene	19.868	202	1581577	22.116	ng/ul	99
83) Butylbenzylphthalate	20.750	149	642674	20.612	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.539	252	466824	20.357	ng/ul	98
85) Benzo(a)anthracene	21.609	228	1474937	21.059	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.521	149	945672	20.072	ng/ul	99
87) Chrysene	21.668	228	1387298	21.279	ng/ul	99
89) Di-n-octyl phthalate	22.474	149	1629800	19.152	ng/ul	100
90) Benzo(b)fluoranthene	23.350	252	1573760	20.724	ng/ul	100
91) Benzo(k)fluoranthene	23.403	252	1466910	20.024	ng/ul	100
93) Benzo(a)pyrene	24.003	252	1381914	20.985	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.715	276	1683994	22.746	ng/ul	99
95) Dibenzo(a,h)anthracene	26.732	278	1530655	22.716	ng/ul	99
96) Benzo(g,h,i)perylene	27.526	276	1506207	23.313	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120522\
 Data File : BM037839.D
 Acq On : 05 Dec 2022 11:55
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062

Quant Time: Dec 05 23:24:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.478 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.001 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion: 88 Resp: 52404
Ion Ratio Lower Upper
88 100
43 34.8 51.3 76.9#
58 77.4 79.0 118.6#



#5
Pyridine
Concen: 21.848 ng/uL
RT: 3.887 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion: 79 Resp: 357449
Ion Ratio Lower Upper
79 100
52 83.8 75.2 112.8
51 39.8 33.9 50.9









#6
Benzaldehyde
Concen: 21.322 ng/ul
RT: 7.228 min Scan# 611
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion: 77 Resp: 223934
Ion Ratio Lower Upper
77 100
105 103.5 81.2 121.8
106 99.9 79.0 118.6



#8
Phenol
Concen: 19.634 ng/ul
RT: 7.275 min Scan# 661
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion: 94 Resp: 437554
Ion Ratio Lower Upper
94 100
65 26.3 21.1 31.7
66 30.7 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.150 ng/ul
RT: 7.510 min Scan# 701
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion: 93 Resp: 358228
Ion Ratio Lower Upper
93 100
63 79.2 65.7 98.5
95 33.2 26.8 40.2



#12
2-Chlorophenol
Concen: 20.382 ng/ul
RT: 7.651 min Scan# 725
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:128 Resp: 361701
Ion Ratio Lower Upper
128 100
64 45.3 37.3 55.9
130 31.3 27.1 40.7









#13
2-Methylphenol
Concen: 18.820 ng/ul
RT: 8.540 min Scan# 876
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:108 Resp: 334473
Ion Ratio Lower Upper
108 100
107 86.9 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.316 ng/ul
RT: 8.616 min Scan# 889
Delta R.T. -0.018 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion: 45 Resp: 605059
Ion Ratio Lower Upper
45 100
77 11.3 8.2 12.4
79 8.2 6.7 10.1









#16
Acetophenone
Concen: 19.060 ng/ul
RT: 8.922 min Scan# 941
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:105 Resp: 521400
Ion Ratio Lower Upper
105 100
77 72.0 59.0 88.4
51 29.0 24.4 36.6



#17
N-Nitroso-di-n-propylamine
Concen: 18.472 ng/ul
RT: 8.910 min Scan# 939
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion: 70 Resp: 254518
Ion Ratio Lower Upper
70 100
42 67.5 55.8 83.6
101 9.0 8.0 12.0
130 22.3 18.6 27.8









#18
4-Methylphenol
Concen: 18.398 ng/ul
RT: 8.869 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:108 Resp: 352789
Ion Ratio Lower Upper
108 100
107 109.6 89.3 133.9



#19
Hexachloroethane
Concen: 21.350 ng/ul
RT: 9.181 min Scan# 985
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:117 Resp: 144719
Ion Ratio Lower Upper
117 100
201 95.9 82.3 123.5
199 59.9 49.8 74.6









#22
Nitrobenzene
Concen: 21.199 ng/ul
RT: 9.310 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion: 77 Resp: 364153
Ion Ratio Lower Upper
77 100
123 50.6 41.4 62.0
65 13.3 10.9 16.3



#23
Isophorone
Concen: 19.394 ng/ul
RT: 9.839 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion: 82 Resp: 716789
Ion Ratio Lower Upper
82 100
95 6.1 5.5 8.3
138 18.9 15.3 22.9









#25
2-Nitrophenol
Concen: 19.474 ng/ul
RT: 10.022 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:139 Resp: 179003
Ion Ratio Lower Upper
139 100
65 45.7 36.6 54.8
109 23.7 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 20.543 ng/ul
RT: 10.081 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:107 Resp: 362975
Ion Ratio Lower Upper
107 100
121 54.1 41.8 62.8
122 91.7 73.1 109.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.460 ng/ul
 RT: 10.316 min Scan# 1178
 Delta R.T. -0.006 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062


Tgt Ion: 93 Resp: 456618
 Ion Ratio Lower Upper
 93 100
 95 32.7 27.9 41.9
 123 11.6 9.7 14.5



#29
 2,4-Dichlorophenol
 Concen: 19.933 ng/ul
 RT: 10.557 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55

Tgt Ion: 162 Resp: 317156
 Ion Ratio Lower Upper
 162 100
 164 66.2 50.7 76.1
 98 36.3 27.3 40.9









#30
Naphthalene
Concen: 20.918 ng/ul
RT: 10.963 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:128 Resp: 1171473
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.6 10.4 15.6



#32
4-Chloroaniline
Concen: 18.979 ng/ul
RT: 11.075 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:127 Resp: 463392
Ion Ratio Lower Upper
127 100
129 32.4 26.9 40.3









#33
Hexachlorobutadiene
Concen: 20.487 ng/ul
RT: 11.245 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:225 Resp: 201916
Ion Ratio Lower Upper
225 100
223 62.4 49.3 73.9
227 61.4 49.4 74.0



#34
Caprolactam
Concen: 15.953 ng/ul
RT: 11.857 min Scan# 1440
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:113 Resp: 92174
Ion Ratio Lower Upper
113 100
55 187.2 186.2 279.2
56 138.3 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 18.633 ng/ul
RT: 12.186 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:107 Resp: 309381
Ion Ratio Lower Upper
107 100
144 27.9 23.0 34.6
142 83.6 69.1 103.7



#36
2-Methylnaphthalene
Concen: 19.584 ng/ul
RT: 12.563 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:142 Resp: 758761
Ion Ratio Lower Upper
142 100
141 86.8 70.1 105.1









#37
1-Methylnaphthalene
Concen: 19.500 ng/ul
RT: 12.780 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:142 Resp: 756196
Ion Ratio Lower Upper
142 100
141 91.4 73.4 110.2
116 3.5 2.6 4.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.415 ng/ul
RT: 12.927 min Scan# 1622
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:216 Resp: 371000
Ion Ratio Lower Upper
216 100
214 78.4 62.7 94.1
179 20.5 15.9 23.9
108 16.1 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 18.277 ng/ul
RT: 12.904 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:237 Resp: 174128
Ion Ratio Lower Upper
237 100
235 63.2 49.2 73.8
272 13.6 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 19.666 ng/ul
RT: 13.163 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:196 Resp: 222429
Ion Ratio Lower Upper
196 100
198 94.6 75.6 113.4
200 30.9 24.4 36.6





#42

2,4,5-Trichlorophenol

Concen: 19.773 ng/ul

RT: 13.233 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM037839.D


Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062






Tgt Ion:196 Resp: 238175

Ion Ratio Lower Upper

196 100

198 94.5 76.1 114.1

200 31.1 24.8 37.2



#43

1,1'-Biphenyl

Concen: 21.728 ng/ul

RT: 13.563 min Scan# 1730

Delta R.T. -0.000 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:154 Resp: 955420


Ion Ratio Lower Upper

154 100

153 38.3 31.4 47.0

76 12.3 10.2 15.2









#44
2-Chloronaphthalene
Concen: 21.685 ng/ul
RT: 13.604 min Scan# 1737
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:162 Resp: 748247
Ion Ratio Lower Upper
162 100
164 32.6 26.5 39.7
127 37.6 30.5 45.7



#45
2-Nitroaniline
Concen: 20.130 ng/ul
RT: 13.810 min Scan# 1772
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion: 65 Resp: 196535
Ion Ratio Lower Upper
65 100
92 70.5 54.7 82.1
138 119.3 96.8 145.2





#47

Dimethylphthalate

Concen: 20.064 ng/ul

RT: 14.174 min Scan# 1834

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:163 Resp: 868092

Ion Ratio Lower Upper

163 100

194 3.9 3.6 5.4

164 10.3 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 18.376 ng/ul

RT: 14.298 min Scan# 1855

Delta R.T. -0.000 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:165 Resp: 162323


Ion Ratio Lower Upper

165 100

89 51.5 40.6 60.8

121 21.2 16.4 24.6









#50
Acenaphthylene
Concen: 20.931 ng/ul
RT: 14.445 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:152 Resp: 1172391
Ion Ratio Lower Upper
152 100
151 19.4 15.8 23.6
153 13.3 10.6 15.8



#51
3-Nitroaniline
Concen: 17.781 ng/ul
RT: 14.627 min Scan# 1911
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:138 Resp: 187475
Ion Ratio Lower Upper
138 100
108 11.8 7.8 11.8#
92 104.7 83.8 125.6









#52
Acenaphthene
Concen: 20.785 ng/ul
RT: 14.786 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:153 Resp: 806718
Ion Ratio Lower Upper
153 100
152 47.1 37.0 55.4
154 88.9 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 12.099 ng/ul
RT: 14.827 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:184 Resp: 62421
Ion Ratio Lower Upper
184 100
63 63.9 43.4 65.0
154 60.5 39.7 59.5#









#55
4-Nitrophenol
Concen: 17.591 ng/ul
RT: 14.927 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:109 Resp: 93487
Ion Ratio Lower Upper
109 100
139 178.7 131.6 197.4
65 153.3 109.4 164.0



#56
Dibenzofuran
Concen: 20.455 ng/ul
RT: 15.116 min Scan# 1994
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:168 Resp: 1068892
Ion Ratio Lower Upper
168 100
139 34.9 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 18.383 ng/ul
RT: 15.080 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:165 Resp: 226790
Ion Ratio Lower Upper
165 100
63 40.0 30.2 45.2
182 2.4 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.809 ng/ul
RT: 15.339 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:232 Resp: 188693
Ion Ratio Lower Upper
232 100
131 46.5 32.7 49.1
130 1.9 1.8 2.6
166 31.8 22.2 33.4









#59
Diethylphthalate
Concen: 19.542 ng/ul
RT: 15.527 min Scan# 2064
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:149 Resp: 867086
Ion Ratio Lower Upper
149 100
177 21.9 18.0 27.0
150 12.3 9.8 14.8



#61
Fluorene
Concen: 19.921 ng/ul
RT: 15.763 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:166 Resp: 876355
Ion Ratio Lower Upper
166 100
165 98.4 77.0 115.6
167 13.6 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 19.950 ng/ul
RT: 15.751 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:204 Resp: 418409
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.4
141 58.7 44.7 67.1



#63
4-Nitroaniline
Concen: 17.339 ng/ul
RT: 15.780 min Scan# 2107
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:138 Resp: 183919
Ion Ratio Lower Upper
138 100
92 45.6 36.6 54.8
108 64.1 48.2 72.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.467 ng/ul
 RT: 15.839 min Scan# 2117
 Delta R.T. -0.000 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062


Tgt Ion:198 Resp: 106521
 Ion Ratio Lower Upper
 198 100
 182 4.8 2.9 4.3#
 77 21.7 17.5 26.3



#67
 N-Nitrosodiphenylamine
 Concen: 21.380 ng/ul
 RT: 15.968 min Scan# 2139
 Delta R.T. -0.000 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55

Tgt Ion:169 Resp: 733955
 Ion Ratio Lower Upper
 169 100
 168 68.7 53.8 80.6
 167 36.8 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 20.502 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:248 Resp: 250632
Ion Ratio Lower Upper
248 100
250 94.5 78.1 117.1
141 68.6 51.6 77.4



#69
Hexachlorobenzene
Concen: 20.487 ng/ul
RT: 16.762 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:284 Resp: 292457
Ion Ratio Lower Upper
284 100
142 35.1 24.5 36.7
249 29.4 22.9 34.3





#70

Atrazine

Concen: 19.579 ng/ul

RT: 16.910 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:200 Resp: 231471

Ion Ratio Lower Upper

200 100

173 27.8 22.0 33.0

215 48.9 40.0 60.0



#71

Pentachlorophenol

Concen: 15.909 ng/ul

RT: 17.110 min Scan# 2333

Delta R.T. -0.000 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:266 Resp: 139679


Ion Ratio Lower Upper

266 100

264 64.1 50.3 75.5

268 64.8 52.2 78.2





#72

Phenanthrene

Concen: 20.999 ng/ul

RT: 17.504 min Scan# 2400

Delta R.T. -0.000 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:178 Resp: 1355729

Ion Ratio Lower Upper

178 100

179 15.8 12.5 18.7

176 19.4 15.3 22.9



#74

Anthracene

Concen: 20.919 ng/ul

RT: 17.592 min Scan# 2415

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:178 Resp: 1377762


Ion Ratio Lower Upper

178 100

179 15.0 12.2 18.4

176 18.8 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.864 ng/uL

RT: 13.527 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:216 Resp: 367243

Ion Ratio Lower Upper

216 100

214 77.8 63.0 94.6

218 48.6 40.0 60.0



#76

Pentachlorobenzene

Concen: 21.724 ng/uL

RT: 15.039 min Scan# 1981

Delta R.T. -0.000 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:250 Resp: 379619


Ion Ratio Lower Upper

250 100

248 62.8 50.5 75.7

252 62.8 51.1 76.7









#77
 Carbazole
 Concen: 20.669 ng/ul
 RT: 17.862 min Scan# 2461
 Delta R.T. -0.000 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062


Tgt Ion:167 Resp: 1252755
 Ion Ratio Lower Upper
 167 100
 166 20.9 17.2 25.8
 139 12.5 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 20.006 ng/ul
 RT: 18.409 min Scan# 2554
 Delta R.T. -0.006 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55

Tgt Ion:149 Resp: 1524503
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 4.5 3.4 5.0





#80

Fluoranthene

Concen: 21.815 ng/ul

RT: 19.503 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:202 Resp: 1510889

Ion Ratio Lower Upper

202 100

101 13.2 9.7 14.5

100 9.7 7.7 11.5



#82

Pyrene

Concen: 22.116 ng/ul

RT: 19.868 min Scan# 2802

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:202 Resp: 1581577


Ion Ratio Lower Upper

202 100

101 14.5 11.0 16.4

100 11.6 9.0 13.6





#83

Butylbenzylphthalate

Concen: 20.612 ng/ul

RT: 20.750 min Scan# 2952

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:149 Resp: 642674

Ion Ratio Lower Upper

149 100

91 67.7 51.7 77.5

206 20.3 17.1 25.7



#84

3,3'-Dichlorobenzidine

Concen: 20.357 ng/ul

RT: 21.539 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:252 Resp: 466824


Ion Ratio Lower Upper

252 100

254 63.5 52.2 78.4

126 11.9 8.8 13.2









#85
Benzo(a)anthracene
Concen: 21.059 ng/ul
RT: 21.609 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:228 Resp: 1474937
Ion Ratio Lower Upper
228 100
229 19.6 16.2 24.2
226 27.1 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.072 ng/ul
RT: 21.521 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:149 Resp: 945672
Ion Ratio Lower Upper
149 100
167 28.4 23.3 34.9
279 3.8 3.6 5.4





#87

Chrysene

Concen: 21.279 ng/ul

RT: 21.668 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:228 Resp: 1387298

Ion Ratio Lower Upper

228 100

226 29.5 24.2 36.2

229 19.2 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 19.152 ng/ul




RT: 22.474 min Scan# 3245


Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Tgt Ion:149 Resp: 1629800









#90
Benzo(b)fluoranthene
Concen: 20.724 ng/ul
RT: 23.350 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55




Instrument :
BNA_M
ClientSampleId :
SSTD020062


Tgt Ion:252 Resp: 1573760
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 11.7 9.3 13.9



#91
Benzo(k)fluoranthene
Concen: 20.024 ng/ul
RT: 23.403 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BM037839.D
Acq: 05 Dec 2022 11:55

Tgt Ion:252 Resp: 1466910
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.7 9.3 13.9





#93

Benzo(a)pyrene

Concen: 20.985 ng/ul

RT: 24.003 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55

Instrument :

BNA_M

ClientSampleId :

SSTD020062





Tgt Ion:252 Resp: 1381914

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.2

125 12.3 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.746 ng/ul

RT: 26.715 min Scan# 3966

Delta R.T. -0.006 min

Lab File: BM037839.D

Acq: 05 Dec 2022 11:55




Tgt Ion:276 Resp: 1683994


Ion Ratio Lower Upper

276 100

138 27.7 22.7 34.1

277 25.0 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 22.716 ng/ul
 RT: 26.732 min Scan# 3969
 Delta R.T. -0.006 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020062

Tgt Ion:278 Resp: 1530655
 Ion Ratio Lower Upper
 278 100
 139 17.3 13.9 20.9
 279 24.3 19.0 28.6



#96
 Benzo(g,h,i)perylene
 Concen: 23.313 ng/ul
 RT: 27.526 min Scan# 4104
 Delta R.T. -0.000 min
 Lab File: BM037839.D
 Acq: 05 Dec 2022 11:55

Tgt Ion:276 Resp: 1506207
 Ion Ratio Lower Upper
 276 100
 138 24.7 19.2 28.8
 277 23.6 19.1 28.7

