

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120522\
 Data File : BM037849.D
 Acq On : 05 Dec 2022 18:07
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020063

Quant Time: Dec 05 23:27:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.092	152	277441	20.000	ng/ul	0.00
20) Naphthalene-d8	10.916	136	1313147	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.721	164	841259	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.462	188	1739706	20.000	ng/ul	0.00
79) Chrysene-d12	21.627	240	1277721	20.000	ng/ul	0.00
88) Perylene-d12	24.115	264	1089420	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	48108	8.135	ng/uL	0.00
4) Pyridine-d5	3.863	84	404098	21.129	ng/ul	0.00
7) Phenol-d5	7.245	99	509565	20.386	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.416	67	291128	19.964	ng/ul	0.00
11) 2-Chlorophenol-d4	7.622	132	406789	21.525	ng/ul	0.00
15) 4-Methylphenol-d8	8.804	113	413933	20.824	ng/ul	0.00
21) Nitrobenzene-d5	9.263	128	216178	20.724	ng/ul	0.00
24) 2-Nitrophenol-d4	9.992	143	233627	21.644	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.533	165	424425	21.202	ng/ul	0.00
31) 4-Chloroaniline-d4	11.051	131	607854	19.797	ng/ul	0.00
46) Dimethylphthalate-d6	14.127	166	1192908	20.650	ng/ul	0.00
49) Acenaphthylene-d8	14.421	160	1455775	20.721	ng/ul	0.00
54) 4-Nitrophenol-d4	14.915	143	233323	19.459	ng/ul	0.00
60) Fluorene-d10	15.710	176	1078566	20.580	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.827	200	182938	19.962	ng/ul	0.00
73) Anthracene-d10	17.562	188	1651129	20.638	ng/ul	0.00
81) Pyrene-d10	19.845	212	1714838	25.910	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.950	264	1134790	20.793	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	51531	8.131	ng/uL#	75
5) Pyridine	3.887	79	399964	21.328	ng/ul	95
6) Benzaldehyde	7.228	77	258734	21.493	ng/ul	96
8) Phenol	7.275	94	523386	20.489	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.510	93	413347	20.284	ng/ul	96
12) 2-Chlorophenol	7.651	128	441653	21.712	ng/ul	95
13) 2-Methylphenol	8.539	108	415386	20.390	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.628	45	725217	19.152	ng/ul	98
16) Acetophenone	8.928	105	640184	20.417	ng/ul	96
17) N-Nitroso-di-n-propyla...	8.910	70	317390	20.096	ng/ul	99
18) 4-Methylphenol	8.869	108	451485	20.541	ng/ul	99
19) Hexachloroethane	9.181	117	169387	21.801	ng/ul	97
22) Nitrobenzene	9.310	77	453974	20.767	ng/ul	100
23) Isophorone	9.839	82	918976	19.539	ng/ul	100
25) 2-Nitrophenol	10.022	139	254287	21.739	ng/ul	98
26) 2,4-Dimethylphenol	10.080	107	475460	21.146	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.316	93	571770	20.132	ng/ul	97
29) 2,4-Dichlorophenol	10.557	162	436820	21.574	ng/ul	99
30) Naphthalene	10.969	128	1478942	20.752	ng/ul	100
32) 4-Chloroaniline	11.075	127	618139	19.894	ng/ul	99
33) Hexachlorobutadiene	11.245	225	263101	20.977	ng/ul	98
34) Caprolactam	11.869	113	135285	18.400	ng/ul#	78
35) 4-Chloro-3-methylphenol	12.186	107	444344	21.029	ng/ul	98
36) 2-Methylnaphthalene	12.563	142	998279	20.247	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120522\
 Data File : BM037849.D
 Acq On : 05 Dec 2022 18:07
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020063

Quant Time: Dec 05 23:27:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 02 01:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.780	142	1012294	20.513	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.927	216	521779	21.657	ng/ul	97
40) Hexachlorocyclopentadiene	12.904	237	244818	18.478	ng/ul	98
41) 2,4,6-Trichlorophenol	13.163	196	348442	22.153	ng/ul	98
42) 2,4,5-Trichlorophenol	13.233	196	371284	22.164	ng/ul	99
43) 1,1'-Biphenyl	13.563	154	1295798	21.191	ng/ul	99
44) 2-Chloronaphthalene	13.604	162	1021569	21.289	ng/ul	99
45) 2-Nitroaniline	13.810	65	284335	20.941	ng/ul	96
47) Dimethylphthalate	14.174	163	1239104	20.594	ng/ul	99
48) 2,6-Dinitrotoluene	14.298	165	254418	20.710	ng/ul	100
50) Acenaphthylene	14.445	152	1627624	20.895	ng/ul	100
51) 3-Nitroaniline	14.627	138	289385	19.736	ng/ul	100
52) Acenaphthene	14.786	153	1119802	20.747	ng/ul	99
53) 2,4-Dinitrophenol	14.833	184	131280	18.297	ng/ul	95
55) 4-Nitrophenol	14.933	109	145719	19.716	ng/ul	92
56) Dibenzofuran	15.121	168	1523338	20.962	ng/ul	100
57) 2,4-Dinitrotoluene	15.080	165	365524	21.305	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.345	232	324336	22.012	ng/ul	98
59) Diethylphthalate	15.533	149	1252112	20.292	ng/ul	99
61) Fluorene	15.768	166	1263652	20.655	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.757	204	605434	20.758	ng/ul	98
63) 4-Nitroaniline	15.786	138	285390	19.347	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.839	198	192933	20.422	ng/ul	100
67) N-Nitrosodiphenylamine	15.968	169	1066510	21.273	ng/ul	99
68) 4-Bromophenyl-phenylether	16.645	248	372949	20.890	ng/ul	97
69) Hexachlorobenzene	16.768	284	444940	21.342	ng/ul	100
70) Atrazine	16.915	200	352763	20.432	ng/ul	99
71) Pentachlorophenol	17.109	266	274881	21.438	ng/ul	100
72) Phenanthrene	17.504	178	1952693	20.710	ng/ul	100
74) Anthracene	17.598	178	1996198	20.754	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.527	216	520289	22.181	ng/ul	98
76) Pentachlorobenzene	15.039	250	562775	22.053	ng/ul	100
77) Carbazole	17.868	167	1801925	20.357	ng/ul	99
78) Di-n-butylphthalate	18.415	149	2230072	20.039	ng/ul	100
80) Fluoranthene	19.509	202	2130024	26.348	ng/ul	98
82) Pyrene	19.868	202	2175315	26.061	ng/ul	100
83) Butylbenzylphthalate	20.750	149	908928	24.975	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.539	252	483132	18.049	ng/ul	98
85) Benzo(a)anthracene	21.615	228	1723464	21.082	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.521	149	1301941	23.675	ng/ul	99
87) Chrysene	21.668	228	1606887	21.116	ng/ul	100
89) Di-n-octyl phthalate	22.468	149	2036587	25.977	ng/ul	100
90) Benzo(b)fluoranthene	23.350	252	1498149	21.414	ng/ul	100
91) Benzo(k)fluoranthene	23.403	252	1404869	20.815	ng/ul	98
93) Benzo(a)pyrene	24.003	252	1248066	20.571	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.721	276	1475400	21.631	ng/ul	98
95) Dibenzo(a,h)anthracene	26.738	278	1338800	21.566	ng/ul	99
96) Benzo(g,h,i)perylene	27.521	276	1314917	22.090	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.131 ng/uL
RT: 3.463 min Scan# 11
Delta R.T. -0.001 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion: 88 Resp: 51531
Ion Ratio Lower Upper
88 100
43 33.9 51.3 76.9#
58 82.0 79.0 118.6



#5
Pyridine
Concen: 21.328 ng/uL
RT: 3.887 min Scan# 85
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion: 79 Resp: 399964
Ion Ratio Lower Upper
79 100
52 88.4 75.2 112.8
51 41.3 33.9 50.9









#6
Benzaldehyde
Concen: 21.493 ng/ul
RT: 7.228 min Scan# 611
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion: 77 Resp: 258734
Ion Ratio Lower Upper
77 100
105 105.2 81.2 121.8
106 103.5 79.0 118.6



#8
Phenol
Concen: 20.489 ng/ul
RT: 7.275 min Scan# 661
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion: 94 Resp: 523386
Ion Ratio Lower Upper
94 100
65 26.2 21.1 31.7
66 31.6 25.2 37.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.284 ng/ul
RT: 7.510 min Scan# 701
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion: 93 Resp: 413347
Ion Ratio Lower Upper
93 100
63 77.7 65.7 98.5
95 34.0 26.8 40.2



#12
2-Chlorophenol
Concen: 21.712 ng/ul
RT: 7.651 min Scan# 725
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:128 Resp: 441653
Ion Ratio Lower Upper
128 100
64 42.3 37.3 55.9
130 32.1 27.1 40.7









#13
2-Methylphenol
Concen: 20.390 ng/ul
RT: 8.539 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:108 Resp: 415386
Ion Ratio Lower Upper
108 100
107 87.8 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.152 ng/ul
RT: 8.628 min Scan# 891
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion: 45 Resp: 725217
Ion Ratio Lower Upper
45 100
77 11.2 8.2 12.4
79 9.2 6.7 10.1









#16
Acetophenone
Concen: 20.417 ng/ul
RT: 8.928 min Scan# 941
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:105 Resp: 640184
Ion Ratio Lower Upper
105 100
77 70.4 59.0 88.4
51 27.4 24.4 36.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.096 ng/ul
RT: 8.910 min Scan# 939
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion: 70 Resp: 317390
Ion Ratio Lower Upper
70 100
42 69.6 55.8 83.6
101 10.7 8.0 12.0
130 24.2 18.6 27.8









#18
4-Methylphenol
Concen: 20.541 ng/ul
RT: 8.869 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:108 Resp: 451485
Ion Ratio Lower Upper
108 100
107 110.5 89.3 133.9



#19
Hexachloroethane
Concen: 21.801 ng/ul
RT: 9.181 min Scan# 985
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:117 Resp: 169387
Ion Ratio Lower Upper
117 100
201 104.1 82.3 123.5
199 66.1 49.8 74.6









#22
Nitrobenzene
Concen: 20.767 ng/ul
RT: 9.310 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion: 77 Resp: 453974
Ion Ratio Lower Upper
77 100
123 51.6 41.4 62.0
65 13.0 10.9 16.3



#23
Isophorone
Concen: 19.539 ng/ul
RT: 9.839 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion: 82 Resp: 918976
Ion Ratio Lower Upper
82 100
95 6.3 5.5 8.3
138 19.1 15.3 22.9









#25
2-Nitrophenol
Concen: 21.739 ng/ul
RT: 10.022 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:139 Resp: 254287
Ion Ratio Lower Upper
139 100
65 45.5 36.6 54.8
109 24.3 21.2 31.8



#26
2,4-Dimethylphenol
Concen: 21.146 ng/ul
RT: 10.080 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:107 Resp: 475460
Ion Ratio Lower Upper
107 100
121 52.3 41.8 62.8
122 93.4 73.1 109.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.132 ng/ul
 RT: 10.316 min Scan# 1178
 Delta R.T. -0.006 min
 Lab File: BM037849.D
 Acq: 05 Dec 2022 18:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020063


Tgt Ion: 93 Resp: 571770
 Ion Ratio Lower Upper
 93 100
 95 33.0 27.9 41.9
 123 11.8 9.7 14.5



#29
 2,4-Dichlorophenol
 Concen: 21.574 ng/ul
 RT: 10.557 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BM037849.D
 Acq: 05 Dec 2022 18:07

Tgt Ion: 162 Resp: 436820
 Ion Ratio Lower Upper
 162 100
 164 64.7 50.7 76.1
 98 33.5 27.3 40.9









#30
Naphthalene
Concen: 20.752 ng/ul
RT: 10.969 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:128 Resp: 1478942
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 19.894 ng/ul
RT: 11.075 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:127 Resp: 618139
Ion Ratio Lower Upper
127 100
129 33.2 26.9 40.3





#33

Hexachlorobutadiene

Concen: 20.977 ng/ul

RT: 11.245 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:225 Resp: 263101

Ion Ratio Lower Upper

225 100

223 62.4 49.3 73.9

227 63.8 49.4 74.0



#34

Caprolactam

Concen: 18.400 ng/ul

RT: 11.869 min Scan# 1442

Delta R.T. 0.011 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion:113 Resp: 135285


Ion Ratio Lower Upper

113 100

55 185.6 186.2 279.2#

56 132.8 119.8 179.8









#35
4-Chloro-3-methylphenol
Concen: 21.029 ng/ul
RT: 12.186 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:107 Resp: 444344
Ion Ratio Lower Upper
107 100
144 26.7 23.0 34.6
142 85.5 69.1 103.7



#36
2-Methylnaphthalene
Concen: 20.247 ng/ul
RT: 12.563 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:142 Resp: 998279
Ion Ratio Lower Upper
142 100
141 88.6 70.1 105.1





#37

1-Methylnaphthalene

Concen: 20.513 ng/ul

RT: 12.780 min Scan# 1597

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:142 Resp: 1012294

Ion Ratio Lower Upper

142 100

141 90.6 73.4 110.2

116 3.4 2.6 4.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.657 ng/ul

RT: 12.927 min Scan# 1622

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Tgt Ion:216 Resp: 521779




Ion Ratio Lower Upper


216 100

214 75.5 62.7 94.1

179 19.4 15.9 23.9

108 15.9 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 18.478 ng/ul
RT: 12.904 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:237 Resp: 244818
Ion Ratio Lower Upper
237 100
235 59.8 49.2 73.8
272 13.2 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 22.153 ng/ul
RT: 13.163 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:196 Resp: 348442
Ion Ratio Lower Upper
196 100
198 92.8 75.6 113.4
200 29.1 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 22.164 ng/ul
RT: 13.233 min Scan# 1674
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:196 Resp: 371284
Ion Ratio Lower Upper
196 100
198 95.6 76.1 114.1
200 29.9 24.8 37.2



#43
1,1'-Biphenyl
Concen: 21.191 ng/ul
RT: 13.563 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:154 Resp: 1295798
Ion Ratio Lower Upper
154 100
153 38.9 31.4 47.0
76 11.8 10.2 15.2





#44

2-Chloronaphthalene

Concen: 21.289 ng/ul

RT: 13.604 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:162 Resp: 1021569

Ion Ratio Lower Upper

162 100

164 32.0 26.5 39.7

127 38.0 30.5 45.7



#45

2-Nitroaniline

Concen: 20.941 ng/ul

RT: 13.810 min Scan# 1772

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion: 65 Resp: 284335


Ion Ratio Lower Upper

65 100

92 73.7 54.7 82.1

138 124.0 96.8 145.2





#47

Dimethylphthalate

Concen: 20.594 ng/ul

RT: 14.174 min Scan# 1834

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:163 Resp: 1239104

Ion Ratio Lower Upper

163 100

194 4.1 3.6 5.4

164 10.4 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 20.710 ng/ul

RT: 14.298 min Scan# 1855

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion:165 Resp: 254418


Ion Ratio Lower Upper

165 100

89 51.1 40.6 60.8

121 20.6 16.4 24.6





#50

Acenaphthylene

Concen: 20.895 ng/ul

RT: 14.445 min Scan# 1880

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:152 Resp: 1627624

Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.6

153 13.3 10.6 15.8



#51

3-Nitroaniline

Concen: 19.736 ng/ul

RT: 14.627 min Scan# 1911

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion:138 Resp: 289385


Ion Ratio Lower Upper

138 100

108 9.3 7.8 11.8

92 104.5 83.8 125.6









#52
Acenaphthene
Concen: 20.747 ng/ul
RT: 14.786 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:153 Resp: 1119802
Ion Ratio Lower Upper
153 100
152 46.8 37.0 55.4
154 88.5 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 18.297 ng/ul
RT: 14.833 min Scan# 1946
Delta R.T. 0.005 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:184 Resp: 131280
Ion Ratio Lower Upper
184 100
63 49.0 43.4 65.0
154 51.1 39.7 59.5









#55
4-Nitrophenol
Concen: 19.716 ng/ul
RT: 14.933 min Scan# 1963
Delta R.T. 0.005 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:109 Resp: 145719
Ion Ratio Lower Upper
109 100
139 177.4 131.6 197.4
65 143.6 109.4 164.0



#56
Dibenzofuran
Concen: 20.962 ng/ul
RT: 15.121 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:168 Resp: 1523338
Ion Ratio Lower Upper
168 100
139 34.5 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 21.305 ng/ul

RT: 15.080 min Scan# 1988

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:165 Resp: 365524

Ion Ratio Lower Upper

165 100

63 39.0 30.2 45.2

182 2.6 1.8 2.8



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.012 ng/ul

RT: 15.345 min Scan# 2033

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Tgt Ion:232 Resp: 324336




Ion Ratio Lower Upper


232 100

131 42.1 32.7 49.1

130 2.3 1.8 2.6

166 29.4 22.2 33.4









#59
Diethylphthalate
Concen: 20.292 ng/ul
RT: 15.533 min Scan# 20
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:149 Resp: 1252112
Ion Ratio Lower Upper
149 100
177 22.0 18.0 27.0
150 12.0 9.8 14.8



#61
Fluorene
Concen: 20.655 ng/ul
RT: 15.768 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:166 Resp: 1263652
Ion Ratio Lower Upper
166 100
165 97.5 77.0 115.6
167 13.0 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.758 ng/ul
RT: 15.757 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:204 Resp: 605434
Ion Ratio Lower Upper
204 100
206 33.6 26.2 39.4
141 54.7 44.7 67.1



#63
4-Nitroaniline
Concen: 19.347 ng/ul
RT: 15.786 min Scan# 2108
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:138 Resp: 285390
Ion Ratio Lower Upper
138 100
92 44.4 36.6 54.8
108 60.9 48.2 72.2









#66
4,6-Dinitro-2-methylphenol
Concen: 20.422 ng/ul
RT: 15.839 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:198 Resp: 192933
Ion Ratio Lower Upper
198 100
182 4.2 2.9 4.3
77 21.8 17.5 26.3



#67
N-Nitrosodiphenylamine
Concen: 21.273 ng/ul
RT: 15.968 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:169 Resp: 1066510
Ion Ratio Lower Upper
169 100
168 66.4 53.8 80.6
167 36.3 29.5 44.3









#68
4-Bromophenyl-phenylether
Concen: 20.890 ng/ul
RT: 16.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:248 Resp: 372949
Ion Ratio Lower Upper
248 100
250 96.8 78.1 117.1
141 68.6 51.6 77.4



#69
Hexachlorobenzene
Concen: 21.342 ng/ul
RT: 16.768 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:284 Resp: 444940
Ion Ratio Lower Upper
284 100
142 30.7 24.5 36.7
249 29.0 22.9 34.3









#70
Atrazine
Concen: 20.432 ng/ul
RT: 16.915 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:200 Resp: 352763
Ion Ratio Lower Upper
200 100
173 27.0 22.0 33.0
215 49.5 40.0 60.0



#71
Pentachlorophenol
Concen: 21.438 ng/ul
RT: 17.109 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:266 Resp: 274881
Ion Ratio Lower Upper
266 100
264 63.0 50.3 75.5
268 65.8 52.2 78.2





#72

Phenanthrene

Concen: 20.710 ng/ul

RT: 17.504 min Scan# 2400

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:178 Resp: 1952693

Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 19.3 15.3 22.9



#74

Anthracene

Concen: 20.754 ng/ul

RT: 17.598 min Scan# 2416

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion:178 Resp: 1996198


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.181 ng/uL
RT: 13.527 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:216 Resp: 520289
Ion Ratio Lower Upper
216 100
214 80.7 63.0 94.6
218 48.6 40.0 60.0



#76
Pentachlorobenzene
Concen: 22.053 ng/uL
RT: 15.039 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:250 Resp: 562775
Ion Ratio Lower Upper
250 100
248 63.0 50.5 75.7
252 63.9 51.1 76.7









#77
 Carbazole
 Concen: 20.357 ng/ul
 RT: 17.868 min Scan# 2461
 Delta R.T. 0.005 min
 Lab File: BM037849.D
 Acq: 05 Dec 2022 18:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020063


Tgt Ion:167 Resp: 1801925
 Ion Ratio Lower Upper
 167 100
 166 20.8 17.2 25.8
 139 12.5 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 20.039 ng/ul
 RT: 18.415 min Scan# 2555
 Delta R.T. -0.000 min
 Lab File: BM037849.D
 Acq: 05 Dec 2022 18:07

Tgt Ion:149 Resp: 2230072
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 4.2 3.4 5.0









#80
Fluoranthene
Concen: 26.348 ng/ul
RT: 19.509 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:202 Resp: 2130024
Ion Ratio Lower Upper
202 100
101 11.4 9.7 14.5
100 8.5 7.7 11.5



#82
Pyrene
Concen: 26.061 ng/ul
RT: 19.868 min Scan# 2802
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:202 Resp: 2175315
Ion Ratio Lower Upper
202 100
101 13.5 11.0 16.4
100 11.4 9.0 13.6





#83

Butylbenzylphthalate

Concen: 24.975 ng/ul

RT: 20.750 min Scan# 2952

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:149 Resp: 908928

Ion Ratio Lower Upper

149 100

91 63.6 51.7 77.5

206 21.1 17.1 25.7



#84

3,3'-Dichlorobenzidine

Concen: 18.049 ng/ul

RT: 21.539 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion:252 Resp: 483132


Ion Ratio Lower Upper

252 100

254 63.9 52.2 78.4

126 10.9 8.8 13.2









#85
Benzo(a)anthracene
Concen: 21.082 ng/ul
RT: 21.615 min Scan# 3099
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:228 Resp: 1723464
Ion Ratio Lower Upper
228 100
229 19.2 16.2 24.2
226 27.1 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 23.675 ng/ul
RT: 21.521 min Scan# 3083
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:149 Resp: 1301941
Ion Ratio Lower Upper
149 100
167 28.8 23.3 34.9
279 4.3 3.6 5.4









#87
Chrysene
Concen: 21.116 ng/ul
RT: 21.668 min Scan# 3108
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:228 Resp: 1606887
Ion Ratio Lower Upper
228 100
226 30.2 24.2 36.2
229 19.6 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 25.977 ng/ul
RT: 22.468 min Scan# 3244
Delta R.T. -0.012 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:149 Resp: 2036587









#90
Benzo(b)fluoranthene
Concen: 21.414 ng/ul
RT: 23.350 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07




Instrument :
BNA_M
ClientSampleId :
SSTD020063


Tgt Ion:252 Resp: 1498149
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 11.2 9.3 13.9



#91
Benzo(k)fluoranthene
Concen: 20.815 ng/ul
RT: 23.403 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:252 Resp: 1404869
Ion Ratio Lower Upper
252 100
253 21.3 17.8 26.8
125 10.5 9.3 13.9





#93

Benzo(a)pyrene

Concen: 20.571 ng/ul

RT: 24.003 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07

Instrument :

BNA_M

ClientSampleId :

SSTD020063





Tgt Ion:252 Resp: 1248066

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 12.0 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.631 ng/ul

RT: 26.721 min Scan# 3967

Delta R.T. -0.000 min

Lab File: BM037849.D

Acq: 05 Dec 2022 18:07




Tgt Ion:276 Resp: 1475400


Ion Ratio Lower Upper

276 100

138 26.7 22.7 34.1

277 24.8 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 21.566 ng/ul
RT: 26.738 min Scan# 3969
Delta R.T. -0.000 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Instrument :
BNA_M
ClientSampleId :
SSTD020063

Tgt Ion:278 Resp: 1338800
Ion Ratio Lower Upper
278 100
139 16.8 13.9 20.9
279 23.2 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 22.090 ng/ul
RT: 27.521 min Scan# 4103
Delta R.T. -0.006 min
Lab File: BM037849.D
Acq: 05 Dec 2022 18:07

Tgt Ion:276 Resp: 1314917
Ion Ratio Lower Upper
276 100
138 22.0 19.2 28.8
277 24.5 19.1 28.7

