

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121421\
 Data File : BM033444.D
 Acq On : 14 Dec 2021 17:00
 Operator : CG/JU
 Sample : SSTD3.220
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2027

Quant Time: Dec 15 00:39:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 15 00:38:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.901	152	610	0.400	ng/ul	0.00
4) Naphthalene-d8	10.688	136	1555	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.519	164	1042	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.257	188	2635	0.400	ng/ul	-0.02
17) Chrysene-d12	21.422	240	3820	0.400	ng/ul	-0.01
23) Perylene-d12	23.745	264	3597	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.362	96	1448	2.323	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.396	88	1966	2.708	ng/ul#	81
14) Pentachlorophenol	16.917	266	3533	4.039	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121421\
Data File : BM033444.D
Acq On : 14 Dec 2021 17:00
Operator : CG/JU
Sample : SSTD3.220
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2027

Quant Time: Dec 15 00:39:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM121421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 15 00:38:17 2021
Response via : Initial Calibration

