

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121624\
 Data File : BM049006.D
 Acq On : 16 Dec 2024 15:58
 Operator : RC/JU
 Sample : P5228-13
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E29R6

Quant Time: Dec 16 17:34:10 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 09 17:52:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.573	152	5504	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.336	136	15435	0.400	ng/ul	#-0.02
9) Acenaphthene-d10	14.205	164	7949	0.400	ng/ul	-0.02
13) Phenanthrene-d10	16.950	188	16182m	0.400	ng/ul	-0.02
17) Chrysene-d12	21.152	240	10668m	0.400	ng/ul	-0.01
23) Perylene-d12	23.309	264	10094m	0.400	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	29009	4.510	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.936	152	5434	0.276	ng/ul	-0.01
18) Fluoranthene-d10	18.987	212	12852	0.403	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121624\
Data File : BM049006.D
Acq On : 16 Dec 2024 15:58
Operator : RC/JU
Sample : P5228-13
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E29R6

Quant Time: Dec 16 17:34:10 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 09 17:52:26 2024
Response via : Initial Calibration

