

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050619\
 Data File : BM020135.D
 Acq On : 06 May 2019 14:48
 Operator : JU/SJ
 Sample : K2647-01 2X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SU-04-050219-A

Quant Time: May 07 05:01:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM043019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:58:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	133286	20.00	ng	0.00
21) Naphthalene-d8	10.58	136	517810	20.00	ng	0.00
39) Acenaphthene-d10	14.43	164	309020	20.00	ng	0.00
64) Phenanthrene-d10	17.18	188	719121	20.00	ng	0.00
76) Chrysene-d12	21.36	240	750323	20.00	ng	0.00
87) Perylene-d12	23.66	264	880864	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	446694	54.78	ng	0.00
7) Phenol-d6	6.95	99	657526	59.03	ng	0.00
23) Nitrobenzene-d5	8.95	82	446898	34.07	ng	0.00
42) 2,4,6-Tribromophenol	15.92	330	256371	53.45	ng	0.00
45) 2-Fluorobiphenyl	13.05	172	831611	33.11	ng	0.00
79) Terphenyl-d14	19.80	244	1257467	29.23	ng	0.00
Target Compounds						
38) 1-Methylnaphthalene	12.47	142	105058	5.484	ng	99
50) Dimethylphthalate	13.89	163	89947	3.425	ng	99
71) Phenanthrene	17.22	178	163743	4.125	ng	97
75) Fluoranthene	19.24	202	226766	4.515	ng	95
78) Pyrene	19.60	202	236731	4.699	ng	98
81) Benzo(a)anthracene	21.35	228	149428	2.840	ng	# 85
83) Chrysene	21.40	228	140081	2.745	ng	# 85
88) Benzo(b)fluoranthene	22.96	252	201159	3.458	ng	# 75
90) Benzo(a)pyrene	23.56	252	140393	2.657	ng	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM050619\
Data File : BM020135.D
Acq On : 06 May 2019 14:48
Operator : JU/SJ
Sample : K2647-01 2X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SU-04-050219-A

Quant Time: May 07 05:01:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM043019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:58:37 2019
Response via : Initial Calibration

