

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071621\
 Data File : BM030993.D
 Acq On : 16 Jul 2021 17:13
 Operator : CG/JU
 Sample : PB137665BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS665

Quant Time: Jul 16 17:46:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM071321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 16 16:49:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.649	152	906	0.40	ng/ul	0.00
4) Naphthalene-d8	10.414	136	3027	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.280	164	1941	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.026	188	4283	0.40	ng/ul	0.00
17) Chrysene-d12	21.217	240	4103	0.40	ng/ul	0.00
23) Perylene-d12	23.401	264	4224	0.40	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.224	96	890	0.96	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.006	152	2225	0.42	ng/ul	0.00
18) Fluoranthene-d10	19.058	212	5747	0.42	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.255	88	2146	2.06	ng/ul#	85
5) Naphthalene	10.463	128	3599	0.40	ng/ul	99
7) 2-Methylnaphthalene	12.082	142	2521	0.39	ng/ul	99
8) 1-Methylnaphthalene	12.303	142	2301	0.37	ng/ul	100
10) Acenaphthylene	13.998	152	2944	0.31	ng/ul#	85
11) Acenaphthene	14.341	153	2702	0.39	ng/ul	97
12) Fluorene	15.330	166	3229	0.40	ng/ul	98
14) Pentachlorophenol	16.676	266	1090	0.72	ng/ul	99
15) Phenanthrene	17.065	178	5405	0.40	ng/ul	99
16) Anthracene	17.158	178	5049	0.39	ng/ul	98
19) Fluoranthene	19.088	202	6834	0.38	ng/ul	99
20) Pyrene	19.450	202	7025	0.38	ng/ul	99
21) Benzo(a)anthracene	21.202	228	7200	0.41	ng/ul	99
22) Chrysene	21.253	228	7548	0.41	ng/ul	98
24) Benzo(b)fluoranthene	22.749	252	7945	0.39	ng/ul	94
25) Benzo(k)fluoranthene	22.793	252	8524	0.41	ng/ul	95
26) Benzo(a)pyrene	23.307	252	6918	0.39	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.572	276	9984	0.41	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.591	278	7986	0.41	ng/ul	93
29) Benzo(g,h,i)perylene	26.233	276	8252	0.40	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071621\
Data File : BM030993.D
Acq On : 16 Jul 2021 17:13
Operator : CG/JU
Sample : PB137665BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS665

Quant Time: Jul 16 17:46:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM071321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 16 16:49:56 2021
Response via : Initial Calibration

