

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080719\
 Data File : BM021891.D
 Acq On : 07 Aug 2019 12:10
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.439

Quant Time: Aug 07 15:07:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 08:28:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	121	0.00
3	Naphthalene	0.400	0.378	5.5	115	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	123	-0.01
5	2-Methylnaphthalene	0.400	0.402	-0.5	122	-0.01
6 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.00
7	Acenaphthylene	0.400	0.397	0.8	118	0.00
8 C	Acenaphthene	0.400	0.394	1.5	119	0.00
9	Fluorene	0.400	0.385	3.8	115	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
11	Pentachlorophenol	0.400	0.446	-11.5	143	0.00
12	Phenanthrene	0.400	0.384	4.0	111	0.00
13	Anthracene	0.400	0.419	-4.7	122	0.00
14 SURR	Fluoranthene-d10	0.400	0.375	6.3	107	0.00
15 C	Fluoranthene	0.400	0.373	6.8	109	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
17	Pyrene	0.400	0.387	3.3	106	0.00
18	Benzo(a)anthracene	0.400	0.407	-1.7	112	0.00
19	Chrysene	0.400	0.365	8.8	102	0.00
20 I	Perylene-d12	0.400	0.400	0.0	115	0.00
21	Benzo(b)fluoranthene	0.400	0.387	3.3	108	0.00
22	Benzo(k)fluoranthene	0.400	0.367	8.3	104	0.00
23 C	Benzo(a)pyrene	0.400	0.384	4.0	109	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.387	3.3	110	0.00
25	Dibenzo(a,h)anthracene	0.400	0.382	4.5	109	0.00
26	Benzo(g,h,i)perylene	0.400	0.366	8.5	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0