

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080719\
 Data File : BM021900.D
 Acq On : 07 Aug 2019 17:40
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.440

Quant Time: Aug 07 18:36:22 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 08:28:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	117	0.00
3	Naphthalene	0.400	0.383	4.3	112	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.399	0.3	118	0.00
5	2-Methylnaphthalene	0.400	0.399	0.3	117	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
7	Acenaphthylene	0.400	0.394	1.5	115	0.00
8 C	Acenaphthene	0.400	0.385	3.8	114	0.00
9	Fluorene	0.400	0.379	5.3	111	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
11	Pentachlorophenol	0.400	0.440	-10.0	137	0.00
12	Phenanthrene	0.400	0.386	3.5	109	0.00
13	Anthracene	0.400	0.416	-4.0	118	0.00
14 SURR	Fluoranthene-d10	0.400	0.374	6.5	104	0.00
15 C	Fluoranthene	0.400	0.376	6.0	106	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	110	0.00
17	Pyrene	0.400	0.386	3.5	105	0.00
18	Benzo(a)anthracene	0.400	0.405	-1.3	110	0.00
19	Chrysene	0.400	0.366	8.5	101	0.00
20 I	Perylene-d12	0.400	0.400	0.0	112	0.00
21	Benzo(b)fluoranthene	0.400	0.397	0.8	108	0.00
22	Benzo(k)fluoranthene	0.400	0.368	8.0	101	0.00
23 C	Benzo(a)pyrene	0.400	0.378	5.5	105	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.390	2.5	108	0.00
25	Dibenzo(a,h)anthracene	0.400	0.389	2.8	108	0.00
26	Benzo(g,h,i)perylene	0.400	0.369	7.8	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0