

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080719\
 Data File : BM021922.D
 Acq On : 08 Aug 2019 07:39
 Operator : HP/JU
 Sample : K4028-16
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BENR1

Manual Integrations
 APPROVED

mohammad
 8/9/2019 6:11:58 AM

Quant Time: Aug 08 09:28:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 08 06:54:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	945	0.40	ng/ul	0.00
2) Naphthalene-d8	10.40	136	4270	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.27	164	2403	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.03	188	5494m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.24	240	4935m	0.40	ng/ul	0.00
20) Perylene-d12	23.45	264	5190m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.02	152	1537	0.22	ng/ul	0.00
14) Fluoranthene-d10	19.07	212	3914m	0.21	ng/ul	-0.01

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080719\
Data File : BM021922.D
Acq On : 08 Aug 2019 07:39
Operator : HP/JU
Sample : K4028-16
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BENR1

Manual Integrations
APPROVED

mohammad
8/9/2019 6:11:58 AM

Quant Time: Aug 08 09:28:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM080319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 08 06:54:24 2019
Response via : Initial Calibration

