

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080924\
 Data File : BM047112.D
 Acq On : 09 Aug 2024 08:52
 Operator : RC/JU
 Sample : P3435-15
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E05Z7

Quant Time: Aug 09 09:27:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 09 01:30:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.387	152	259246	20.000	ng/u1	0.00
20) Naphthalene-d8	10.139	136	1048506	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.039	164	668532	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.798	188	1365311	20.000	ng/u1	0.00
79) Chrysene-d12	21.033	240	1175482	20.000	ng/u1	0.00
88) Perylene-d12	23.739	264	1390003	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.016	96	11843	1.810	ng/uL	0.00
4) Pyridine-d5	3.405	84	146033	8.182	ng/u1	0.00
7) Phenol-d5	6.593	99	255397	12.969	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.740	67	173184	13.365	ng/u1	0.00
11) 2-Chlorophenol-d4	6.928	132	211167	12.658	ng/u1	0.00
15) 4-Methylphenol-d8	8.104	113	210892	13.518	ng/u1	0.00
21) Nitrobenzene-d5	8.528	128	104848	12.462	ng/u1	0.00
24) 2-Nitrophenol-d4	9.239	143	111339	12.118	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.781	165	214736	12.173	ng/u1	0.00
31) 4-Chloroaniline-d4	10.292	131	270862	11.683	ng/u1	0.00
46) Dimethylphthalate-d6	13.469	166	688158	13.462	ng/u1	0.00
49) Acenaphthylene-d8	13.727	160	787673	13.805	ng/u1	0.00
54) 4-Nitrophenol-d4	14.286	143	79600	8.054	ng/u1	0.00
60) Fluorene-d10	15.045	176	604604	13.827	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.186	200	59170	6.860	ng/u1	0.00
73) Anthracene-d10	16.898	188	936184	14.464	ng/u1	0.00
81) Pyrene-d10	19.209	212	1039964	15.455	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.550	264	1069031	14.606	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080924\
Data File : BM047112.D
Acq On : 09 Aug 2024 08:52
Operator : RC/JU
Sample : P3435-15
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E05Z7

Quant Time: Aug 09 09:27:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 09 01:30:34 2024
Response via : Initial Calibration

