

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010216\
 Data File : BM003915.D
 Acq On : 31 Dec 2015 15:40
 Operator : SJ/UM
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM010216

Quant Time: Jan 04 13:45:45 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 04 13:33:21 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	173#	0.00
2	1,4-Dioxane	0.481	0.425	11.6	166#	0.00
3	n-Nitrosodimethylamine	0.517	0.528	-2.1	177#	-0.02
4 S	2-Fluorophenol	1.023	0.989	3.3	178#	0.00
5 S	Phenol-d6	1.253	1.277	-1.9	181#	0.00
6	bis(2-Chloroethyl)ether	1.145	1.155	-0.9	172#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	172#	0.00
8 S	Nitrobenzene-d5	0.246	0.256	-4.1	181#	-0.01
9	Nitrobenzene	0.266	0.280	-5.3	185#	-0.01
10	Naphthalene	1.094	1.068	2.4	169#	-0.02
11	Hexachlorobutadiene	0.167	0.162	3.0	171#	0.00
12	2-Methylnaphthalene	0.739	0.733	0.8	171#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	173#	0.00
14 S	2,4,6-Tribromophenol	0.116	0.117	-0.9	186#	0.00
15 S	2-Fluorobiphenyl	1.612	1.570	2.6	170#	0.00
16	Acenaphthylene	2.086	2.199	-5.4	181#	0.00
17	Acenaphthene	1.457	1.376	5.6	173#	0.00
18	Fluorene	1.734	1.733	0.1	170#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	170#	0.00
20	4-Bromophenyl-phenylether	0.155	0.148	4.5	173#	0.00
21	Hexachlorobenzene	0.164	0.154	6.1	176#	0.00
22	Pentachlorophenol	0.043	0.044	-2.3	221#	0.00
23	Phenanthrene	1.037	1.023	1.4	176#	0.00
24	Anthracene	1.006	1.001	0.5	182#	0.00
25	Fluoranthene	1.113	1.160	-4.2	184#	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	182#	0.02
27	Benzydine	0.418	0.451	-7.9	213#	0.00
28	Pvrene	1.796	1.776	1.1	179#	0.00
29 S	Terphenyl-d14	0.809	0.812	-0.4	184#	0.00
30	Benzo(a)anthracene	1.421	1.372	3.4	180#	0.02
31	3,3'-Dichlorobenzidine	0.361	0.422	-16.9	229#	0.00
32	Chrysene	1.475	1.586	-7.5	193#	0.00
33	Bis(2-ethylhexyl)phthalate	0.629	0.720	-14.5	257#	0.00
34	Indeno(1,2,3-cd)pyrene	1.057	1.056	0.1	181#	0.01
35 I	Perylene-d12	1.000	1.000	0.0	188#	0.00
36	Benzo(b)fluoranthene	1.528	1.439	5.8	189#	0.00
37	Benzo(k)fluoranthene	1.480	1.489	-0.6	193#	0.01
38 C	Benzo(a)pyrene	1.301	1.273	2.2	195#	0.00
39	Dibenzo(a,h)anthracene	1.110	1.063	4.2	180#	0.00
40	Benzo(a,h,i)perylene	1.176	1.098	6.6	177#	0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010216\
Data File : BM003915.D
Acq On : 31 Dec 2015 15:40
Operator : SJ/UM
Sample : SSTDICV001
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM010216

Quant Time: Jan 04 13:45:45 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 04 13:33:21 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				