

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010325\
 Data File : BM049332.D
 Acq On : 03 Jan 2025 17:09
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM010325

Quant Time: Jan 03 17:45:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 03 17:24:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	37.956	5.1	117	0.00
3	Pyridine	40.000	38.489	3.8	114	0.00
4	n-Nitrosodimethylamine	40.000	38.499	3.8	114	0.00
5 S	2-Fluorophenol	80.000	78.975	1.3	115	0.00
6	Aniline	40.000	39.918	0.2	117	0.00
7 S	Phenol-d6	80.000	80.551	-0.7	117	0.00
8	2-Chlorophenol	40.000	39.711	0.7	117	0.00
9	Benzaldehyde	40.000	39.128	2.2	121	0.00
10 C	Phenol	40.000	39.862	0.3	117	0.00
11	bis(2-Chloroethyl)ether	40.000	39.286	1.8	115	0.00
12	1,3-Dichlorobenzene	40.000	39.417	1.5	117	0.00
13 C	1,4-Dichlorobenzene	40.000	39.486	1.3	117	0.00
14	1,2-Dichlorobenzene	40.000	39.303	1.7	116	0.00
15	Benzyl Alcohol	40.000	40.391	-1.0	118	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.888	2.8	116	0.00
17	2-Methylphenol	40.000	40.040	-0.1	116	0.00
18	Hexachloroethane	40.000	38.981	2.5	114	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.832	-2.1	117	0.00
20	3+4-Methylphenols	40.000	40.683	-1.7	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.00
22	Acetophenone	40.000	39.694	0.8	117	0.00
23 S	Nitrobenzene-d5	80.000	79.381	0.8	115	0.00
24	Nitrobenzene	40.000	39.126	2.2	116	0.00
25	Isophorone	40.000	39.882	0.3	118	0.00
26 C	2-Nitrophenol	40.000	41.515	-3.8	119	0.00
27	2,4-Dimethylphenol	40.000	40.396	-1.0	118	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.567	1.1	117	0.00
29 C	2,4-Dichlorophenol	40.000	40.879	-2.2	119	0.00
30	1,2,4-Trichlorobenzene	40.000	39.322	1.7	117	0.00
31	Naphthalene	40.000	39.264	1.8	117	0.00
32	Benzoic acid	40.000	41.115	-2.8	122	0.00
33	4-Chloroaniline	40.000	40.631	-1.6	119	0.00
34 C	Hexachlorobutadiene	40.000	39.314	1.7	118	0.00
35	Caprolactam	40.000	41.563	-3.9	122	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.336	-0.8	119	0.00
37	2-Methylnaphthalene	40.000	40.269	-0.7	119	0.00
38	1-Methylnaphthalene	40.000	39.874	0.3	118	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.220	2.0	119	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.393	1.5	115	0.00
42 S	2,4,6-Tribromophenol	80.000	81.271	-1.6	122	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.166	-0.4	119	0.00
44	2,4,5-Trichlorophenol	40.000	39.934	0.2	120	0.00
45 S	2-Fluorobiphenyl	80.000	79.492	0.6	119	0.00
46	1,1'-Biphenyl	40.000	39.537	1.2	119	0.00
47	2-Chloronaphthalene	40.000	39.444	1.4	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010325\
 Data File : BM049332.D
 Acq On : 03 Jan 2025 17:09
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM010325

Quant Time: Jan 03 17:45:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 03 17:24:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.984	-2.5	121	0.00
49	Acenaphthylene	40.000	39.505	1.2	119	0.00
50	Dimethylphthalate	40.000	39.231	1.9	120	0.00
51	2,6-Dinitrotoluene	40.000	40.593	-1.5	121	0.00
52 C	Acenaphthene	40.000	40.080	-0.2	121	0.00
53	3-Nitroaniline	40.000	41.146	-2.9	120	0.00
54 P	2,4-Dinitrophenol	40.000	38.946	2.6	118	0.00
55	Dibenzofuran	40.000	39.169	2.1	119	0.00
56 P	4-Nitrophenol	40.000	39.805	0.5	122	0.00
57	2,4-Dinitrotoluene	40.000	41.458	-3.6	121	0.00
58	Fluorene	40.000	39.598	1.0	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.087	-0.2	122	0.00
60	Diethylphthalate	40.000	39.356	1.6	120	0.00
61	4-Chlorophenyl-phenylether	40.000	39.709	0.7	120	0.00
62	4-Nitroaniline	40.000	42.290	-5.7	122	0.00
63	Azobenzene	40.000	39.285	1.8	119	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.653	-1.6	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.470	-3.7	121	0.00
67	4-Bromophenyl-phenylether	40.000	40.721	-1.8	120	0.00
68	Hexachlorobenzene	40.000	40.113	-0.3	119	0.00
69	Atrazine	40.000	38.659	3.4	121	0.00
70 C	Pentachlorophenol	40.000	42.567	-6.4	122	0.00
71	Phenanthrene	40.000	40.736	-1.8	120	0.00
72	Anthracene	40.000	41.452	-3.6	121	0.00
73	Carbazole	40.000	41.251	-3.1	121	0.00
74	Di-n-butylphthalate	40.000	42.066	-5.2	122	0.00
75 C	Fluoranthene	40.000	41.376	-3.4	124	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	123	0.00
77	Benzydine	40.000	29.255	26.9#	76	0.00
78	Pyrene	40.000	38.918	2.7	124	0.00
79 S	Terphenyl-d14	80.000	79.289	0.9	121	0.00
80	Butylbenzylphthalate	40.000	40.320	-0.8	126	0.00
81	Benzo(a)anthracene	40.000	39.078	2.3	124	0.00
82	3,3'-Dichlorobenzidine	40.000	40.491	-1.2	125	0.00
83	Chrysene	40.000	38.705	3.2	124	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.986	-2.5	125	0.00
85 c	Di-n-octyl phthalate	40.000	40.871	-2.2	127	0.00
86 I	Perylene-d12	20.000	20.000	0.0	125	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.983	-2.5	127	0.00
88	Benzo(b)fluoranthene	40.000	40.644	-1.6	127	0.00
89	Benzo(k)fluoranthene	40.000	39.906	0.2	123	0.00
90 C	Benzo(a)pyrene	40.000	40.460	-1.2	125	0.00
91	Dibenzo(a,h)anthracene	40.000	40.697	-1.7	126	0.00
92	Benzo(g,h,i)perylene	40.000	40.371	-0.9	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010325\
Data File : BM049332.D
Acq On : 03 Jan 2025 17:09
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM010325

Quant Time: Jan 03 17:45:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 03 17:24:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0