


Data Path : Z:\HPCHEM1\BNA M\DATA\BM010417\
 Data File : BM008724.D
 Acq On : 04 Jan 2017 09:42
 Operator : UM/SJ
 Sample : DFTPP50
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP50

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\Methods\SOM02.2-EPA-BM121416.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 19 15:00:56 2016



AutoFind: Scans 2527, 2528, 2529; Background Corrected with Scan 2521

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.0	40002	PASS
68	69	0.00	2	1.8	811	PASS
69	198	0.00	100	35.9	44960	PASS
70	69	0.00	2	0.3	141	PASS
127	198	10	80	40.7	50912	PASS
197	198	0.00	2	0.6	703	PASS
198	198	100	100	100.0	125178	PASS
199	198	5	9	6.6	8212	PASS
275	198	10	60	30.4	38082	PASS
365	198	1	100	4.9	6077	PASS
441	443	0.01	100	79.0	17792	PASS
442	198	50	100	92.6	115880	PASS
443	442	15	24	19.4	22528	PASS

Scan 2521 (17.415 min): BM008724.D
DFTPP50

m/z	Abundance
40.00	570.0
44.00	1260.0
73.00	553.0
207.10	1287.0
281.00	493.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.10	638.0
39.10	3212.0
40.00	623.0
41.00	325.0
44.00	1195.0
50.10	12199.0
51.10	46456.0
52.10	2352.0
55.00	456.0
56.00	1469.0
57.00	3095.0
61.10	789.0
62.00	524.0
63.10	2402.0
64.00	438.0
65.00	1064.0
68.10	918.0
69.00	52776.0
69.90	425.0
73.00	968.0
74.10	5712.0
75.00	9092.0
76.10	3132.0
77.10	57400.0
78.10	4196.0
79.00	4766.0
80.00	3336.0
81.00	4727.0
82.00	1354.0
83.00	1005.0
85.00	599.0
86.00	1506.0
87.00	842.0
91.00	1497.0
92.00	1225.0
93.00	7390.0
93.90	478.0
96.20	302.0
98.00	6437.0
99.00	4683.0
100.00	411.0
101.00	2896.0
102.90	688.0
104.00	1670.0
105.10	1194.0
106.00	502.0
107.00	19800.0
108.00	2618.0
109.00	700.0
110.00	29872.0
111.10	4357.0
111.90	774.0
116.10	1283.0
117.00	19160.0
118.00	1405.0
122.00	1171.0
123.00	1844.0
124.00	745.0
125.00	869.0
126.10	316.0
127.00	55584.0
128.00	4513.0
129.00	23968.0
130.00	1822.0
130.90	304.0
132.00	300.0
133.90	894.0
135.00	1922.0
136.00	1043.0
137.00	1042.0
140.00	364.0
141.00	3505.0
142.00	1311.0
143.00	668.0
146.10	468.0
147.00	1811.0
148.00	3970.0
149.10	692.0
150.00	315.0
151.10	362.0
152.00	366.0
153.00	1215.0
154.00	824.0
155.10	1598.0
156.10	2906.0
157.00	481.0
158.00	663.0
159.00	512.0
160.00	1150.0
161.00	1233.0
162.00	429.0
165.00	1573.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

166.00	993.0
167.00	8387.0
168.00	3901.0
169.00	759.0
169.90	368.0
170.80	314.0
172.00	548.0
173.00	833.0
174.10	1552.0
175.00	2534.0
176.10	1076.0
177.00	1594.0
178.00	576.0
179.00	4806.0
180.00	3611.0
181.00	1578.0
182.00	341.0
183.90	371.0
185.10	2324.0
186.00	18296.0
187.00	5754.0
188.10	673.0
189.00	1685.0
190.00	360.0
191.00	805.0
191.90	2025.0
193.00	2076.0
194.10	566.0
196.00	2938.0
197.10	713.0
198.00	126040.0
199.00	7915.0
200.10	829.0
201.30	443.0
203.00	1373.0
204.00	5445.0
205.00	8161.0
206.10	34400.0
207.00	6356.0
208.00	1955.0
209.00	510.0
210.00	453.0
210.90	1376.0
211.80	327.0
215.00	524.0
216.10	898.0
217.00	10525.0
218.00	1406.0
221.00	4603.0
221.90	1299.0
223.00	2210.0
224.00	20896.0
225.10	4677.0
225.90	700.0
227.00	9740.0
228.00	1124.0
229.00	1569.0
231.10	609.0
234.00	614.0
235.00	653.0
236.00	641.0
237.00	1499.0
239.00	701.0
241.10	558.0
242.00	1115.0
243.00	1082.0
244.10	15877.0
245.00	2380.0
246.00	2800.0
246.90	621.0
248.90	428.0
253.10	346.0
254.10	632.0
255.00	83200.0
256.00	12340.0
257.00	1192.0
258.00	5339.0
259.00	804.0
265.00	2349.0
265.90	320.0
271.10	340.0
273.00	2357.0
274.00	6255.0
275.00	35400.0
276.00	4559.0
277.00	2883.0
278.00	512.0
281.00	656.0
283.00	411.0
285.00	624.0
293.00	523.0
293.90	353.0
295.00	359.0
296.00	12255.0
297.00	1869.0
303.10	1578.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

313.00	316.0
314.10	560.0
315.00	1192.0
316.00	708.0
321.00	358.0
323.10	3541.0
324.10	574.0
327.00	582.0
328.00	566.0
333.10	368.0
334.10	2243.0
335.00	749.0
341.00	502.0
346.00	862.0
351.90	1076.0
352.90	625.0
354.00	1166.0
355.10	641.0
365.00	5384.0
365.90	835.0
372.10	1464.0
372.90	427.0
382.90	452.0
402.10	761.0
402.90	634.0
404.00	471.0
420.90	855.0
422.00	703.0
423.10	5444.0
424.00	1702.0
441.10	14457.0
442.00	91784.0
443.10	18368.0
444.00	1731.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.10	865.0
39.10	2692.0
40.00	669.0
41.10	363.0
44.00	1042.0
49.20	564.0
50.10	11595.0
51.10	45600.0
52.10	2298.0
55.00	347.0
56.00	1486.0
57.00	3405.0
60.90	681.0
62.00	767.0
63.00	1516.0
64.00	423.0
65.10	907.0
68.00	750.0
69.00	51816.0
73.10	1101.0
74.00	5994.0
75.00	8224.0
76.00	2504.0
77.10	59616.0
78.10	4263.0
79.10	4117.0
80.00	2944.0
81.00	4252.0
82.00	1366.0
83.00	944.0
85.00	835.0
86.00	1632.0
87.00	703.0
88.00	363.0
91.00	1299.0
91.90	989.0
93.00	8513.0
94.00	466.0
96.10	445.0
98.00	5676.0
99.00	4581.0
101.00	2359.0
103.10	753.0
104.00	2002.0
104.90	1513.0
105.80	418.0
107.00	19552.0
108.10	2563.0
109.20	515.0
110.00	32344.0
111.00	5253.0
112.00	1076.0
115.90	1082.0
117.00	19536.0
118.00	1633.0
120.00	430.0
122.00	1251.0
123.00	2232.0
124.00	825.0
125.00	794.0
127.00	60480.0
128.00	4983.0
129.00	25248.0
130.00	2622.0
130.90	433.0
134.00	846.0
135.00	2613.0
136.00	1130.0
137.00	1353.0
140.10	312.0
141.00	3806.0
142.00	1211.0
143.00	793.0
145.90	775.0
147.00	2077.0
148.00	5001.0
149.10	1172.0
151.20	737.0
151.90	301.0
152.90	978.0
154.00	777.0
155.00	2323.0
156.00	2298.0
157.10	671.0
157.90	912.0
158.90	689.0
160.10	1347.0
161.00	1778.0
162.10	687.0
165.10	1335.0
166.00	917.0
167.00	7253.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

168.00	4911.0
169.00	514.0
170.10	331.0
170.90	310.0
172.00	800.0
172.90	748.0
174.00	1601.0
175.10	3006.0
175.90	713.0
177.00	1745.0
177.90	527.0
179.00	5686.0
180.00	4477.0
180.90	1517.0
181.90	346.0
183.90	589.0
185.00	3272.0
186.00	21120.0
187.00	6230.0
188.00	533.0
189.00	1360.0
191.00	729.0
192.00	2009.0
193.10	2803.0
194.00	404.0
195.10	325.0
196.00	4270.0
197.10	985.0
198.00	150976.0
199.00	8968.0
200.00	748.0
201.50	617.0
203.00	1289.0
204.00	6473.0
205.10	10414.0
206.00	42816.0
207.00	7006.0
208.00	2252.0
209.00	754.0
210.00	516.0
210.30	521.0
211.10	1441.0
212.00	505.0
215.00	607.0
216.00	1338.0
217.00	11522.0
218.00	1250.0
221.10	5766.0
221.80	1619.0
223.00	2422.0
224.10	25384.0
225.00	6639.0
226.00	772.0
227.00	11503.0
227.90	1228.0
228.90	2510.0
231.10	875.0
234.10	663.0
235.00	748.0
236.00	561.0
237.00	1016.0
239.00	378.0
240.00	370.0
241.00	532.0
242.10	1441.0
243.00	1414.0
244.10	19600.0
245.10	2148.0
246.00	4839.0
247.00	801.0
249.00	978.0
252.90	626.0
254.10	900.0
255.00	105488.0
256.00	14709.0
257.00	1051.0
258.00	6107.0
258.90	1219.0
265.00	3063.0
272.00	410.0
273.00	3154.0
274.00	8456.0
275.00	48280.0
276.00	6769.0
277.00	4555.0
278.00	625.0
281.00	683.0
283.00	847.0
285.00	595.0
293.00	1088.0
295.20	546.0
296.00	14450.0
297.00	2266.0
303.00	1573.0
304.00	515.0
313.10	425.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

314.00	752.0
314.90	1717.0
316.00	733.0
321.00	401.0
321.90	377.0
323.00	4074.0
324.00	689.0
326.90	1223.0
328.00	549.0
334.10	3062.0
335.00	702.0
341.10	442.0
345.90	717.0
352.00	1248.0
353.00	672.0
354.00	1454.0
365.00	7285.0
366.00	767.0
371.00	354.0
372.00	2284.0
373.00	602.0
383.00	659.0
389.90	379.0
402.00	893.0
403.00	1292.0
404.00	611.0
421.10	851.0
422.10	1260.0
423.10	7577.0
424.00	1765.0
441.10	21088.0
442.10	142272.0
443.10	27216.0
444.10	3003.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
37.90	356.0
39.10	2134.0
40.00	621.0
43.20	332.0
44.10	1287.0
50.00	7642.0
51.10	27952.0
52.00	1474.0
55.20	401.0
55.90	684.0
57.10	1973.0
60.90	483.0
62.00	689.0
63.10	1170.0
65.10	560.0
67.90	766.0
69.00	30288.0
73.00	964.0
74.00	3532.0
75.00	5001.0
76.10	1568.0
77.10	32800.0
78.00	2664.0
79.00	2530.0
80.00	1784.0
81.10	2837.0
82.00	670.0
83.00	387.0
85.00	723.0
86.00	866.0
86.90	439.0
91.10	598.0
92.00	889.0
93.00	4876.0
94.10	329.0
96.00	302.0
98.10	3704.0
99.00	3642.0
100.00	320.0
101.00	1343.0
103.00	556.0
104.00	1216.0
105.00	946.0
106.10	378.0
107.10	12259.0
108.00	2075.0
109.20	385.0
110.00	20720.0
111.10	2862.0
112.10	332.0
116.10	754.0
117.00	12723.0
118.00	805.0
121.90	971.0
123.00	1337.0
123.90	475.0
124.90	709.0
127.00	36672.0
128.00	2941.0
129.00	14259.0
130.00	1029.0
133.90	618.0
135.00	1260.0
136.00	570.0
136.90	783.0
140.00	313.0
140.90	2335.0
142.00	605.0
143.00	590.0
146.10	327.0
147.00	1453.0
148.00	3133.0
149.00	507.0
153.10	919.0
153.90	750.0
155.00	1331.0
156.00	1816.0
157.10	578.0
158.10	548.0
158.90	312.0
160.10	884.0
161.00	1140.0
165.00	1232.0
166.00	707.0
167.00	4870.0
168.00	2580.0
169.00	528.0
171.90	577.0
172.90	350.0
174.00	1101.0
175.00	2273.0
176.10	461.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

176.90	1125.0
178.10	350.0
178.90	3879.0
180.00	2616.0
181.00	1414.0
181.80	323.0
183.90	342.0
185.10	1971.0
186.00	14779.0
187.10	4308.0
188.00	525.0
188.90	987.0
191.00	680.0
192.00	1349.0
193.00	1755.0
194.00	336.0
196.00	2616.0
197.10	412.0
198.00	98520.0
199.00	7754.0
200.00	650.0
201.30	420.0
203.00	932.0
204.00	4274.0
205.00	6919.0
206.00	27248.0
207.10	4941.0
208.00	1146.0
208.90	405.0
209.90	457.0
211.00	888.0
212.10	347.0
214.90	600.0
215.90	576.0
217.00	8181.0
217.90	1168.0
221.00	4778.0
222.10	803.0
223.00	2176.0
224.00	16848.0
225.00	4316.0
226.10	354.0
227.00	7947.0
228.00	1049.0
229.00	1389.0
231.10	617.0
233.90	412.0
234.90	528.0
237.10	689.0
239.00	309.0
240.90	590.0
242.00	829.0
243.10	1021.0
244.00	12265.0
245.00	1728.0
246.00	2982.0
246.90	434.0
249.00	527.0
252.90	333.0
253.20	329.0
254.20	462.0
255.00	72712.0
256.00	9872.0
257.10	847.0
258.00	4634.0
259.00	558.0
265.00	2023.0
265.80	301.0
273.00	2365.0
274.00	6268.0
275.00	30568.0
276.00	3933.0
277.00	3478.0
277.90	616.0
281.00	580.0
282.90	466.0
284.00	361.0
284.90	480.0
292.90	622.0
296.00	11296.0
297.10	1541.0
303.10	852.0
303.90	483.0
313.90	443.0
315.00	1616.0
316.10	823.0
320.90	333.0
323.00	3469.0
323.90	970.0
327.00	582.0
327.90	433.0
332.90	406.0
334.10	2225.0
335.10	569.0
340.90	490.0
346.00	623.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

351.90	1077.0
353.00	627.0
354.00	1075.0
355.00	450.0
365.00	5564.0
366.00	600.0
372.00	1761.0
373.10	393.0
383.00	549.0
389.90	342.0
401.90	714.0
403.00	906.0
403.90	327.0
421.10	846.0
421.90	731.0
423.10	5790.0
424.00	1198.0
441.00	17832.0
442.00	113584.0
443.10	22000.0
444.00	1936.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50