

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010523\
 Data File : BM038375.D
 Acq On : 05 Jan 2023 12:22
 Operator : CG/JU
 Sample : PB149946BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK946

Quant Time: Jan 05 13:25:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:11:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	71264	20.000	ng/u1	0.00
20) Naphthalene-d8	10.810	136	266845	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.633	164	161024	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.374	188	364395	20.000	ng/u1	0.00
79) Chrysene-d12	21.544	240	349637	20.000	ng/u1	0.00
88) Perylene-d12	23.980	264	454633	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	11416	5.195	ng/uL	0.00
4) Pyridine-d5	3.799	84	149091	23.451	ng/u1	0.00
7) Phenol-d5	7.157	99	148964	22.946	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	111850	23.950	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	116009	24.940	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	110266	22.004	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	50247	24.001	ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	58290	24.550	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.433	165	103552	22.584	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	131364	20.762	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	301562	24.729	ng/u1	0.00
49) Acenaphthylene-d8	14.327	160	341938	24.061	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	39902	18.176	ng/u1	0.00
60) Fluorene-d10	15.621	176	264391	23.840	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	48876	18.507	ng/u1	0.00
73) Anthracene-d10	17.474	188	413267	23.510	ng/u1	0.00
81) Pyrene-d10	19.756	212	500723	25.736	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.821	264	525988	23.294	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010523\
Data File : BM038375.D
Acq On : 05 Jan 2023 12:22
Operator : CG/JU
Sample : PB149946BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK946

Quant Time: Jan 05 13:25:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 02:11:50 2023
Response via : Initial Calibration

