

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010622\
 Data File : BM033663.D
 Acq On : 06 Jan 2022 14:24
 Operator : CG/JU
 Sample : SSTD16023
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160023

Quant Time: Jan 06 15:20:56 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 15:17:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.984	152	199998	20.000	ng/ul	0.00
20) Naphthalene-d8	10.807	136	877345	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.624	164	580269	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.359	188	1223969	20.000	ng/ul	0.00
79) Chrysene-d12	21.530	240	929486	20.000	ng/ul	0.01
88) Perylene-d12	23.959	264	839807	20.000	ng/ul	0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.319	96	306350	57.562	ng/uL	0.00
4) Pyridine-d5	3.749	84	2270326	147.128	ng/ul	0.00
7) Phenol-d5	7.148	99	2782084	146.904	ng/ul	0.02
9) Bis-(2-Chloroethyl)eth...	7.313	67	1611396	129.968	ng/ul	0.01
11) 2-Chlorophenol-d4	7.519	132	2216971	167.200	ng/ul	0.01
15) 4-Methylphenol-d8	8.707	113	2274839	153.462	ng/ul	0.02
21) Nitrobenzene-d5	9.160	128	1147116	161.137	ng/ul	0.02
24) 2-Nitrophenol-d4	9.883	143	1289342	176.530	ng/ul	0.01
28) 2,4-Dichlorophenol-d3	10.431	165	2258851	163.651	ng/ul	0.02
31) 4-Chloroaniline-d4	10.942	131	3180504	155.414	ng/ul	0.02
46) Dimethylphthalate-d6	14.048	166	6736468	155.375	ng/ul	0.02
49) Acenaphthylene-d8	14.324	160	7994214	148.738	ng/ul	0.01
54) 4-Nitrophenol-d4	14.836	143	1336626	169.793	ng/ul	0.04
60) Fluorene-d10	15.618	176	5556532	143.279	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.742	200	1277722	172.960	ng/ul	0.03
73) Anthracene-d10	17.471	188	8614969	142.420	ng/ul	0.02
81) Pyrene-d10	19.748	212	9175718	176.630	ng/ul	0.01
92) Benzo(a)pyrene-d12	23.812	264	6944367	152.557	ng/ul	0.02

Target Compounds					Qvalue	
2) 1,4-Dioxane	3.360	88	333337	56.270	ng/uL#	86
5) Pyridine	3.772	79	2297948	144.593	ng/ul	86
6) Benzaldehyde	7.113	77	1274615	120.806	ng/ul	84
8) Phenol	7.178	94	2799383	143.505	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.413	93	2114543	144.089	ng/ul	91
12) 2-Chlorophenol	7.554	128	2261870	164.965	ng/ul#	88
13) 2-Methylphenol	8.437	108	2153509	152.268	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.531	45	2876057	113.579	ng/ul#	94
16) Acetophenone	8.825	105	3387687	137.959	ng/ul#	87
17) N-Nitroso-di-n-propyla...	8.842	70	1738687	129.537	ng/ul#	84
18) 4-Methylphenol	8.789	108	2342700	151.399	ng/ul	98
19) Hexachloroethane	9.084	117	970544	140.108	ng/ul#	82
22) Nitrobenzene	9.207	77	2681034	128.635	ng/ul#	89
23) Isophorone	9.754	82	5128194	143.690	ng/ul	96
25) 2-Nitrophenol	9.919	139	1325544	171.188	ng/ul#	83
26) 2,4-Dimethylphenol	9.983	107	2774413	147.774	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.219	93	2928527	145.900	ng/ul	98
29) 2,4-Dichlorophenol	10.460	162	2221352	158.980	ng/ul	97
30) Naphthalene	10.860	128	7081103	144.832	ng/ul	99
32) 4-Chloroaniline	10.972	127	3160242	153.648	ng/ul	98
33) Hexachlorobutadiene	11.154	225	1534043	149.989	ng/ul	96
34) Caprolactam	11.801	113	436572	92.649	ng/ul	80
35) 4-Chloro-3-methylphenol	12.101	107	2644576	161.241	ng/ul	89
36) 2-Methylnaphthalene	12.466	142	4870804	146.904	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010622\
 Data File : BM033663.D
 Acq On : 06 Jan 2022 14:24
 Operator : CG/JU
 Sample : SSTD16023
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD160023


Quant Time: Jan 06 15:20:56 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 15:17:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.683	142	4930077	143.078	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.836	216	2698916	154.373	ng/ul	98
40) Hexachlorocyclopentadiene	12.819	237	1895167	166.088	ng/ul	99
41) 2,4,6-Trichlorophenol	13.072	196	1824587	177.298	ng/ul	98
42) 2,4,5-Trichlorophenol	13.142	196	1992746	178.741	ng/ul	92
43) 1,1'-Biphenyl	13.472	154	6502836	146.485	ng/ul	98
44) 2-Chloronaphthalene	13.513	162	5074816	148.584	ng/ul	97
45) 2-Nitroaniline	13.713	65	1829617	150.209	ng/ul#	79
47) Dimethylphthalate	14.095	163	6611462	153.581	ng/ul	99
48) 2,6-Dinitrotoluene	14.213	165	1475302	177.287	ng/ul	91
50) Acenaphthylene	14.360	152	8131839	145.315	ng/ul	97
51) 3-Nitroaniline	14.536	138	1143772	139.637	ng/ul	86
52) Acenaphthene	14.695	153	5405127	145.535	ng/ul	98
53) 2,4-Dinitrophenol	14.736	184	978538	200.095	ng/ul	86
55) 4-Nitrophenol	14.854	109	1250402	146.963	ng/ul	84
56) Dibenzofuran	15.030	168	7656737	142.166	ng/ul	98
57) 2,4-Dinitrotoluene	14.995	165	2156410	176.397	ng/ul#	78
58) 2,3,4,6-Tetrachlorophenol	15.254	232	1734950	183.061	ng/ul	97
59) Diethylphthalate	15.454	149	7062505	158.071	ng/ul	98
61) Fluorene	15.677	166	6164092	139.197	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.665	204	3245767	146.933	ng/ul	96
63) 4-Nitroaniline	15.701	138	902140	107.252	ng/ul	87
66) 4,6-Dinitro-2-methylph...	15.760	198	1231938	167.410	ng/ul#	93
67) N-Nitrosodiphenylamine	15.877	169	5492673	152.626	ng/ul	98
68) 4-Bromophenyl-phenylether	16.560	248	2022428	164.051	ng/ul	95
69) Hexachlorobenzene	16.677	284	2356203	165.881	ng/ul	94
70) Atrazine	16.842	200	2260327	158.291	ng/ul	100
71) Pentachlorophenol	17.012	266	1628025	204.398	ng/ul	97
72) Phenanthrene	17.412	178	9648629	136.647	ng/ul	96
74) Anthracene	17.507	178	9654039	134.856	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.436	216	2928931	165.920	ng/ul	98
76) Pentachlorobenzene	14.954	250	2762482	158.468	ng/ul	99
77) Carbazole	17.765	167	8890713	137.407	ng/ul	96
78) Di-n-butylphthalate	18.336	149	11271630	155.255	ng/ul	97
80) Fluoranthene	19.418	202	10741647	175.423	ng/ul	99
82) Pyrene	19.783	202	10515281	164.035	ng/ul	99
83) Butylbenzylphthalate	20.665	149	5096733	207.563	ng/ul	86
84) 3,3'-Dichlorobenzidine	21.442	252	2792190	132.167	ng/ul	96
85) Benzo(a)anthracene	21.512	228	9250333	151.473	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.442	149	7083814	200.724	ng/ul#	95
87) Chrysene	21.571	228	8766551	146.059	ng/ul	97
89) Di-n-octyl phthalate	22.377	149	11073421	179.840	ng/ul	100
90) Benzo(b)fluoranthene	23.224	252	8767971	152.383	ng/ul	99
91) Benzo(k)fluoranthene	23.277	252	8050132	151.102	ng/ul	99
93) Benzo(a)pyrene	23.871	252	8234481	149.149	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.529	276	9227874	153.689	ng/ul	97
95) Dibenzo(a,h)anthracene	26.553	278	7919866	151.641	ng/ul	97
96) Benzo(g,h,i)perylene	27.324	276	7804186	152.390	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 06 15:20:56 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM010622.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 06 15:17:08 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 56.270 ng/uL
RT: 3.360 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion: 88 Resp: 333337
Ion Ratio Lower Upper
88 100
43 29.7 36.2 54.4#
58 78.0 68.5 102.7



#5
Pyridine
Concen: 144.593 ng/uL
RT: 3.772 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion: 79 Resp: 2297948
Ion Ratio Lower Upper
79 100
52 81.7 78.1 117.1
51 39.4 36.0 54.0









#6
Benzaldehyde
Concen: 120.806 ng/ul
RT: 7.113 min Scan# 61
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion: 77 Resp: 1274615
Ion Ratio Lower Upper
77 100
105 95.9 65.6 98.4
106 90.1 60.6 90.8



#8
Phenol
Concen: 143.505 ng/ul
RT: 7.178 min Scan# 670
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion: 94 Resp: 2799383
Ion Ratio Lower Upper
94 100
65 31.6 31.7 47.5#
66 45.5 45.4 68.0









#10
Bis(2-Chloroethyl)ether
Concen: 144.089 ng/ul
RT: 7.413 min Scan# 710
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion: 93 Resp: 2114543
Ion Ratio Lower Upper
93 100
63 78.4 71.7 107.5
95 32.3 26.3 39.5



#12
2-Chlorophenol
Concen: 164.965 ng/ul
RT: 7.554 min Scan# 734
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:128 Resp: 2261870
Ion Ratio Lower Upper
128 100
64 48.4 49.4 74.2#
130 32.8 26.8 40.2









#13
2-Methylphenol
Concen: 152.268 ng/ul
RT: 8.437 min Scan# 881
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:108 Resp: 2153509
Ion Ratio Lower Upper
108 100
107 90.1 70.2 105.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 113.579 ng/ul
RT: 8.531 min Scan# 900
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion: 45 Resp: 2876057
Ion Ratio Lower Upper
45 100
77 15.5 9.9 14.9#
79 11.9 8.3 12.5









#16
Acetophenone
Concen: 137.959 ng/ul
RT: 8.825 min Scan# 91
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:105 Resp: 3387687
Ion Ratio Lower Upper
105 100
77 79.5 73.1 109.7
51 29.5 30.5 45.7#



#17
N-Nitroso-di-n-propylamine
Concen: 129.537 ng/ul
RT: 8.842 min Scan# 953
Delta R.T. 0.047 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion: 70 Resp: 1738687
Ion Ratio Lower Upper
70 100
42 55.9 56.9 85.3#
101 10.4 6.9 10.3#
130 21.7 13.8 20.8#









#18
4-Methylphenol
Concen: 151.399 ng/ul
RT: 8.789 min Scan# 944
Delta R.T. 0.035 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:108 Resp: 2342700
Ion Ratio Lower Upper
108 100
107 111.3 87.4 131.2



#19
Hexachloroethane
Concen: 140.108 ng/ul
RT: 9.084 min Scan# 994
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:117 Resp: 970544
Ion Ratio Lower Upper
117 100
201 102.0 69.6 104.4
199 64.6 40.6 60.8#









#22
Nitrobenzene
Concen: 128.635 ng/ul
RT: 9.207 min Scan# 1108
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion: 77 Resp: 2681034
Ion Ratio Lower Upper
77 100
123 45.1 29.0 43.6#
65 14.9 11.4 17.2



#23
Isophorone
Concen: 143.690 ng/ul
RT: 9.754 min Scan# 1108
Delta R.T. 0.035 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion: 82 Resp: 5128194
Ion Ratio Lower Upper
82 100
95 8.1 6.5 9.7
138 17.2 11.8 17.8









#25
2-Nitrophenol
Concen: 171.188 ng/ul
RT: 9.919 min Scan# 1136
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:139 Resp: 1325544
Ion Ratio Lower Upper
139 100
65 57.0 59.0 88.6#
109 38.7 37.2 55.8



#26
2,4-Dimethylphenol
Concen: 147.774 ng/ul
RT: 9.983 min Scan# 1147
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:107 Resp: 2774413
Ion Ratio Lower Upper
107 100
121 48.0 34.8 52.2
122 79.4 60.0 90.0





#27

Bis(2-Chloroethoxy)methane

Concen: 145.900 ng/ul

RT: 10.219 min Scan# 1187

Delta R.T. 0.018 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :





BNA_M

ClientSampleId :

SSTD160023

Tgt Ion: 93 Resp: 2928527

Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.3	37.9
123	11.2	7.9	11.9



#29

2,4-Dichlorophenol

Concen: 158.980 ng/ul

RT: 10.460 min Scan# 1228




Delta R.T. 0.018 min


Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Tgt Ion:162 Resp: 2221352

Ion	Ratio	Lower	Upper
162	100		
164	65.3	50.0	75.0
98	39.3	30.6	45.8









#30
Naphthalene
Concen: 144.832 ng/ul
RT: 10.860 min Scan# 1315
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:128 Resp: 7081103
Ion Ratio Lower Upper
128 100
129 11.3 9.1 13.7
127 13.7 10.2 15.4



#32
4-Chloroaniline
Concen: 153.648 ng/ul
RT: 10.972 min Scan# 1315
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:127 Resp: 3160242
Ion Ratio Lower Upper
127 100
129 32.7 25.4 38.0









#33
Hexachlorobutadiene
Concen: 149.989 ng/ul
RT: 11.154 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:225 Resp: 1534043
Ion Ratio Lower Upper
225 100
223 62.3 52.2 78.2
227 62.7 52.8 79.2



#34
Caprolactam
Concen: 92.649 ng/ul
RT: 11.801 min Scan# 1456
Delta R.T. 0.088 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:113 Resp: 436572
Ion Ratio Lower Upper
113 100
55 168.3 157.9 236.9
56 135.6 131.8 197.6









#35
4-Chloro-3-methylphenol
Concen: 161.241 ng/ul
RT: 12.101 min Scan# 1507
Delta R.T. 0.029 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:107 Resp: 2644576
Ion Ratio Lower Upper
107 100
144 23.2 17.1 25.7
142 71.9 49.4 74.2



#36
2-Methylnaphthalene
Concen: 146.904 ng/ul
RT: 12.466 min Scan# 1569
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:142 Resp: 4870804
Ion Ratio Lower Upper
142 100
141 90.5 71.1 106.7





#37

1-Methylnaphthalene

Concen: 143.078 ng/ul

RT: 12.683 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:142 Resp: 4930077

Ion Ratio Lower Upper

142 100

141 93.3 75.5 113.3

116 4.5 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 154.373 ng/ul

RT: 12.836 min Scan# 1632

Delta R.T. 0.012 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Tgt Ion:216 Resp: 2698916




Ion Ratio Lower Upper


216 100

214 78.2 61.3 91.9

179 22.4 18.0 27.0

108 18.0 16.6 25.0









#40
Hexachlorocyclopentadiene
Concen: 166.088 ng/ul
RT: 12.819 min Scan# 1628
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:237 Resp: 1895167
Ion Ratio Lower Upper
237 100
235 62.4 50.6 76.0
272 12.1 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 177.298 ng/ul
RT: 13.072 min Scan# 1672
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:196 Resp: 1824587
Ion Ratio Lower Upper
196 100
198 97.0 78.9 118.3
200 31.6 23.7 35.5









#42
2,4,5-Trichlorophenol
Concen: 178.741 ng/ul
RT: 13.142 min Scan# 1
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:196 Resp: 1992746
Ion Ratio Lower Upper
196 100
198 94.8 82.2 123.2
200 30.2 27.9 41.9



#43
1,1'-Biphenyl
Concen: 146.485 ng/ul
RT: 13.472 min Scan# 1740
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:154 Resp: 6502836
Ion Ratio Lower Upper
154 100
153 42.3 33.1 49.7
76 12.2 10.7 16.1









#44
2-Chloronaphthalene
Concen: 148.584 ng/ul
RT: 13.513 min Scan# 1745
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:162 Resp: 5074816
Ion Ratio Lower Upper
162 100
164 32.9 26.2 39.4
127 44.9 33.7 50.5



#45
2-Nitroaniline
Concen: 150.209 ng/ul
RT: 13.713 min Scan# 1781
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion: 65 Resp: 1829617
Ion Ratio Lower Upper
65 100
92 67.6 47.8 71.6
138 101.2 61.0 91.6#









#47
Dimethylphthalate
Concen: 153.581 ng/ul
RT: 14.095 min Scan# 1846
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:163 Resp: 6611462
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.5 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 177.287 ng/ul
RT: 14.213 min Scan# 1866
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:165 Resp: 1475302
Ion Ratio Lower Upper
165 100
89 61.5 55.7 83.5
121 22.0 20.0 30.0









#50
Acenaphthylene
Concen: 145.315 ng/ul
RT: 14.360 min Scan# 1888
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:152 Resp: 8131839
Ion Ratio Lower Upper
152 100
151 21.4 15.8 23.8
153 13.7 10.1 15.1



#51
3-Nitroaniline
Concen: 139.637 ng/ul
RT: 14.536 min Scan# 1921
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:138 Resp: 1143772
Ion Ratio Lower Upper
138 100
108 11.8 10.8 16.2
92 119.7 110.6 166.0





#52

Acenaphthene

Concen: 145.535 ng/ul

RT: 14.695 min Scan# 1948

Delta R.T. 0.012 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:153 Resp: 5405127

Ion Ratio Lower Upper

153 100

152 48.3 39.4 59.2

154 88.9 72.7 109.1



#53

2,4-Dinitrophenol

Concen: 200.095 ng/ul

RT: 14.736 min Scan# 1955

Delta R.T. 0.024 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24




Tgt Ion:184 Resp: 978538


Ion Ratio Lower Upper

184 100

63 74.0 73.8 110.8

154 72.8 63.5 95.3









#55
4-Nitrophenol
Concen: 146.963 ng/ul
RT: 14.854 min Scan# 1967
Delta R.T. 0.047 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:109 Resp: 1250402
Ion Ratio Lower Upper
109 100
139 109.4 77.6 116.4
65 119.0 112.2 168.4



#56
Dibenzofuran
Concen: 142.166 ng/ul
RT: 15.030 min Scan# 2005
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:168 Resp: 7656737
Ion Ratio Lower Upper
168 100
139 43.9 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 176.397 ng/ul
RT: 14.995 min Scan# 1999
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:165 Resp: 2156410
Ion Ratio Lower Upper
165 100
63 46.4 51.8 77.8#
182 2.8 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 183.061 ng/ul
RT: 15.254 min Scan# 2043
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:232 Resp: 1734950
Ion Ratio Lower Upper
232 100
131 49.6 41.4 62.2
130 2.6 2.0 3.0
166 32.7 27.4 41.0









#59
Diethylphthalate
Concen: 158.071 ng/ul
RT: 15.454 min Scan# 2074
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:149 Resp: 7062505
Ion Ratio Lower Upper
149 100
177 22.3 17.1 25.7
150 13.2 9.9 14.9



#61
Fluorene
Concen: 139.197 ng/ul
RT: 15.677 min Scan# 2115
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:166 Resp: 6164092
Ion Ratio Lower Upper
166 100
165 98.7 77.9 116.9
167 13.7 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 146.933 ng/ul
RT: 15.665 min Scan# 2112
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:204 Resp: 3245767
Ion Ratio Lower Upper
204 100
206 32.5 27.8 41.8
141 65.6 54.9 82.3



#63
4-Nitroaniline
Concen: 107.252 ng/ul
RT: 15.701 min Scan# 2119
Delta R.T. 0.035 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:138 Resp: 902140
Ion Ratio Lower Upper
138 100
92 54.5 48.0 72.0
108 94.6 89.5 134.3









#66
4,6-Dinitro-2-methylphenol
Concen: 167.410 ng/ul
RT: 15.760 min Scan# 2129
Delta R.T. 0.029 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:198 Resp: 1231938
Ion Ratio Lower Upper
198 100
182 3.6 4.2 6.2#
77 28.6 26.3 39.5



#67
N-Nitrosodiphenylamine
Concen: 152.626 ng/ul
RT: 15.877 min Scan# 2149
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:169 Resp: 5492673
Ion Ratio Lower Upper
169 100
168 68.4 52.9 79.3
167 37.5 29.9 44.9









#68
4-Bromophenyl-phenylether
Concen: 164.051 ng/ul
RT: 16.560 min Scan# 21
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:248 Resp: 2022428
Ion Ratio Lower Upper
248 100
250 97.3 79.0 118.4
141 79.1 69.7 104.5



#69
Hexachlorobenzene
Concen: 165.881 ng/ul
RT: 16.677 min Scan# 2285
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:284 Resp: 2356203
Ion Ratio Lower Upper
284 100
142 37.1 34.3 51.5
249 33.4 26.1 39.1





#70

Atrazine

Concen: 158.291 ng/ul

RT: 16.842 min Scan# 21

Delta R.T. 0.029 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:200 Resp: 2260327

Ion Ratio Lower Upper

200 100

173 25.9 20.6 30.8

215 46.2 36.9 55.3



#71

Pentachlorophenol

Concen: 204.398 ng/ul

RT: 17.012 min Scan# 2342

Delta R.T. 0.006 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24




Tgt Ion:266 Resp: 1628025


Ion Ratio Lower Upper

266 100

264 62.9 49.1 73.7

268 63.3 47.8 71.8









#72
Phenanthrene
Concen: 136.647 ng/ul
RT: 17.412 min Scan# 2410
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:178 Resp: 9648629
Ion Ratio Lower Upper
178 100
179 16.3 11.7 17.5
176 21.0 15.1 22.7



#74
Anthracene
Concen: 134.856 ng/ul
RT: 17.507 min Scan# 2426
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:178 Resp: 9654039
Ion Ratio Lower Upper
178 100
179 16.9 12.3 18.5
176 20.6 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 165.920 ng/uL

RT: 13.436 min Scan# 1734

Delta R.T. 0.012 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:216 Resp: 2928931

Ion Ratio Lower Upper

216 100

214 79.2 65.4 98.0

218 48.8 38.8 58.2



#76

Pentachlorobenzene

Concen: 158.468 ng/uL

RT: 14.954 min Scan# 1992

Delta R.T. 0.012 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24




Tgt Ion:250 Resp: 2762482


Ion Ratio Lower Upper

250 100

248 63.6 51.6 77.4

252 63.9 52.0 78.0









#77
Carbazole
Concen: 137.407 ng/ul
RT: 17.765 min Scan# 2468
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:167 Resp: 8890713
Ion Ratio Lower Upper
167 100
166 23.1 16.9 25.3
139 15.1 10.5 15.7



#78
Di-n-butylphthalate
Concen: 155.255 ng/ul
RT: 18.336 min Scan# 2567
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:149 Resp:11271630
Ion Ratio Lower Upper
149 100
150 10.2 7.0 10.6
104 6.1 5.1 7.7





#80

Fluoranthene

Concen: 175.423 ng/ul

RT: 19.418 min Scan# 21

Delta R.T. 0.012 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :





BNA_M

ClientSampleId :

SSTD160023

Tgt Ion:202 Resp:10741647

Ion	Ratio	Lower	Upper
202	100		
101	10.6	8.2	12.4
100	8.2	6.9	10.3



#82

Pyrene

Concen: 164.035 ng/ul

RT: 19.783 min Scan# 2813




Delta R.T. 0.018 min


Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Tgt Ion:202 Resp:10515281

Ion	Ratio	Lower	Upper
202	100		
101	12.9	9.8	14.6
100	11.0	8.6	13.0





#83

Butylbenzylphthalate

Concen: 207.563 ng/ul

RT: 20.665 min Scan# 2963

Delta R.T. 0.006 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:149 Resp: 5096733

Ion Ratio Lower Upper

149 100

91 71.3 69.3 103.9

206 21.1 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 132.167 ng/ul

RT: 21.442 min Scan# 3095

Delta R.T. 0.012 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24




Tgt Ion:252 Resp: 2792190


Ion Ratio Lower Upper

252 100

254 65.5 49.4 74.2

126 9.5 8.2 12.4









#85
Benzo(a)anthracene
Concen: 151.473 ng/ul
RT: 21.512 min Scan# 3107
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:228 Resp: 9250333
Ion Ratio Lower Upper
228 100
229 21.3 15.0 22.4
226 29.8 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 200.724 ng/ul
RT: 21.442 min Scan# 3095
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:149 Resp: 7083814
Ion Ratio Lower Upper
149 100
167 30.0 21.8 32.8
279 5.5 3.2 4.8#









#87
Chrysene
Concen: 146.059 ng/ul
RT: 21.571 min Scan# 3115
Delta R.T. 0.012 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:228 Resp: 8766551
Ion Ratio Lower Upper
228 100
226 32.8 24.5 36.7
229 21.2 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 179.840 ng/ul
RT: 22.377 min Scan# 3254
Delta R.T. 0.006 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:149 Resp:11073421









#90
Benzo(b)fluoranthene
Concen: 152.383 ng/ul
RT: 23.224 min Scan# 31
Delta R.T. 0.018 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24




Instrument :
BNA_M
ClientSampleId :
SSTD160023


Tgt Ion:252 Resp: 8767971
Ion Ratio Lower Upper
252 100
253 22.6 17.7 26.5
125 9.5 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 151.102 ng/ul
RT: 23.277 min Scan# 3407
Delta R.T. 0.024 min
Lab File: BM033663.D
Acq: 06 Jan 2022 14:24

Tgt Ion:252 Resp: 8050132
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 9.2 7.4 11.2





#93

Benzo(a)pyrene

Concen: 149.149 ng/ul

RT: 23.871 min Scan# 3

Delta R.T. 0.029 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:252 Resp: 8234481

Ion Ratio Lower Upper

252 100

253 22.2 17.4 26.0

125 10.5 7.8 11.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 153.689 ng/ul

RT: 26.529 min Scan# 3960

Delta R.T. 0.041 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24




Tgt Ion:276 Resp: 9227874


Ion Ratio Lower Upper

276 100

138 22.9 16.5 24.7

277 25.5 20.6 31.0





#95

Dibenzo(a,h)anthracene

Concen: 151.641 ng/ul

RT: 26.553 min Scan# 3957

Delta R.T. 0.041 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Instrument :

BNA_M

ClientSampleId :

SSTD160023





Tgt Ion:278 Resp: 7919866

Ion Ratio Lower Upper

278 100

139 14.6 11.0 16.4

279 24.1 18.1 27.1



#96

Benzo(g,h,i)perylene

Concen: 152.390 ng/ul

RT: 27.324 min Scan# 4095

Delta R.T. 0.059 min

Lab File: BM033663.D

Acq: 06 Jan 2022 14:24

Tgt Ion:276 Resp: 7804186

Ion Ratio Lower Upper

276 100

138 19.8 15.6 23.4

277 23.4 18.7 28.1

