

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
 Data File : BM000125.D
 Acq On : 07 Jan 2015 12:54
 Operator : TP/IZ
 Sample : SSTD3.219
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.219

Quant Time: Jan 08 10:31:01 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 10:05:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3427	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	11826	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	7097	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	11684	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	10294	0.40	ng/ul	0.00
20) Perylene-d12	23.70	264	8645	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.87	266	8122	4.44	ng/ul	Ovalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
Data File : BM000125.D
Acq On : 07 Jan 2015 12:54
Operator : TP/IZ
Sample : SSTD3.219
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.219

Quant Time: Jan 08 10:31:01 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 08 10:05:57 2015
Response via : Initial Calibration

