

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
 Data File : BM000129.D
 Acq On : 07 Jan 2015 15:16
 Operator : TP/IZ
 Sample : MDL-03-W
 Misc : MDL-WATER-0.07PPM-SIM
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-03-W

Manual Integrations
 APPROVED

TEJASKUMAR
 1/9/2015 5:44:47 PM

Quant Time: Jan 08 11:55:03 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 10:41:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3151	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	10285	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	5826	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.23	188	10786	0.40	ng/ul	0.02
16) Chrysene-d12	21.41	240	9189	0.40	ng/ul	0.00
20) Perylene-d12	23.70	264	8389	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	5097	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.25	212	8469	0.36	ng/ul	0.00

Target Compounds						Ovalue
3) Naphthalene	10.68	128	1573	0.06	ng/ul#	89
5) 2-Methylnaphthalene	12.29	142	1015	0.06	ng/ul	97
7) Acenaphthylene	14.20	152	1570	0.05	ng/ul#	89
8) Acenaphthene	14.54	153	1004	0.06	ng/ul	97
9) Fluorene	15.54	166	1208	0.06	ng/ul#	90
11) Pentachlorophenol	16.88	266	58	0.03	ng/ul#	70
12) Phenanthrene	17.26	178	1888	0.06	ng/ul#	89
13) Anthracene	17.35	178	1782	0.06	ng/ul#	86
15) Fluoranthene	19.27	202	2252	0.06	ng/ul	91
17) Pyrene	19.64	202	2297	0.06	ng/ul#	92
18) Benzo(a)anthracene	21.39	228	1848	0.06	ng/ul	93
19) Chrysene	21.44	228	2010m	0.06	ng/ul	
21) Benzo(b)fluoranthene	23.05	252	1775	0.05	ng/ul	67
22) Benzo(k)fluoranthene	23.05	252	1775	0.06	ng/ul#	68
23) Benzo(a)pyrene	23.60	252	1808	0.06	ng/ul#	58
24) Indeno(1,2,3-cd)pyrene	26.05	276	2025	0.06	ng/ul	98
25) Dibenzo(a,h)anthracene	26.07	278	1611	0.06	ng/ul#	64
26) Benzo(g,h,i)perylene	26.75	276	1824	0.06	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
Data File : BM000129.D
Acq On : 07 Jan 2015 15:16
Operator : TP/IZ
Sample : MDL-03-W
Misc : MDL-WATER-0.07PPM-SIM
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-03-W

Manual Integrations
APPROVED

TEJASKUMAR
1/9/2015 5:44:47 PM

Quant Time: Jan 08 11:55:03 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 08 10:41:47 2015
Response via : Initial Calibration

