

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
 Data File : BM000133.D
 Acq On : 07 Jan 2015 17:38
 Operator : TP/IZ
 Sample : MDL-07-W
 Misc : MDL-WATER-0.07PPM-SIM
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-07-W

Quant Time: Jan 08 11:58:42 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 10:41:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	2794	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	9176	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	5482	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.23	188	10376	0.40	ng/ul	0.02
16) Chrysene-d12	21.41	240	8708	0.40	ng/ul	0.00
20) Perylene-d12	23.70	264	7955	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.22	152	4714	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.25	212	8362	0.37	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Naphthalene	10.68	128	1377	0.06	ng/ul#	87
5) 2-Methylnaphthalene	12.30	142	868	0.05	ng/ul	96
7) Acenaphthylene	14.20	152	1426	0.05	ng/ul#	88
8) Acenaphthene	14.54	153	937	0.05	ng/ul	96
9) Fluorene	15.54	166	1058	0.05	ng/ul#	95
11) Pentachlorophenol	16.90	266	44	0.03	ng/ul#	82
12) Phenanthrene	17.26	178	1694	0.05	ng/ul#	83
13) Anthracene	17.35	178	1645	0.05	ng/ul#	77
15) Fluoranthene	19.28	202	2030	0.06	ng/ul	91
17) Pyrene	19.65	202	2037	0.06	ng/ul#	93
18) Benzo(a)anthracene	21.39	228	1554	0.05	ng/ul	91
19) Chrysene	21.44	228	1764	0.06	ng/ul	95
21) Benzo(b)fluoranthene	23.00	252	1665	0.05	ng/ul#	63
22) Benzo(k)fluoranthene	23.05	252	1694	0.06	ng/ul#	64
23) Benzo(a)pyrene	23.61	252	1534	0.05	ng/ul#	48
24) Indeno(1,2,3-cd)pyrene	26.05	276	1778	0.05	ng/ul	97
25) Dibenzo(a,h)anthracene	26.08	278	1346	0.05	ng/ul#	54
26) Benzo(g,h,i)perylene	26.75	276	1541	0.05	ng/ul#	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
Data File : BM000133.D
Acq On : 07 Jan 2015 17:38
Operator : TP/IZ
Sample : MDL-07-W
Misc : MDL-WATER-0.07PPM-SIM
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-07-W

Quant Time: Jan 08 11:58:42 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 08 10:41:47 2015
Response via : Initial Calibration

