

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
 Data File : BM000143.D
 Acq On : 08 Jan 2015 01:18
 Operator : TP/IZ
 Sample : SSTD0.421
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.421

Quant Time: Jan 08 12:13:21 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 08 10:41:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3102	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	10088	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	5924	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	10901	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	9472	0.40	ng/ul	-0.01
20) Perylene-d12	23.69	264	8624	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.22	152	5178	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.25	212	9263	0.39	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Naphthalene	10.67	128	10560	0.39	ng/ul	97
5) 2-Methylnaphthalene	12.28	142	6980	0.39	ng/ul	100
7) Acenaphthylene	14.20	152	11370	0.38	ng/ul	97
8) Acenaphthene	14.52	153	7114	0.39	ng/ul#	83
9) Fluorene	15.52	166	8459	0.40	ng/ul	96
11) Pentachlorophenol	16.87	266	505	0.30	ng/ul	94
12) Phenanthrene	17.26	178	12994	0.39	ng/ul	99
13) Anthracene	17.35	178	12511	0.40	ng/ul	98
15) Fluoranthene	19.27	202	14512	0.39	ng/ul	95
17) Pyrene	19.63	202	14903	0.38	ng/ul	97
18) Benzo(a)anthracene	21.39	228	12001	0.38	ng/ul	96
19) Chrysene	21.43	228	13494	0.39	ng/ul	95
21) Benzo(b)fluoranthene	22.99	252	13766	0.41	ng/ul	99
22) Benzo(k)fluoranthene	23.04	252	11657	0.36	ng/ul	98
23) Benzo(a)pyrene	23.58	252	11806	0.38	ng/ul	99
24) Indeno(1,2,3-cd)pyrene	26.02	276	13679	0.39	ng/ul	99
25) Dibenzo(a,h)anthracene	26.05	278	10959	0.39	ng/ul	99
26) Benzo(g,h,i)perylene	26.73	276	12138	0.39	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010715\
Data File : BM000143.D
Acq On : 08 Jan 2015 01:18
Operator : TP/IZ
Sample : SSTD0.421
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.421

Quant Time: Jan 08 12:13:21 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM010715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 08 10:41:47 2015
Response via : Initial Calibration

