

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010816\
 Data File : BM003929.D
 Acq On : 08 Jan 2016 19:09
 Operator : SJ/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 09 00:54:45 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 04 13:33:21 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.481	0.438	8.9	82	0.00
3	n-Nitrosodimethylamine	0.517	0.451	12.8	72	0.00
4 S	2-Fluorophenol	1.023	0.928	9.3	80	0.00
5 S	Phenol-d6	1.253	1.158	7.6	78	0.00
6	bis(2-Chloroethyl)ether	1.145	1.108	3.2	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
8 S	Nitrobenzene-d5	0.246	0.210	14.6	71	0.00
9	Nitrobenzene	0.266	0.241	9.4	76	0.00
10	Naphthalene	1.094	1.055	3.6	79	0.00
11	Hexachlorobutadiene	0.167	0.160	4.2	80	0.01
12	2-Methylnaphthalene	0.739	0.675	8.7	75	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	82	-0.02
14 S	2,4,6-Tribromophenol	0.116	0.105	9.5	80	0.00
15 S	2-Fluorobiphenyl	1.612	1.453	9.9	75	0.00
16	Acenaphthylene	2.086	1.931	7.4	76	-0.02
17	Acenaphthene	1.457	1.441	1.1	87	0.00
18	Fluorene	1.734	1.668	3.8	78	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
20	4-Bromophenyl-phenylether	0.155	0.174	-12.3	90	-0.02
21	Hexachlorobenzene	0.164	0.194	-18.3	98	-0.02
22	Pentachlorophenol	0.043	0.034	20.9	74	0.00
23	Phenanthrene	1.037	1.134	-9.4	86	-0.02
24	Anthracene	1.006	1.096	-8.9	88	0.00
25	Fluoranthene	1.113	1.145	-2.9	80	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	97	0.01
27	Benzydine	0.418	0.000	100.0#	0#	-19.33#
28	Pvrene	1.796	1.553	13.5	83	0.00
29 S	Terphenyl-d14	0.809	0.715	11.6	86	-0.01
30	Benzo(a)anthracene	1.421	1.247	12.2	87	0.01
31	3,3'-Dichlorobenzidine	0.361	0.205	43.2#	59	0.00
32	Chrysene	1.475	1.375	6.8	89	-0.01
33	Bis(2-ethylhexyl)phthalate	0.629	0.568	9.7	108	-0.01
34	Indeno(1,2,3-cd)pyrene	1.057	0.903	14.6	83	0.00
35 I	Perylene-d12	1.000	1.000	0.0	95	-0.01
36	Benzo(b)fluoranthene	1.528	1.482	3.0	99	0.00
37	Benzo(k)fluoranthene	1.480	1.320	10.8	87	0.00
38 C	Benzo(a)pyrene	1.301	1.198	7.9	93	0.00
39	Dibenzo(a,h)anthracene	1.110	0.959	13.6	82	-0.01
40	Benzo(a,h,i)perylene	1.176	1.027	12.7	84	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010816\
Data File : BM003929.D
Acq On : 08 Jan 2016 19:09
Operator : SJ/UM
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Jan 09 00:54:45 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 04 13:33:21 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				