

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010916\
 Data File : BM003931.D
 Acq On : 10 Jan 2016 09:28
 Operator : SJ/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 11 00:49:57 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 04 13:33:21 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.481	0.460	4.4	83	0.00
3	n-Nitrosodimethylamine	0.517	0.439	15.1	68	0.00
4 S	2-Fluorophenol	1.023	0.952	6.9	79	0.00
5 S	Phenol-d6	1.253	1.137	9.3	74	0.00
6	bis(2-Chloroethyl)ether	1.145	1.084	5.3	74	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
8 S	Nitrobenzene-d5	0.246	0.211	14.2	64	0.00
9	Nitrobenzene	0.266	0.241	9.4	68	-0.01
10	Naphthalene	1.094	1.052	3.8	71	-0.02
11	Hexachlorobutadiene	0.167	0.162	3.0	73	0.00
12	2-Methylnaphthalene	0.739	0.635	14.1	63	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	63	0.00
14 S	2,4,6-Tribromophenol	0.116	0.090	22.4	52	0.00
15 S	2-Fluorobiphenyl	1.612	1.567	2.8	61	0.00
16	Acenaphthylene	2.086	1.930	7.5	57	0.00
17	Acenaphthene	1.457	1.427	2.1	65	0.00
18	Fluorene	1.734	1.648	5.0	59	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
20	4-Bromophenyl-phenylether	0.155	0.140	9.7	58	0.00
21	Hexachlorobenzene	0.164	0.147	10.4	60	0.00
22	Pentachlorophenol	0.043	0.031	27.9#	55	0.00
23	Phenanthrene	1.037	0.952	8.2	58	0.00
24	Anthracene	1.006	0.946	6.0	61	0.00
25	Fluoranthene	1.113	1.020	8.4	58	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
27	Benzydine	0.418	0.475	-13.6	91	0.00
28	Pvrene	1.796	1.418	21.0	58	0.00
29 S	Terphenyl-d14	0.809	0.638	21.1	58	0.00
30	Benzo(a)anthracene	1.421	1.272	10.5	68	0.00
31	3,3'-Dichlorobenzidine	0.361	0.304	15.8	67	0.00
32	Chrysene	1.475	1.438	2.5	71	0.00
33	Bis(2-ethylhexyl)phthalate	0.629	0.436	30.7#	63	0.00
34	Indeno(1,2,3-cd)pyrene	1.057	1.289	-21.9	90	0.01
35 I	Perylene-d12	1.000	1.000	0.0	89	0.00
36	Benzo(b)fluoranthene	1.528	1.315	13.9	82	0.00
37	Benzo(k)fluoranthene	1.480	1.333	9.9	82	0.00
38 C	Benzo(a)pyrene	1.301	1.185	8.9	86	0.00
39	Dibenzo(a,h)anthracene	1.110	1.122	-1.1	91	0.00
40	Benzo(a,h,i)perylene	1.176	1.206	-2.6	93	0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010916\
Data File : BM003931.D
Acq On : 10 Jan 2016 09:28
Operator : SJ/UM
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Jan 11 00:49:57 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 04 13:33:21 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				