

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010916\
 Data File : BM003933.D
 Acq On : 10 Jan 2016 11:19
 Operator : SJ/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 11 00:59:50 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 04 13:33:21 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	71	0.00
2	1,4-Dioxane	1.000	1.021	-2.1	72	0.00
3	n-Nitrosodimethylamine	1.000	0.849	15.1	60	0.00
4 S	2-Fluorophenol	1.000	0.895	10.5	67	0.00
5 S	Phenol-d6	1.000	0.901	9.9	65	0.00
6	bis(2-Chloroethyl)ether	1.000	0.944	5.6	66	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	69	0.00
8 S	Nitrobenzene-d5	1.000	0.835	16.5	58	0.00
9	Nitrobenzene	1.000	0.885	11.5	62	0.00
10	Naphthalene	1.000	0.978	2.2	68	0.00
11	Hexachlorobutadiene	1.000	0.966	3.4	68	0.00
12	2-Methylnaphthalene	1.000	0.904	9.6	62	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	66	0.00
14 S	2,4,6-Tribromophenol	1.000	0.816	18.4	55	0.00
15 S	2-Fluorobiphenyl	1.000	0.933	6.7	62	0.00
16	Acenaphthylene	1.000	0.917	8.3	60	0.00
17	Acenaphthene	1.000	0.929	7.1	65	0.00
18	Fluorene	1.000	0.966	3.4	63	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	64	0.00
20	4-Bromophenyl-phenylether	1.000	0.975	2.5	67	0.00
21	Hexachlorobenzene	1.000	0.930	7.0	66	0.00
22	Pentachlorophenol	1.000	0.745	25.5#	47	0.00
23	Phenanthrene	1.000	1.046	-4.6	66	0.00
24	Anthracene	1.000	0.923	7.7	64	0.00
25	Fluoranthene	1.000	0.951	4.9	63	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	71	0.00
27	Benzydine	1.000	1.173	-17.3	85	0.00
28	Pvrene	1.000	0.912	8.8	64	0.00
29 S	Terphenyl-d14	1.000	0.935	6.5	67	0.00
30	Benzo(a)anthracene	1.000	0.906	9.4	66	0.02
31	3,3'-Dichlorobenzidine	1.000	0.867	13.3	63	0.00
32	Chrysene	1.000	1.106	-10.6	77	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	0.904	9.6	71	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	0.924	7.6	66	0.01
35 I	Perylene-d12	5.000	5.000	0.0	76	0.00
36	Benzo(b)fluoranthene	1.000	0.995	0.5	81	0.00
37	Benzo(k)fluoranthene	1.000	0.857	14.3	66	0.01
38 C	Benzo(a)pyrene	1.000	0.894	10.6	72	0.00
39	Dibenzo(a,h)anthracene	1.000	0.860	14.0	65	0.00
40	Benzo(a,h,i)perylene	1.000	0.885	11.5	68	0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM010916\
Data File : BM003933.D
Acq On : 10 Jan 2016 11:19
Operator : SJ/UM
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Jan 11 00:59:50 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM010216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 04 13:33:21 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				