


Data Path : Z:\HPCHEM1\BNA M\DATA\BM010916\
 Data File : BM003930.D
 Acq On : 10 Jan 2016 08:53
 Operator : SJ/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM010216.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 04 13:33:21 2016



AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.2	15923	PASS
68	69	0.00	2	1.2	231	PASS
69	198	0.00	100	35.8	18906	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.2	25421	PASS
197	198	0.00	2	0.2	114	PASS
198	198	100	100	100.0	52754	PASS
199	198	5	9	6.8	3574	PASS
275	198	10	60	25.2	13300	PASS
365	198	1	100	2.8	1501	PASS
441	442	0.01	24	15.2	6206	PASS
442	198	50	100	77.3	40760	PASS
443	442	15	24	19.7	8016	PASS