

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010923\
 Data File : BM038411.D
 Acq On : 09 Jan 2023 12:45
 Operator : CG/JU
 Sample : PB150061BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK061

Quant Time: Jan 09 13:59:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:11:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.992	152	64633	20.000	ng/u1	0.00
20) Naphthalene-d8	10.804	136	241738	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.627	164	146852	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.368	188	324521	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	300863	20.000	ng/u1	0.00
88) Perylene-d12	23.968	264	369202	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	11436	5.738	ng/uL	0.00
4) Pyridine-d5	3.793	84	167747	29.092	ng/u1	0.00
7) Phenol-d5	7.151	99	165169	28.052	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.322	67	128395	30.313	ng/u1	0.00
11) 2-Chlorophenol-d4	7.522	132	125651	29.784	ng/u1	0.00
15) 4-Methylphenol-d8	8.704	113	124901	27.481	ng/u1	-0.01
21) Nitrobenzene-d5	9.163	128	57189	30.154	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.886	143	64312	29.899	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.428	165	114968	27.678	ng/u1	0.00
31) 4-Chloroaniline-d4	10.951	131	149770	26.130	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	349500	31.426	ng/u1	0.00
49) Acenaphthylene-d8	14.321	160	382855	29.540	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	46282	23.117	ng/u1	0.00
60) Fluorene-d10	15.616	176	295511	29.218	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	56965	24.220	ng/u1	0.00
73) Anthracene-d10	17.468	188	483344	30.875	ng/u1	0.00
81) Pyrene-d10	19.756	212	577253	34.479	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	571363	31.159	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010923\
Data File : BM038411.D
Acq On : 09 Jan 2023 12:45
Operator : CG/JU
Sample : PB150061BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK061

Quant Time: Jan 09 13:59:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 02:11:50 2023
Response via : Initial Calibration

