

Data Path : Z:\HPCHEM1\BNA M\DATA\BM011017\
 Data File : BM008751.D
 Acq On : 09 Jan 2017 22:46
 Operator : UM/SJ
 Sample : I1061-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA972

Quant Time: Jan 10 01:15:31 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM010617.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 10 00:37:03 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.95	152	107100	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.76	136	434535	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.59	164	326660	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.33	188	725831	20.00	ng/ul	0.00
75) Chrysene-d12	21.49	240	928308	20.00	ng/ul	0.00
83) Perylene-d12	23.86	264	856115	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.41	96	3906	1.68	ng/uL	0.00
5) Phenol-d5	7.10	99	59487	6.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.29	67	183068	30.66	ng/ul	0.00
9) 2-Chlorophenol-d4	7.49	132	152597	24.90	ng/ul	0.00
13) 4-Methylphenol-d8	8.65	113	106609	15.62	ng/ul	0.00
19) Nitrobenzene-d5	9.12	128	90054	33.58	ng/ul	0.00
22) 2-Nitrophenol-d4	9.84	143	104076	35.71	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.37	165	194291	28.95	ng/ul	0.00
29) 4-Chloroaniline-d4	10.89	131	7658	1.03	ng/ul	0.00
43) Dimethylphthalate-d6	13.99	166	711144	33.00	ng/ul	0.00
46) Acenaphthylene-d8	14.28	160	839896	33.05	ng/ul	0.00
51) 4-Nitrophenol-d4	14.76	143	20478	5.81	ng/ul	0.00
57) Fluorene-d10	15.57	176	629648	31.68	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.69	200	116228	31.07	ng/ul	0.00
70) Anthracene-d10	17.43	188	1061033	33.93	ng/ul	0.00
76) Pyrene-d10	19.71	212	1264131	30.91	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.70	264	1158318	33.17	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM011017\
Data File : BM008751.D
Acq On : 09 Jan 2017 22:46
Operator : UM/SJ
Sample : I1061-06
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA972

Quant Time: Jan 10 01:15:31 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM010617.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 10 00:37:03 2017
Response via : Initial Calibration

