

Data Path : Z:\HPCHEM1\BNA M\DATA\BM011017\
 Data File : BM008749.D
 Acq On : 09 Jan 2017 21:35
 Operator : UM/SJ
 Sample : I1061-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA968

Quant Time: Jan 10 01:10:45 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM010617.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 10 00:37:03 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.96	152	96821	20.00	ng/ul	0.00
18) Naphthalene-d8	10.76	136	392517	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.59	164	288178	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.33	188	663320	20.00	ng/ul	0.00
75) Chrysene-d12	21.49	240	877326	20.00	ng/ul	0.00
83) Perylene-d12	23.85	264	804320	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.42	96	3276	1.56	ng/uL	0.00
5) Phenol-d5	7.10	99	56611	7.12	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.29	67	160116	29.66	ng/ul	0.00
9) 2-Chlorophenol-d4	7.49	132	132875	23.99	ng/ul	0.00
13) 4-Methylphenol-d8	8.65	113	92160	14.94	ng/ul	0.00
19) Nitrobenzene-d5	9.12	128	75809	31.30	ng/ul	0.00
22) 2-Nitrophenol-d4	9.84	143	81104	30.81	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.37	165	170884	28.19	ng/ul	0.00
29) 4-Chloroaniline-d4	10.89	131	4927	0.73	ng/ul	0.00
43) Dimethylphthalate-d6	13.99	166	587081	30.88	ng/ul	0.00
46) Acenaphthylene-d8	14.28	160	682675	30.45	ng/ul	0.00
51) 4-Nitrophenol-d4	14.76	143	19123	6.15	ng/ul	0.00
57) Fluorene-d10	15.57	176	521338	29.74	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.68	200	94564	27.66	ng/ul	0.00
70) Anthracene-d10	17.43	188	878296	30.73	ng/ul	0.00
76) Pyrene-d10	19.71	212	1103988	28.56	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.70	264	1020201	31.10	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Phenol	7.13	94	47379	5.54	ng/ul#	78
44) Dimethylphthalate	14.04	163	62083	3.27	ng/ul#	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM011017\
Data File : BM008749.D
Acq On : 09 Jan 2017 21:35
Operator : UM/SJ
Sample : I1061-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA968

Quant Time: Jan 10 01:10:45 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM010617.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jan 10 00:37:03 2017
Response via : Initial Calibration





#6

Phenol

Concen: 5.54 ng/ul

RT: 7.13 min Scan# 773

Delta R.T. 0.00 min

Lab File: BM008749.D

Acq: 09 Jan 2017 21:35

Instrument :

BNA_M

ClientSampleId :

DA968





Tot Ion: 94 Resp: 47379

Ion Ratio Lower Upper

94 100

65 39.0 23.1 34.7#

66 57.8 33.7 50.5#



#44

Dimethylphthalate

Concen: 3.27 ng/ul

RT: 14.04 min Scan# 1947

Delta R.T. -0.01 min

Lab File: BM008749.D

Acq: 09 Jan 2017 21:35

Tgt Ion: 163 Resp: 62083

Ion Ratio Lower Upper

163 100

194 3.0 3.4 5.0#

164 11.8 8.2 12.4

