




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011022\
 Data File : BM033685.D
 Acq On : 10 Jan 2022 14:14
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 11 03:03:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 05 17:07:26 2022
 Response via : Initial Calibration






TIC: BM033685.D\data.ms

(77) Benzidine**19.577min (-0.006) 40159.95 ng****response 1184827**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	13.77
183.00	12.10	12.20
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011022\
Data File : BM033685.D
Acq On : 10 Jan 2022 14:14
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 11 03:03:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM010522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 05 17:07:26 2022
Response via : Initial Calibration



TIC: BM033685.D\data.ms

(70) Pentachlorophenol (C)

16.995min (-0.011) 0.00 ng

response 349684

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	63.88
264.00	61.40	63.74
0.00	0.00	0.00