

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011024\
 Data File : BM043813.D
 Acq On : 10 Jan 2024 15:06
 Operator : MA/JU
 Sample : SSTD3.289
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2041

Quant Time: Jan 10 15:38:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 10 14:15:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	7932	0.400	ng/ul	0.00
4) Naphthalene-d8	10.763	136	23842	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.597	164	14231	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.350	188	30095	0.400	ng/ul	0.00
17) Chrysene-d12	21.539	240	20787	0.400	ng/ul	0.00
23) Perylene-d12	23.974	264	21760	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.360	96	30007	3.007	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.393	88	32778	3.238	ng/ul#	80
14) Pentachlorophenol	17.004	266	30491	4.343	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011024\
Data File : BM043813.D
Acq On : 10 Jan 2024 15:06
Operator : MA/JU
Sample : SST3.289
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2041

Quant Time: Jan 10 15:38:05 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 10 14:15:09 2024
Response via : Initial Calibration

