

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011024\
 Data File : BM043814.D
 Acq On : 10 Jan 2024 15:43
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV042

Quant Time: Jan 10 16:11:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 10 15:40:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.954	152	8289	0.400	ng/ul	0.00
4) Naphthalene-d8	10.774	136	24410	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	14878	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.354	188	32664	0.400	ng/ul	0.00
17) Chrysene-d12	21.545	240	22617	0.400	ng/ul	0.00
23) Perylene-d12	23.974	264	24632	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	4439	0.443	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.368	152	15453	0.438	ng/ul	0.00
18) Fluoranthene-d10	19.380	212	33355	0.456	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.398	88	4634	0.435	ng/ul	93
5) Naphthalene	10.823	128	32803	0.468	ng/ul	99
7) 2-Methylnaphthalene	12.440	142	21612	0.469	ng/ul	100
8) 1-Methylnaphthalene	12.654	142	21901	0.468	ng/ul	99
10) Acenaphthylene	14.324	152	38182	0.479	ng/ul	100
11) Acenaphthene	14.662	153	26949	0.465	ng/ul	100
12) Fluorene	15.656	166	31548	0.475	ng/ul	99
14) Pentachlorophenol	17.016	266	3719	0.359	ng/ul	97
15) Phenanthrene	17.396	178	48958	0.472	ng/ul	100
16) Anthracene	17.493	178	46000	0.466	ng/ul	100
19) Fluoranthene	19.413	202	54978	0.491	ng/ul	100
20) Pyrene	19.775	202	56591	0.485	ng/ul	100
21) Benzo(a)anthracene	21.530	228	44577	0.454	ng/ul	100
22) Chrysene	21.583	228	48111	0.455	ng/ul	100
24) Benzo(b)fluoranthene	23.228	252	47984	0.478	ng/ul	97
25) Benzo(k)fluoranthene	23.278	252	48347	0.437	ng/ul	99
26) Benzo(a)pyrene	23.866	252	41509	0.441	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.505	276	54512	0.458	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.525	278	42348	0.459	ng/ul	99
29) Benzo(g,h,i)perylene	27.280	276	44639	0.449	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011024\
Data File : BM043814.D
Acq On : 10 Jan 2024 15:43
Operator : MA/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SICV042

Quant Time: Jan 10 16:11:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM011024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 10 15:40:11 2024
Response via : Initial Calibration

